

Surrogate Summary

SDG No.: Q3551

Client: Spectra East Inc.

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3551-01	MANHOLE	1,2-Dichloroethane-d4	30	28.4	95		91	110
		Toluene-d8	30	29.4	98		91	112
		4-Bromofluorobenzene	30	26.0	87		63	112
VN1114WBL01	VN1114WBL01	1,2-Dichloroethane-d4	30	30.6	102		91	110
		Toluene-d8	30	29.6	98		91	112
		4-Bromofluorobenzene	30	26.5	88		63	112
VN1114WBS01	VN1114WBS01	1,2-Dichloroethane-d4	30	29.7	99		91	110
		Toluene-d8	30	30.5	102		91	112
		4-Bromofluorobenzene	30	29.6	99		63	112
VN1114WBSD0	VN1114WBSD01	1,2-Dichloroethane-d4	30	29.5	98		91	110
		Toluene-d8	30	30.6	102		91	112
		4-Bromofluorobenzene	30	29.9	100		63	112