

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3551</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Spectra East Inc.</u>	Datafile :	<u>VN088283.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1114WBS01	Methylene Chloride	20	20.5	ug/L	103			60	140	
	trans-1,2-Dichloroethene	20	19.1	ug/L	96			70	130	
	Chloroform	20	19.4	ug/L	97			70	135	
	1,1,1-Trichloroethane	20	19.0	ug/L	95			70	130	
	Trichloroethene	20	19.3	ug/L	97			65	135	
	Toluene	20	19.7	ug/L	99			70	130	
	Tetrachloroethene	20	18.7	ug/L	94			70	130	
	Ethyl Benzene	20	19.2	ug/L	96			60	140	
	m/p-Xylenes	40	38.7	ug/L	97			87	111	
	o-Xylene	20	18.9	ug/L	95			87	111	
	1,3-Dichlorobenzene	20	19.4	ug/L	97			70	130	