

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3551</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Spectra East Inc.</u>	Datafile :	<u>VN088284.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1114WBSD01	Methylene Chloride	20	19.6	ug/L	98	5		60	140	20
	trans-1,2-Dichloroethene	20	19.2	ug/L	96	0		70	130	20
	Chloroform	20	19.5	ug/L	98	1		70	135	20
	1,1,1-Trichloroethane	20	19.3	ug/L	97	2		70	130	20
	Trichloroethene	20	19.5	ug/L	98	1		65	135	20
	Toluene	20	20.0	ug/L	100	1		70	130	20
	Tetrachloroethene	20	20.0	ug/L	100	6		70	130	20
	Ethyl Benzene	20	19.6	ug/L	98	2		60	140	20
	m/p-Xylenes	40	38.6	ug/L	97	0		87	111	20
	o-Xylene	20	19.6	ug/L	98	3		87	111	20
	1,3-Dichlorobenzene	20	19.9	ug/L	100	3		70	130	20