

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SPEC01
Lab Code: ACE SDG NO.: Q3551
Lab File ID: VN088282.D Date Analyzed: 11/14/2025
Instrument ID: MSVOA_N Time Analyzed: 09:07
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	56753	7.79	292349	9.08	258679	11.85
UPPER LIMIT	113506	8.294	584698	9.582	517358	12.347
LOWER LIMIT	28376.5	7.294	146175	8.582	129340	11.347
EPA SAMPLE NO.						
MANHOLE	53480	7.79	279713	9.08	245018	11.84
VN1114WBL01	53945	7.79	285107	9.08	248105	11.84
VN1114WBS01	59304	7.79	313554	9.08	280371	11.84
VN1114WBSD01	57783	7.79	307983	9.08	272076	11.85

IS1 = Bromochloromethane

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.