

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3552
Project: Edison Landfill	Sample ID: Q3534-01
Client ID: MW-6DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.047	U	0.047	U	1	0		11/07/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Edison Landfill Client ID: MW-6MSD	SDG No.: Q3552 Sample ID: Q3534-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	1.80		1.80		1	0		11/07/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Edison Landfill Client ID: MANHOLEDUP	SDG No.: Q3552 Sample ID: Q3551-04 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	25.9		25.8		1	0.19		11/06/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3552
Project:	Edison Landfill	Sample ID:	Q3552-01
Client ID:	MW-4DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	6.83		6.85		1	0.29		11/06/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3552
Project:	Edison Landfill	Sample ID:	Q3552-04
Client ID:	MW-10DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		11/05/2025
TKN	mg/L	+/-20	2.90		2.70		1	7		11/12/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3552
Project: Edison Landfill	Sample ID: Q3552-04
Client ID: MW-10MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.94		0.95		2	0.21		11/05/2025
Fluoride	mg/L	+/-20	2.10		2.10		1	0		11/05/2025
Sulfate	mg/L	+/-20	1210	OR	1210	OR	1	0		11/05/2025
Chloride	mg/L	+/-20	0.19	U	0.19	U	1	0		11/05/2025
Bromide	mg/L	+/-20	37.5	OR	37.2	OR	1	1		11/05/2025
Orthophosphate as P	mg/L	+/-20	1.30		1.10		1	17		11/05/2025
Nitrate	mg/L	+/-20	2.50		2.40		1	4		11/05/2025
Nitrite	mg/L	+/-20	2.90		2.80		1	4		11/05/2025
TKN	mg/L	+/-20	7.30		7.20		1	1		11/12/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3552
Project:	Edison Landfill	Sample ID:	Q3552-05
Client ID:	MW-3DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	1410		1410		1	0.07		11/06/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3552
Project: Edison Landfill	Sample ID: Q3551-01
Client ID: MANHOLEMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	25.9		26.0		1	0.39		11/07/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3552
Project:	Edison Landfill	Sample ID:	Q3554-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	38.6		38.8		1	0.52		11/07/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Edison Landfill Client ID: DSN003DUP	SDG No.: Q3552 Sample ID: Q3630-05 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	23.9		22.9		1	4.27		11/14/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3552
Project: Edison Landfill	Sample ID: Q3630-05
Client ID: DSN003MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	68.4		67.4		1	1.47		11/14/2025