

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** REMI02  
**Lab Code:** ACE **SDG NO.:** Q3552  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 11/06/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO114951.D **Time Analyzed:** 19:28

| COMPOUND             | RT    | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|--------|--------------------|-------------------|------|
|                      |       | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.486 | 5.387     | 5.587  | 512.950            | 500.000           | 2.6  |
| Aroclor-1016-2       | 5.508 | 5.408     | 5.608  | 504.850            | 500.000           | 1.0  |
| Aroclor-1016-3       | 5.569 | 5.470     | 5.670  | 531.690            | 500.000           | 6.3  |
| Aroclor-1016-4       | 5.665 | 5.566     | 5.766  | 520.950            | 500.000           | 4.2  |
| Aroclor-1016-5       | 5.957 | 5.858     | 6.058  | 497.090            | 500.000           | -0.6 |
| Aroclor-1260-1       | 7.070 | 6.971     | 7.171  | 491.430            | 500.000           | -1.7 |
| Aroclor-1260-2       | 7.324 | 7.225     | 7.425  | 525.920            | 500.000           | 5.2  |
| Aroclor-1260-3       | 7.682 | 7.582     | 7.782  | 524.330            | 500.000           | 4.9  |
| Aroclor-1260-4       | 7.903 | 7.803     | 8.003  | 511.630            | 500.000           | 2.3  |
| Aroclor-1260-5       | 8.210 | 8.110     | 8.310  | 510.830            | 500.000           | 2.2  |
| Decachlorobiphenyl   | 9.928 | 9.830     | 10.030 | 50.820             | 50.000            | 1.6  |
| Tetrachloro-m-xylene | 4.347 | 4.248     | 4.448  | 52.160             | 50.000            | 4.3  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** REMI02  
**Lab Code:** ACE **SDG NO.:** Q3552  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 11/06/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO114951.D **Time Analyzed:** 19:28

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.726 | 4.628     | 4.828 | 493.530            | 500.000           | -1.3 |
| Aroclor-1016-2       | 4.745 | 4.647     | 4.847 | 495.870            | 500.000           | -0.8 |
| Aroclor-1016-3       | 4.920 | 4.821     | 5.021 | 502.760            | 500.000           | 0.6  |
| Aroclor-1016-4       | 4.961 | 4.864     | 5.064 | 498.680            | 500.000           | -0.3 |
| Aroclor-1016-5       | 5.174 | 5.076     | 5.276 | 501.150            | 500.000           | 0.2  |
| Aroclor-1260-1       | 6.204 | 6.106     | 6.306 | 481.910            | 500.000           | -3.6 |
| Aroclor-1260-2       | 6.392 | 6.294     | 6.494 | 484.950            | 500.000           | -3.0 |
| Aroclor-1260-3       | 6.544 | 6.446     | 6.646 | 488.220            | 500.000           | -2.4 |
| Aroclor-1260-4       | 7.015 | 6.917     | 7.117 | 496.560            | 500.000           | -0.7 |
| Aroclor-1260-5       | 7.256 | 7.158     | 7.358 | 494.130            | 500.000           | -1.2 |
| Decachlorobiphenyl   | 8.639 | 8.541     | 8.741 | 52.800             | 50.000            | 5.6  |
| Tetrachloro-m-xylene | 3.647 | 3.548     | 3.748 | 49.810             | 50.000            | -0.4 |