

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076367.D **Time Analyzed:** 18:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.786	5.695	5.895	532.580	500.000	6.5
Aroclor-1016-2	5.807	5.717	5.917	445.900	500.000	-10.8
Aroclor-1016-3	5.870	5.780	5.980	467.410	500.000	-6.5
Aroclor-1016-4	5.967	5.877	6.077	478.000	500.000	-4.4
Aroclor-1016-5	6.259	6.169	6.369	478.990	500.000	-4.2
Aroclor-1260-1	7.378	7.287	7.487	525.640	500.000	5.1
Aroclor-1260-2	7.631	7.540	7.740	512.000	500.000	2.4
Aroclor-1260-3	7.989	7.899	8.099	524.980	500.000	5.0
Aroclor-1260-4	8.217	8.127	8.327	518.980	500.000	3.8
Aroclor-1260-5	8.542	8.452	8.652	500.170	500.000	0.0
Decachlorobiphenyl	10.404	10.319	10.519	47.610	50.000	-4.8
Tetrachloro-m-xylene	4.635	4.542	4.742	49.100	50.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076367.D **Time Analyzed:** 18:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.869	4.778	4.978	482.510	500.000	-3.5
Aroclor-1016-2	4.928	4.836	5.036	463.880	500.000	-7.2
Aroclor-1016-3	5.048	4.954	5.154	430.090	500.000	-14.0
Aroclor-1016-4	5.090	4.996	5.196	484.090	500.000	-3.2
Aroclor-1016-5	5.302	5.210	5.410	452.950	500.000	-9.4
Aroclor-1260-1	6.329	6.238	6.438	539.390	500.000	7.9
Aroclor-1260-2	6.516	6.425	6.625	537.630	500.000	7.5
Aroclor-1260-3	6.668	6.578	6.778	531.440	500.000	6.3
Aroclor-1260-4	7.137	7.047	7.247	583.090	500.000	16.6
Aroclor-1260-5	7.378	7.287	7.487	543.820	500.000	8.8
Decachlorobiphenyl	8.770	8.682	8.882	45.790	50.000	-8.4
Tetrachloro-m-xylene	3.777	3.681	3.881	49.890	50.000	-0.2