

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q3552	OrderDate:	11/5/2025 2:23:00 PM
Client:	Remington & Vernick Engineers	Project:	Edison Landfill
Contact:	Kyle Carlson	Location:	D41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3552-01	MW-4	WATER	PCB	8082A	11/04/25	11/06/25	11/06/25	11/05/25
Q3552-02	MW-8	WATER	PCB	8082A	11/04/25	11/06/25	11/06/25	11/05/25
Q3552-03	MW-9	WATER	PCB	8082A	11/04/25	11/06/25	11/06/25	11/05/25
Q3552-04	MW-10	WATER	PCB	8082A	11/05/25	11/06/25	11/06/25	11/05/25
Q3552-05	MW-3	WATER	PCB	8082A	11/04/25	11/06/25	11/06/25	11/05/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3552

Order ID: Q3552

Client: Remington & Vernick Engineers

Project ID: Edison Landfill

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: **Q3552**

Client: **Remington & Vernick Engineers**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114870.D	PIBLK-PO114870.D	Tetrachloro-m-xyl	1	20	15.5	77		70 (60)	130 (140)
		Decachlorobiphen	1	20	16.8	84		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	16.2	81		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.3	86		70 (60)	130 (140)
I.BLK-PO114940.D	PIBLK-PO114940.D	Tetrachloro-m-xyl	1	20	18.7	94		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.3	96		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.9	94		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.8	104		70 (60)	130 (140)
PB170425BL	PB170425BL	Tetrachloro-m-xyl	1	20	20.4	102		30 (30)	150 (173)
		Decachlorobiphen	1	20	21.2	106		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	20.6	103		30 (30)	150 (173)
		Decachlorobiphen	2	20	23.0	115		30 (10)	150 (173)
PB170425BS	PB170425BS	Tetrachloro-m-xyl	1	20	21.3	106		30 (30)	150 (173)
		Decachlorobiphen	1	20	21.9	110		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	19.8	99		30 (30)	150 (173)
		Decachlorobiphen	2	20	23.5	117		30 (10)	150 (173)
PB170425BSD	PB170425BSD	Tetrachloro-m-xyl	1	20	21.6	108		30 (30)	150 (173)
		Decachlorobiphen	1	20	22.4	112		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	20.3	101		30 (30)	150 (173)
		Decachlorobiphen	2	20	23.7	119		30 (10)	150 (173)
I.BLK-PO114955.D	PIBLK-PO114955.D	Tetrachloro-m-xyl	1	20	19.1	96		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.6	98		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.2	96		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.9	105		70 (60)	130 (140)
I.BLK-PP075887.D	PIBLK-PP075887.D	Tetrachloro-m-xyl	1	20	19.9	100		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.4	102		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.8	94		70 (60)	130 (140)
		Decachlorobiphen	2	20	19.3	96		70 (60)	130 (140)
I.BLK-PP076356.D	PIBLK-PP076356.D	Tetrachloro-m-xyl	1	20	19.9	100		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.1	100		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.0	90		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.1	85		70 (60)	130 (140)
Q3552-01	MW-4	Tetrachloro-m-xyl	1	20	21.5	108		30 (30)	150 (173)
		Decachlorobiphen	1	20	14.0	70		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	15.3	76		30 (30)	150 (173)
		Decachlorobiphen	2	20	12.5	62		30 (10)	150 (173)
Q3552-02	MW-8	Tetrachloro-m-xyl	1	20	33.4	167	*	30 (30)	150 (173)
		Decachlorobiphen	1	20	30.5	152	*	30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	33.2	166	*	30 (30)	150 (173)
		Decachlorobiphen	2	20	27.2	136		30 (10)	150 (173)
Q3552-03	MW-9	Tetrachloro-m-xyl	1	20	28.8	144		30 (30)	150 (173)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q3552

Client: Remington & Vernick Engineers

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3552-03	MW-9	Decachlorobiphen	1	20	17.3	86		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	22.2	111		30 (30)	150 (173)
Q3552-04	MW-10	Decachlorobiphen	2	20	16.5	83		30 (10)	150 (173)
		Tetrachloro-m-xyl	1	20	29.3	146		30 (30)	150 (173)
		Decachlorobiphen	1	20	27.9	140		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	26.1	130		30 (30)	150 (173)
		Decachlorobiphen	2	20	25.9	130		30 (10)	150 (173)
Q3552-05	MW-3	Tetrachloro-m-xyl	1	20	27.8	139		30 (30)	150 (173)
		Decachlorobiphen	1	20	22.2	111		30 (10)	150 (173)
		Tetrachloro-m-xyl	2	20	22.2	111		30 (30)	150 (173)
		Decachlorobiphen	2	20	20.4	102		30 (10)	150 (173)
I.BLK-PP076371.D	PIBLK-PP076371.D	Tetrachloro-m-xyl	1	20	20.4	102		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.6	98		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.2	91		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.7	88		70 (60)	130 (140)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3552</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Remington & Vernick Engineers</u>	Datafile :	<u>PO114942.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170425BS (Column 1)	AR1016	2.5	2.20	ug/L	88				40 (77)	140 (107)	
	AR1260	2.5	2.10	ug/L	84				40 (66)	140 (113)	
PB170425BS (Column 2)	AR1016	2.5	2.10	ug/L	84				40 (77)	140 (107)	
	AR1260	2.5	2.10	ug/L	84				40 (66)	140 (113)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3552</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Remington & Vernick Engineers</u>	Datafile :	<u>PO114943.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170425BSD (Column 1)	AR1016	2.5	2.20	ug/L	88	0			40 (77)	140 (107)	20 (20)
	AR1260	2.5	2.10	ug/L	84	0			40 (66)	140 (113)	20 (20)
PB170425BSD (Column 2)	AR1016	2.5	2.10	ug/L	84	0			40 (77)	140 (107)	20 (20)
	AR1260	2.5	2.10	ug/L	84	0			40 (66)	140 (113)	20 (20)

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170425BL

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Lab Sample ID: PB170425BL

Lab File ID: PO114941.D

Matrix: (soil/water) WATER

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 11/06/2025

Date Analyzed (1): 11/06/2025

Date Analyzed (2): 11/06/2025

Time Analyzed (1): 15:32

Time Analyzed (2): 15:32

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB170425BS	PB170425BS	PO114942.D	11/06/2025	11/06/2025
PB170425BSD	PB170425BSD	PO114943.D	11/06/2025	11/06/2025
MW-4	Q3552-01	PP076362.D	11/06/2025	11/06/2025
MW-8	Q3552-02	PP076363.D	11/06/2025	11/06/2025
MW-9	Q3552-03	PP076364.D	11/06/2025	11/06/2025
MW-10	Q3552-04	PP076365.D	11/06/2025	11/06/2025
MW-3	Q3552-05	PP076366.D	11/06/2025	11/06/2025

COMMENTS: _____



SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	11/04/25
Project:	Edison Landfill		Date Received:	11/05/25
Client Sample ID:	MW-4		SDG No.:	Q3552
Lab Sample ID:	Q3552-01		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	500 mL	Final Vol: 5000 uL	Test:	PCB
Prep Method:	3510C	Prep Date: 11/06/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/06/25 16:44	PB170425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/06/25 16:44	PB170425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/06/25 16:44	PB170425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/06/25 16:44	PB170425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/06/25 16:44	PB170425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/06/25 16:44	PB170425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/06/25 16:44	PB170425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/06/25 16:44	PB170425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/06/25 16:44	PB170425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.5			30 (30) - 150 (173)	108%	SPK: 20		
2051-24-3	Decachlorobiphenyl	14.0			30 (10) - 150 (173)	70%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
 Data File : PP076362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 16:44
 Operator : YP\AJ
 Sample : Q3552-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MW-4

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/07/2025
 Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:27:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.635	3.770	30770556	101.1E6	21.532m	15.269m#
2) SA Decachlor...	10.405	8.770	15041827	102.7E6	13.973	12.491

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 16:44
Operator : YP\AJ
Sample : Q3552-01
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MW-4

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:27:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP076362.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 4.635 min
 Delta R.T.: -0.007 min
 Response: 30770556
 Conc: 21.53 ng/ml

Instrument :
 ECD_P
 ClientSampleId :
 MW-4

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/07/2025
 Supervised By :Yogesh Patel 11/07/2025


Response_ Signal: PP076362.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.770 min
 Delta R.T.: -0.011 min
 Response: 101056363
 Conc: 15.27 ng/ml m


Response_ Signal: PP076362.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.405 min
 Delta R.T.: -0.014 min
 Response: 15041827
 Conc: 13.97 ng/ml

Response_ Signal: PP076362.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.770 min
 Delta R.T.: -0.012 min
 Response: 102715841
 Conc: 12.49 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers	Date Collected:	11/04/25
Project:	Edison Landfill	Date Received:	11/05/25
Client Sample ID:	MW-8	SDG No.:	Q3552
Lab Sample ID:	Q3552-02	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	500 mL	Test:	PCB
Prep Method:	3510C	Final Vol:	5000 uL
		Prep Date:	11/06/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/06/25 17:00	PB170425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/06/25 17:00	PB170425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/06/25 17:00	PB170425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/06/25 17:00	PB170425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/06/25 17:00	PB170425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/06/25 17:00	PB170425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/06/25 17:00	PB170425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/06/25 17:00	PB170425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/06/25 17:00	PB170425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	33.4	*		30 (30) - 150 (173)	167%	SPK: 20		
2051-24-3	Decachlorobiphenyl	30.5	*		30 (10) - 150 (173)	152%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 17:00
Operator : YP\AJ
Sample : Q3552-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
MW-8

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:28:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.639	3.778	47786490	219.8E6	33.438	33.214
2) SA Decachlor...	10.406	8.773	32812391	223.6E6	30.480	27.190

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 17:00
Operator : YP\AJ
Sample : Q3552-02
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
MW-8

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:28:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP076363.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 47786490
Conc: 33.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
MW-8


Time Response_ Signal: PP076363.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.004 min
Response: 219821566
Conc: 33.21 ng/ml


Time Response_ Signal: PP076363.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.406 min
Delta R.T.: -0.013 min
Response: 32812391
Conc: 30.48 ng/ml

Time Response_ Signal: PP076363.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 223583823
Conc: 27.19 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers	Date Collected:	11/04/25
Project:	Edison Landfill	Date Received:	11/05/25
Client Sample ID:	MW-9	SDG No.:	Q3552
Lab Sample ID:	Q3552-03	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	500 mL	Test:	PCB
Prep Method:	3510C	Final Vol:	5000 uL
		Prep Date:	11/06/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/06/25 17:17	PB170425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/06/25 17:17	PB170425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/06/25 17:17	PB170425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/06/25 17:17	PB170425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/06/25 17:17	PB170425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/06/25 17:17	PB170425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/06/25 17:17	PB170425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/06/25 17:17	PB170425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/06/25 17:17	PB170425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	28.8			30 (30) - 150 (173)	144%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.3			30 (10) - 150 (173)	86%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
 Data File : PP076364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 17:17
 Operator : YP\AJ
 Sample : Q3552-03
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MW-9

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/07/2025
 Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:28:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.638	3.773	41169904	146.7E6	28.808m	22.169m
2) SA Decachlor...	10.408	8.774	18618595	136.0E6	17.295	16.533

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 17:17
Operator : YP\AJ
Sample : Q3552-03
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MW-9


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:28:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.638 min
Delta R.T.: -0.004 min
Response: 41169904
Conc: 28.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
MW-9


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025




#1 Tetrachloro-m-xylene

R.T.: 3.773 min
Delta R.T.: -0.009 min
Response: 146720762
Conc: 22.17 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.408 min
Delta R.T.: -0.011 min
Response: 18618595
Conc: 17.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.774 min
Delta R.T.: -0.008 min
Response: 135951060
Conc: 16.53 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	11/05/25
Project:	Edison Landfill		Date Received:	11/05/25
Client Sample ID:	MW-10		SDG No.:	Q3552
Lab Sample ID:	Q3552-04		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	500 mL	Final Vol: 5000 uL	Test:	PCB
Prep Method:	3510C	Prep Date: 11/06/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/06/25 17:33	PB170425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/06/25 17:33	PB170425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/06/25 17:33	PB170425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/06/25 17:33	PB170425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/06/25 17:33	PB170425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/06/25 17:33	PB170425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/06/25 17:33	PB170425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/06/25 17:33	PB170425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/06/25 17:33	PB170425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	29.3			30 (30) - 150 (173)	146%	SPK: 20		
2051-24-3	Decachlorobiphenyl	27.9			30 (10) - 150 (173)	140%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
 Data File : PP076365.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 17:33
 Operator : YP\AJ
 Sample : Q3552-04
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MW-10

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/07/2025
 Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:28:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.638	3.775	41858580	172.6E6	29.290m	26.078m
2) SA Decachlor...	10.406	8.773	30092729	213.0E6	27.954	25.899

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076365.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 17:33
Operator : YP\AJ
Sample : Q3552-04
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MW-10


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:28:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.638 min
Delta R.T.: -0.004 min
Response: 41858580
Conc: 29.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
MW-10


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025




#1 Tetrachloro-m-xylene

R.T.: 3.775 min
Delta R.T.: -0.006 min
Response: 172594863
Conc: 26.08 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.406 min
Delta R.T.: -0.013 min
Response: 30092729
Conc: 27.95 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 212964105
Conc: 25.90 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	11/04/25
Project:	Edison Landfill		Date Received:	11/05/25
Client Sample ID:	MW-3		SDG No.:	Q3552
Lab Sample ID:	Q3552-05		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	485 mL	Final Vol: 5000 uL	Test:	PCB
Prep Method:	3510C	Prep Date: 11/06/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.10	U	1	0.10	0.52	ug/L	11/06/25 17:49	PB170425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.52	ug/L	11/06/25 17:49	PB170425
11141-16-5	Aroclor-1232	0.099	U	1	0.099	0.52	ug/L	11/06/25 17:49	PB170425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.52	ug/L	11/06/25 17:49	PB170425
12672-29-6	Aroclor-1248	0.073	U	1	0.073	0.52	ug/L	11/06/25 17:49	PB170425
11097-69-1	Aroclor-1254	0.097	U	1	0.097	0.52	ug/L	11/06/25 17:49	PB170425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.52	ug/L	11/06/25 17:49	PB170425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.52	ug/L	11/06/25 17:49	PB170425
11096-82-5	Aroclor-1260	0.084	U	1	0.084	0.52	ug/L	11/06/25 17:49	PB170425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.8			30 (30) - 150 (173)	139%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.2			30 (10) - 150 (173)	111%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
 Data File : PP076366.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 17:49
 Operator : YP\AJ
 Sample : Q3552-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MW-3

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/07/2025
 Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:28:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.636	3.778	39666227	147.2E6	27.756m	22.239m
2) SA Decachlor...	10.405	8.774	23879995	168.1E6	22.182	20.439

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076366.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 17:49
Operator : YP\AJ
Sample : Q3552-05
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MW-3


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:28:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.636 min
Delta R.T.: -0.006 min
Response: 39666227
Conc: 27.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
MW-3


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 147182192
Conc: 22.24 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.405 min
Delta R.T.: -0.014 min
Response: 23879995
Conc: 22.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.774 min
Delta R.T.: -0.009 min
Response: 168067902
Conc: 20.44 ng/ml



CALIBRATION SUMMARY

Contract: REMI02

SDG NO.: Q3552

Calibration Date(s): **11/04/2025** **11/04/2025**

Calibration Times: 09:39 18:31

LAB FILE ID: RT 1000 = PO114871.D RT 750 = PO114872.D
RT 500 = PO114873.D RT 250 = PO114874.D RT 050 = PO114875.D

[illegible]

[illegible]

Calibration Times: 09:39 18:31

LAB FILE ID: RT 1000 = PO114871.D RT 750 = PO114872.D
RT 500 = PO114873.D RT 250 = PO114874.D RT 050 = PO114875.D

[illegible]

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: REMI02
SDG NO.: Q3552

Calibration Date(s): 11/04/2025 11/04/2025
Calibration Times: 09:39 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO114871.D</u>		CF 750 = <u>PO114872.D</u>			
CF 500 = <u>PO114873.D</u>		CF 250 = <u>PO114874.D</u>		CF 050 = <u>PO114875.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	235183609	242405119	252300612	252330724	247880040	246020021
Aroclor-1016-2	(2)	380671765	391174169	403435360	408228592	401844700	397070917
Aroclor-1016-3	(3)	214892716	220811199	230510330	231508016	217282560	223000964
Aroclor-1016-4	(4)	180735079	184273587	191842228	192742164	194782360	188875084
Aroclor-1016-5	(5)	166479276	173610481	179587174	180772736	190536000	178197133
Aroclor-1260-1	(1)	291080181	301659089	311914366	329562032	293951540	305633442
Aroclor-1260-2	(2)	393658847	402595176	414352186	418536840	350242960	395877202
Aroclor-1260-3	(3)	331096941	340266576	347857958	348147780	291720640	331817979
Aroclor-1260-4	(4)	303823322	310642864	321126472	321507392	268000240	305020058
Aroclor-1260-5	(5)	714393776	724690244	737226162	719415648	670367440	713218654
Decachlorobiphenyl		4451734610	4516548787	4612561720	4524869360	3995527600	4420248415
Tetrachloro-m-xylene		7020403280	7162783533	7307727940	7156075200	6543618600	7038121711
Aroclor-1242-1	(1)	180395803	184480117	189796832	192180524	184350480	186240751
Aroclor-1242-2	(2)	290607708	295647931	305830174	302003732	299791360	298776181
Aroclor-1242-3	(3)	164449629	167684461	173929682	174410736	176369820	171368866
Aroclor-1242-4	(4)	136844790	140633719	146402846	139660276	152376540	143183634
Aroclor-1242-5	(5)	138362947	141839337	145002188	138385972	141196180	140957325
Decachlorobiphenyl		4729850460	4804993427	4858375920	4746069120	4160138600	4659885505
Tetrachloro-m-xylene		7348004600	7445204373	7532777260	7262019280	6694579200	7256516943
Aroclor-1248-1	(1)	130126160	135244476	137909000	141742196	154206600	139845686
Aroclor-1248-2	(2)	182893435	187680161	194064692	196392780	211353600	194476934
Aroclor-1248-3	(3)	203368877	206306241	213983954	212927108	217587500	210834736
Aroclor-1248-4	(4)	240181487	246210361	255848644	255957264	264773260	252594203
Aroclor-1248-5	(5)	225312398	230738423	239469078	240115800	243213880	235769916
Decachlorobiphenyl		4590720290	4647354867	4744288880	4661202760	4211361200	4570985599
Tetrachloro-m-xylene		7316541360	7442934853	7528054520	7402608560	7218400400	7381707939
Aroclor-1254-1	(1)	240591852	247416493	258765478	257075152	263299240	253429643
Aroclor-1254-2	(2)	346021204	353529297	371589652	368897016	412184820	370444398
Aroclor-1254-3	(3)	372893964	380446979	397009938	384293176	389678940	384864599
Aroclor-1254-4	(4)	315539209	321424513	336840998	326904952	342871860	328716306
Aroclor-1254-5	(5)	295373361	300690525	312155076	300474784	300573260	301853401
Decachlorobiphenyl		4791769470	4896862187	5061968640	4831244600	4661831200	4848735219
Tetrachloro-m-xylene		7410585440	7526809333	7802184660	7481754040	7408041400	7525874975
Aroclor-1268-1	(1)	876279188	885717404	893041022	887529864	813508780	871215252



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	748334854	756337347	765525536	759915712	702804360	746583562	3
Aroclor-1268-3	(3)	663299706	672085827	679866912	671661004	617886840	660960058	4
Aroclor-1268-4	(4)	274155738	277531193	282310020	282161716	286808980	280593529	2
Aroclor-1268-5	(5)	1966581606	1979440064	1985729368	1956798376	1807416340	1939193151	4
Decachlorobiphenyl		8288385350	8396434587	8504946880	8419357920	7722111200	8266247187	4
Tetrachloro-m-xylene		7475792880	7636683133	7655291620	7597134240	7121440400	7497268455	3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: REMI02
SDG NO.: Q3552

Calibration Date(s): 11/04/2025 11/04/2025
Calibration Times: 09:39 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO114871.D</u>		CF 750 = <u>PO114872.D</u>			
CF 500 = <u>PO114873.D</u>		CF 250 = <u>PO114874.D</u>		CF 050 = <u>PO114875.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	178795127	185535456	193106802	200932132	192907960	190255495
Aroclor-1016-2	(2)	260787154	267435697	277593168	286495324	262010880	270864445
Aroclor-1016-3	(3)	140310175	144533207	150201526	154962192	149858780	147973176
Aroclor-1016-4	(4)	108607533	112021955	117721682	122827936	125909940	117417809
Aroclor-1016-5	(5)	142235561	147493437	153611756	157990124	162236540	152713484
Aroclor-1260-1	(1)	272865489	283696256	302549666	306994580	314431380	296107474
Aroclor-1260-2	(2)	410849028	420744668	440173148	441650472	436738740	430031211
Aroclor-1260-3	(3)	328919579	344471297	362206406	366201912	353957800	351151399
Aroclor-1260-4	(4)	295043629	302500340	313130718	316509244	287959420	303028670
Aroclor-1260-5	(5)	819932115	830411719	851082708	846532168	760181660	821628074
Decachlorobiphenyl		3768213910	3862452533	4051635320	3997632280	3839185200	3903823849
Tetrachloro-m-xylene		4444199150	4532189360	4610968540	4573862320	4172678600	4466779594
Aroclor-1242-1	(1)	138347633	144438088	149108388	148877560	150336280	146221590
Aroclor-1242-2	(2)	199908386	204961789	213926082	211880332	205062860	207147890
Aroclor-1242-3	(3)	108216818	112229521	116418372	113485296	109537600	111977521
Aroclor-1242-4	(4)	101919263	105301932	110004430	109236748	109193360	107131147
Aroclor-1242-5	(5)	135980897	140103160	145653410	150301700	142549620	142917757
Decachlorobiphenyl		4093300380	4301738307	4403735460	4338775160	4380898600	4303689581
Tetrachloro-m-xylene		4800115840	4683347560	4961464680	4527345840	4437059000	4681866584
Aroclor-1248-1	(1)	101027399	105625799	111262386	113093188	115539800	109309714
Aroclor-1248-2	(2)	136540550	142594189	148530518	152692968	164144180	148900481
Aroclor-1248-3	(3)	143887712	149887199	155474250	158349236	169440400	155407759
Aroclor-1248-4	(4)	171266913	177666692	184171754	187513908	195642380	183252329
Aroclor-1248-5	(5)	174649655	180220384	187038042	189642400	190092660	184328628
Decachlorobiphenyl		4076042330	4166088587	4290124020	4403270520	4144788800	4216062851
Tetrachloro-m-xylene		4758509080	4829545453	4892334280	4811209560	4727034000	4803726475
Aroclor-1254-1	(1)	275439068	284238351	300243386	299923620	341055960	300180077
Aroclor-1254-2	(2)	245333366	253279232	267984098	268318040	313122940	269607535
Aroclor-1254-3	(3)	398238109	410614839	430545692	426530308	445294740	422244738
Aroclor-1254-4	(4)	284785979	294895399	305027804	301955520	300306220	297394184
Aroclor-1254-5	(5)	385708423	392326712	409498118	398089984	402006780	397526003
Decachlorobiphenyl		4273373360	4367226933	4604811080	4533103920	4272134600	4410129979
Tetrachloro-m-xylene		4791085320	4887004773	5044010700	4876428000	4843036000	4888312959
Aroclor-1268-1	(1)	948830555	965398547	981732904	993582148	991082540	976125339



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	784312077	796694211	811227528	817006596	768929160	795633914	2
Aroclor-1268-3	(3)	672635507	681509284	695723330	696225136	632763280	675771307	4
Aroclor-1268-4	(4)	270862616	275402580	280086838	282372280	267609400	275266743	2
Aroclor-1268-5	(5)	1768544208	1786014605	1818457210	1831952004	1717683400	1784530285	2
Decachlorobiphenyl		7349980390	7437822467	7706201340	7861559120	8321979800	7735508623	5
Tetrachloro-m-xylene		4791521570	4902547960	4915920660	4883273320	4724156000	4843483902	2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
Instrument ID: ECD_O **Date(s) Analyzed:** 11/04/2025 11/04/2025
GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.55	4.45	4.65	84124000
		2	4.63	4.53	4.73	60183600
		3	4.71	4.61	4.81	193335000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.71	4.61	4.81	150015000
		2	5.22	5.12	5.32	72567400
		3	5.51	5.41	5.61	158675000
		4	5.67	5.57	5.77	73795600
		5	5.76	5.66	5.86	51580600
Aroclor-1262	500	1	7.70	7.60	7.80	453404000
		2	8.23	8.13	8.33	707198000
		3	8.53	8.43	8.63	451420000
		4	8.62	8.52	8.72	336186000
		5	9.24	9.14	9.34	241816000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
Instrument ID: ECD_O **Date(s) Analyzed:** 11/04/2025 11/04/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.86	3.76	3.96	61799200
		2	3.95	3.85	4.05	45637000
		3	4.02	3.92	4.12	140257000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.02	3.92	4.12	111493000
		2	4.75	4.65	4.85	112095000
		3	4.92	4.82	5.02	59654200
		4	5.01	4.91	5.11	50889000
		5	5.18	5.08	5.28	56587600
Aroclor-1262	500	1	6.76	6.66	6.86	494206000
		2	7.26	7.16	7.36	795090000
		3	7.54	7.44	7.64	329028000
		4	7.61	7.51	7.71	538192000
		5	8.10	8.00	8.20	238020000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114871.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 09:39
 Operator : YP/AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 10:39:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 10:34:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	702.0E6	444.4E6	97.995	98.158
2) SA Decachlor...	9.931	8.642	445.2E6	376.8E6	98.226	96.376
Target Compounds						
3) L1 AR-1016-1	5.487	4.728	235.2E6	178.8E6	964.887	961.518
4) L1 AR-1016-2	5.509	4.747	380.7E6	260.8E6	970.969	968.784
5) L1 AR-1016-3	5.571	4.922	214.9E6	140.3E6	964.936	965.952
6) L1 AR-1016-4	5.666	4.964	180.7E6	108.6E6	970.188	959.731
7) L1 AR-1016-5	5.958	5.176	166.5E6	142.2E6	962.123	961.547
31) L7 AR-1260-1	7.072	6.207	291.1E6	272.9E6	965.449	948.413
32) L7 AR-1260-2	7.325	6.394	393.7E6	410.8E6	974.390	965.542
33) L7 AR-1260-3	7.683	6.547	331.1E6	328.9E6	975.314	951.837
34) L7 AR-1260-4	7.904	7.018	303.8E6	295.0E6	972.313	970.260
35) L7 AR-1260-5	8.212	7.259	714.4E6	819.9E6	984.271	981.358



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114871.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:39
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:39:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114872.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 09:58
 Operator : YP/AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 10:42:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 10:34:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	537.2E6	339.9E6	74.991	75.051
2) SA Decachlor...	9.931	8.642	338.7E6	289.7E6	74.828	74.390
Target Compounds						
3) L1 AR-1016-1	5.486	4.728	181.8E6	139.2E6	747.252	748.882
4) L1 AR-1016-2	5.508	4.747	293.4E6	200.6E6	748.878	746.734
5) L1 AR-1016-3	5.571	4.921	165.6E6	108.4E6	745.744	747.508
6) L1 AR-1016-4	5.666	4.964	138.2E6	84016466	744.572	744.934
7) L1 AR-1016-5	5.958	5.176	130.2E6	110.6E6	751.666	748.544
31) L7 AR-1260-1	7.071	6.206	226.2E6	212.8E6	750.268	742.996
32) L7 AR-1260-2	7.326	6.394	301.9E6	315.6E6	748.253	744.378
33) L7 AR-1260-3	7.683	6.546	255.2E6	258.4E6	751.161	748.419
34) L7 AR-1260-4	7.904	7.017	233.0E6	226.9E6	747.063	747.386
35) L7 AR-1260-5	8.211	7.258	543.5E6	622.8E6	749.228	746.944



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:58
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:42:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114873.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 10:16
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 10:36:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 10:34:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	365.4E6	230.5E6	50.000	50.000
2) SA Decachlor...	9.930	8.641	230.6E6	202.6E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.487	4.728	126.2E6	96553401	500.000	500.000
4) L1 AR-1016-2	5.508	4.747	201.7E6	138.8E6	500.000	500.000
5) L1 AR-1016-3	5.570	4.921	115.3E6	75100763	500.000	500.000
6) L1 AR-1016-4	5.666	4.964	95921114	58860841	500.000	500.000
7) L1 AR-1016-5	5.958	5.176	89793587	76805878	500.000	500.000
31) L7 AR-1260-1	7.071	6.206	156.0E6	151.3E6	500.000	500.000
32) L7 AR-1260-2	7.325	6.394	207.2E6	220.1E6	500.000	500.000
33) L7 AR-1260-3	7.682	6.546	173.9E6	181.1E6	500.000	500.000
34) L7 AR-1260-4	7.903	7.017	160.6E6	156.6E6	500.000	500.000
35) L7 AR-1260-5	8.210	7.258	368.6E6	425.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:16
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:36:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114874.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 10:34
 Operator : YP/AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC250

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 10:51:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 10:51:10 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.348	3.649	178.9E6	114.3E6	24.980	25.185
2)	SA Decachlor...	9.931	8.641	113.1E6	99940807	24.991	25.495
Target Compounds							
3)	L1 AR-1016-1	5.487	4.728	63082681	50233033	256.898	264.953
4)	L1 AR-1016-2	5.508	4.747	102.1E6	71623831	257.800	262.284
5)	L1 AR-1016-3	5.571	4.922	57877004	38740548	257.884	262.645
6)	L1 AR-1016-4	5.666	4.964	48185541	30706984	257.129	266.335
7)	L1 AR-1016-5	5.958	5.176	45193184	39497531	258.081	262.734
31)	L7 AR-1260-1	7.071	6.206	82390508	76748645	267.021	260.958m
32)	L7 AR-1260-2	7.325	6.393	104.6E6	110.4E6	256.906	256.862m
33)	L7 AR-1260-3	7.682	6.546	87036945	91550478	254.611	261.237
34)	L7 AR-1260-4	7.904	7.017	80376848	79127311	255.753	257.915
35)	L7 AR-1260-5	8.211	7.258	179.9E6	211.6E6	248.441	252.850

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:34
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660ICC250

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:51:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:51:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 10:53
 Operator : YP/AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 11:09:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 11:09:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.348	3.649	32718093	20863393	4.649	4.671
2)	SA Decachlor...	9.933	8.641	19977638	19195926	4.520	4.917
Target Compounds							
3)	L1 AR-1016-1	5.487	4.727	12394002	9645398	50.378	50.697
4)	L1 AR-1016-2	5.509	4.747	20092235	13100544	50.601	48.366
5)	L1 AR-1016-3	5.572	4.921	10864128	7492939	48.718	51.057m
6)	L1 AR-1016-4	5.666	4.964	9739118	6295497	51.564	53.616
7)	L1 AR-1016-5	5.959	5.175	9526800	8111827	53.462	53.507m
31)	L7 AR-1260-1	7.071	6.206	14697577	15721569	48.664m	53.302m
32)	L7 AR-1260-2	7.325	6.392	17512148	21836937	45.034m	51.096m
33)	L7 AR-1260-3	7.682	6.546	14586032	17697890	44.584m	50.400
34)	L7 AR-1260-4	7.904	7.017	13400012	14397971	44.028m	47.514
35)	L7 AR-1260-5	8.211	7.258	33518372	38009083	46.996	46.261

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:53
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:09:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:09:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114876.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 11:11
 Operator : YP/AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 11:57:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 11:57:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	367.5E6	224.3E6	50.000	50.000
2) SA Decachlor...	9.932	8.641	223.4E6	199.5E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.549	3.860	42062015	30899570	500.000	500.000
9) L2 AR-1221-2	4.632	3.945	30091751	22818522	500.000	500.000
10) L2 AR-1221-3	4.706	4.021	96667420	70128648	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114876.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:11
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:57:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:57:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
 Data File : PO114877.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 11:30
 Operator : YP/AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:15:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 11:57:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.649	373.4E6	232.5E6	50.000	50.000
2) SA Decachlor...	9.931	8.641	232.4E6	203.6E6	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.706	4.020	75007315	55746703	500.000	500.000
12) L3 AR-1232-2	5.224	4.747	36283683	56047490	500.000	500.000
13) L3 AR-1232-3	5.508	4.921	79337557	29827092	500.000	500.000
14) L3 AR-1232-4	5.666	5.005	36897763	25444495	500.000	500.000
15) L3 AR-1232-5	5.757	5.176	25790349	28293807	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114877.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:30
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:15:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:57:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114878.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 11:48
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:52:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 12:48:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	734.8E6	480.0E6	98.758	98.347
2) SA Decachlor...	9.932	8.642	473.0E6	409.3E6	98.660	96.347
Target Compounds						
16) L4 AR-1242-1	5.487	4.728	180.4E6	138.3E6	974.605	962.566
17) L4 AR-1242-2	5.509	4.747	290.6E6	199.9E6	974.478	966.127
18) L4 AR-1242-3	5.570	4.921	164.4E6	108.2E6	971.984	963.489
19) L4 AR-1242-4	5.667	5.005	136.8E6	101.9E6	966.255	961.849
20) L4 AR-1242-5	6.392	5.526	138.4E6	136.0E6	976.570	965.656



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:48
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:52:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114879.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 12:06
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:55:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 12:48:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	558.4E6	351.3E6	75.032	72.950
2) SA Decachlor...	9.930	8.641	360.4E6	322.6E6	75.113	75.624
Target Compounds						
16) L4 AR-1242-1	5.487	4.728	138.4E6	108.3E6	748.334	752.466
17) L4 AR-1242-2	5.508	4.747	221.7E6	153.7E6	745.677	745.260
18) L4 AR-1242-3	5.570	4.921	125.8E6	84172141	745.539	749.608
19) L4 AR-1242-4	5.666	5.006	105.5E6	78976449	746.496	746.880
20) L4 AR-1242-5	6.391	5.525	106.4E6	105.1E6	750.553	747.461



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114879.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:06
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:55:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114880.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 12:25
 Operator : YP/AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:49:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 12:48:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.348	3.649	376.6E6	248.1E6	50.000	50.000
2)	SA Decachlor...	9.932	8.642	242.9E6	220.2E6	50.000	50.000
Target Compounds							
16)	L4 AR-1242-1	5.488	4.728	94898416	74554194	500.000	500.000
17)	L4 AR-1242-2	5.509	4.747	152.9E6	107.0E6	500.000	500.000
18)	L4 AR-1242-3	5.571	4.922	86964841	58209186	500.000	500.000
19)	L4 AR-1242-4	5.667	5.006	73201423	55002215	500.000	500.000
20)	L4 AR-1242-5	6.393	5.526	72501094	72826705	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:25
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:49:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114881.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 12:42
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 12:59:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 12:48:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.347	3.649	181.6E6	113.2E6	24.544	23.863
2)	SA Decachlor...	9.930	8.642	118.7E6	108.5E6	24.798	25.317
Target Compounds							
16)	L4 AR-1242-1	5.485	4.728	48045131	37219390	257.320	256.344
17)	L4 AR-1242-2	5.507	4.747	75500933	52970083	252.915	255.070
18)	L4 AR-1242-3	5.569	4.921	43602684	28371324	256.308	251.994
19)	L4 AR-1242-4	5.665	5.006	34915069	27309187	247.826	256.146
20)	L4 AR-1242-5	6.391	5.526	34596493	37575425	245.544	262.747



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:42
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:59:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
 Data File : PO114882.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 13:00
 Operator : YP/AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 13:15:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 13:15:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	33472896	22185295	4.613	4.739
2) SA Decachlor...	9.932	8.641	20800693	21904493	4.464	5.090
Target Compounds						
16) L4 AR-1242-1	5.487	4.729	9217524	7516814	49.493	51.407
17) L4 AR-1242-2	5.509	4.747	14989568	10253143	50.170	49.497
18) L4 AR-1242-3	5.571	4.922	8818491	5476880	51.459	48.911
19) L4 AR-1242-4	5.667	5.007	7618827	5459668	53.210	50.962
20) L4 AR-1242-5	6.392	5.526	7059809	7127481	50.085	49.823m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114882.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:00
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 13:15:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 13:15:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114883.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 13:19
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:15:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:11:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	731.7E6	475.9E6	98.575	98.613
2) SA Decachlor...	9.931	8.641	459.1E6	407.6E6	98.355	97.441
Target Compounds						
21) L5 AR-1248-1	5.486	4.727	130.1E6	101.0E6	970.963	951.788
22) L5 AR-1248-2	5.757	4.963	182.9E6	136.5E6	970.365	957.940
23) L5 AR-1248-3	5.957	5.005	203.4E6	143.9E6	974.566	961.296
24) L5 AR-1248-4	6.354	5.175	240.2E6	171.3E6	968.415	963.693
25) L5 AR-1248-5	6.391	5.567	225.3E6	174.6E6	969.541	965.748



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114883.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:19
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:15:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114884.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 13:37
 Operator : YP/AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:18:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:11:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.650	558.2E6	362.2E6	75.139	75.043
2) SA Decachlor...	9.934	8.642	348.6E6	312.5E6	74.784	74.797
Target Compounds						
21) L5 AR-1248-1	5.487	4.728	101.4E6	79219349	754.563	747.551
22) L5 AR-1248-2	5.759	4.965	140.8E6	106.9E6	747.878	750.206
23) L5 AR-1248-3	5.958	5.007	154.7E6	112.4E6	744.299	750.689
24) L5 AR-1248-4	6.355	5.177	184.7E6	133.3E6	746.353	749.852
25) L5 AR-1248-5	6.393	5.568	173.1E6	135.2E6	746.437	748.274



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114884.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:37
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:18:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114885.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 13:56
 Operator : YP/AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:12:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:11:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.650	376.4E6	244.6E6	50.000	50.000
2) SA Decachlor...	9.932	8.642	237.2E6	214.5E6	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.487	4.728	68954500	55631193	500.000	500.000
22) L5 AR-1248-2	5.759	4.964	97032346	74265259	500.000	500.000
23) L5 AR-1248-3	5.959	5.006	107.0E6	77737125	500.000	500.000
24) L5 AR-1248-4	6.354	5.177	127.9E6	92085877	500.000	500.000
25) L5 AR-1248-5	6.393	5.568	119.7E6	93519021	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:56
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:12:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
 Data File : PO114886.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 14:14
 Operator : YP/AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:29:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:11:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	185.1E6	120.3E6	24.933	24.939
2) SA Decachlor...	9.931	8.642	116.5E6	110.1E6	25.002	26.000
Target Compounds						
21) L5 AR-1248-1	5.487	4.728	35435549	28273297	260.067	262.392
22) L5 AR-1248-2	5.758	4.964	49098195	38173242	258.061	263.101
23) L5 AR-1248-3	5.958	5.006	53231777	39587309	254.519	260.615
24) L5 AR-1248-4	6.354	5.176	63989316	46878477	256.419	260.212
25) L5 AR-1248-5	6.393	5.567	60028950	47410600	256.634	259.234



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114886.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:14
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:29:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114887.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 14:33
 Operator : YP/AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 14:49:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 04 14:48:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	36092002	23635170	4.889	4.920
2) SA Decachlor...	9.931	8.642	21056806	20723944	4.607	4.915
Target Compounds						
21) L5 AR-1248-1	5.487	4.727	7710330	5776990	55.135	52.850
22) L5 AR-1248-2	5.758	4.963	10567680	8207209	54.339	55.119
23) L5 AR-1248-3	5.958	5.005	10879375	8472020	51.601	54.515
24) L5 AR-1248-4	6.355	5.176	13238663	9782119	52.411	53.381
25) L5 AR-1248-5	6.393	5.567	12160694	9504633	51.579	51.564



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:33
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:49:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 14:51
 Operator : YP/AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:36:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.650	741.1E6	479.1E6	94.981	94.986
2) SA Decachlor...	9.932	8.642	479.2E6	427.3E6	94.662	92.802
Target Compounds						
26) L6 AR-1254-1	6.331	5.527	240.6E6	275.4E6	929.768	917.386
27) L6 AR-1254-2	6.547	5.675	346.0E6	245.3E6	931.192	915.477
28) L6 AR-1254-3	6.907	6.076	372.9E6	398.2E6	939.256	924.961
29) L6 AR-1254-4	7.188	6.304	315.5E6	284.8E6	936.760	933.639
30) L6 AR-1254-5	7.604	6.721	295.4E6	385.7E6	946.239	941.905



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:51
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114889.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 15:09
 Operator : YP/AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:36:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.650	564.5E6	366.5E6	72.353	72.665
2) SA Decachlor...	9.933	8.642	367.3E6	327.5E6	72.554	71.130
Target Compounds						
26) L6 AR-1254-1	6.331	5.527	185.6E6	213.2E6	717.106	710.020
27) L6 AR-1254-2	6.547	5.675	265.1E6	190.0E6	713.548	708.846
28) L6 AR-1254-3	6.907	6.077	285.3E6	308.0E6	718.711	715.281
29) L6 AR-1254-4	7.189	6.304	241.1E6	221.2E6	715.674	725.087
30) L6 AR-1254-5	7.604	6.721	225.5E6	294.2E6	722.455	718.550



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 15:09
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114890.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 15:28
 Operator : YP/AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:36:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	390.1E6	252.2E6	50.000	50.000
2) SA Decachlor...	9.932	8.642	253.1E6	230.2E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.331	5.527	129.4E6	150.1E6	500.000	500.000
27) L6 AR-1254-2	6.547	5.674	185.8E6	134.0E6	500.000	500.000
28) L6 AR-1254-3	6.907	6.076	198.5E6	215.3E6	500.000	500.000
29) L6 AR-1254-4	7.189	6.304	168.4E6	152.5E6	500.000	500.000
30) L6 AR-1254-5	7.603	6.720	156.1E6	204.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 15:28
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 16:04
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:36:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	187.0E6	121.9E6	23.973	24.169
2) SA Decachlor...	9.932	8.642	120.8E6	113.3E6	23.861	24.611
Target Compounds						
26) L6 AR-1254-1	6.331	5.527	64268788	74980905	248.367	249.734
27) L6 AR-1254-2	6.548	5.674	92224254	67079510	248.188	250.312
28) L6 AR-1254-3	6.907	6.076	96073294	106.6E6	241.992	247.668
29) L6 AR-1254-4	7.189	6.304	81726238	75488880	242.626	247.482
30) L6 AR-1254-5	7.604	6.720	75118696	99522496	240.645	243.035


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 16:04
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114892.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 16:23
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:37:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:33:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	37040207	24215180	4.747	4.801
2) SA Decachlor...	9.931	8.642	23309156	21360673	4.605m	4.639
Target Compounds						
26) L6 AR-1254-1	6.330	5.526	13164962	17052798	50.876m	56.797
27) L6 AR-1254-2	6.547	5.674	20609241	15656147	55.462m	58.422
28) L6 AR-1254-3	6.907	6.076	19483947	22264737	49.077	51.713
29) L6 AR-1254-4	7.188	6.303	17143593	15015311	50.895m	49.226
30) L6 AR-1254-5	7.604	6.720	15028663	20100339	48.145	49.085

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 16:23
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_0
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:37:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114893.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 17:00
 Operator : YP/AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:08:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:07:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.368	3.649	362.5E6	234.2E6	50.000	50.000
2) SA Decachlor...	9.952	8.641	223.0E6	207.7E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	7.703	6.759	226.7E6	247.1E6	500.000	500.000
37) L8 AR-1262-2	8.231	7.258	353.6E6	397.5E6	500.000	500.000
38) L8 AR-1262-3	8.534	7.542	225.7E6	164.5E6	500.000	500.000
39) L8 AR-1262-4	8.616	7.605	168.1E6	269.1E6	500.000	500.000
40) L8 AR-1262-5	9.236	8.097	120.9E6	119.0E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:00
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:08:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:07:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 17:18
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.349	3.649	747.6E6	479.2E6	97.655	97.469
2) SA Decachlor...	9.931	8.642	828.8E6	735.0E6	97.454	95.377
Target Compounds						
41) L9 AR-1268-1	8.511	7.542	876.3E6	948.8E6	981.231	966.485
42) L9 AR-1268-2	8.599	7.606	748.3E6	784.3E6	977.544	966.821
43) L9 AR-1268-3	8.820	7.814	663.3E6	672.6E6	975.632	966.815
44) L9 AR-1268-4	9.216	8.097	274.2E6	270.9E6	971.116	967.067
45) L9 AR-1268-5	9.609	8.387	1966.6E6	1768.5E6	990.357	972.552



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:18
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114895.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 17:36
 Operator : YP/AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	572.8E6	367.7E6	74.818	74.796
2) SA Decachlor...	9.932	8.641	629.7E6	557.8E6	74.043	72.388
Target Compounds						
41) L9 AR-1268-1	8.510	7.542	664.3E6	724.0E6	743.849	737.521
42) L9 AR-1268-2	8.599	7.606	567.3E6	597.5E6	740.998	736.564
43) L9 AR-1268-3	8.819	7.814	504.1E6	511.1E6	741.416	734.677
44) L9 AR-1268-4	9.215	8.096	208.1E6	206.6E6	737.304	737.457
45) L9 AR-1268-5	9.609	8.387	1484.6E6	1339.5E6	747.625	736.619



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:36
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 17:55
 Operator : YP/AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	382.8E6	245.8E6	50.000	50.000
2) SA Decachlor...	9.931	8.641	425.2E6	385.3E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.510	7.542	446.5E6	490.9E6	500.000	500.000
42) L9 AR-1268-2	8.598	7.606	382.8E6	405.6E6	500.000	500.000
43) L9 AR-1268-3	8.820	7.814	339.9E6	347.9E6	500.000	500.000
44) L9 AR-1268-4	9.215	8.096	141.2E6	140.0E6	500.000	500.000
45) L9 AR-1268-5	9.609	8.387	992.9E6	909.2E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:55
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114897.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 18:13
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	189.9E6	122.1E6	24.810	24.834
2) SA Decachlor...	9.932	8.641	210.5E6	196.5E6	24.748	25.504
Target Compounds						
41) L9 AR-1268-1	8.511	7.541	221.9E6	248.4E6	248.457	253.017m
42) L9 AR-1268-2	8.599	7.605	190.0E6	204.3E6	248.168	251.781
43) L9 AR-1268-3	8.820	7.813	167.9E6	174.1E6	246.983	250.180
44) L9 AR-1268-4	9.216	8.096	70540429	70593070	249.869	252.040
45) L9 AR-1268-5	9.610	8.387	489.2E6	458.0E6	246.358	251.855

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 18:13
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114898.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 18:31
 Operator : YP/AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:23:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:21:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	35607202	23620780	4.651	4.805
2) SA Decachlor...	9.931	8.640	38610556	41609899	4.540	5.400
Target Compounds						
41) L9 AR-1268-1	8.511	7.542	40675439	49554127	45.547	50.476
42) L9 AR-1268-2	8.599	7.606	35140218	38446458	45.903	47.393
43) L9 AR-1268-3	8.820	7.814	30894342	31638164	45.442	45.475
44) L9 AR-1268-4	9.214	8.096	14340449	13380470	50.797	47.773
45) L9 AR-1268-5	9.609	8.387	90370817	85884170	45.510	47.229



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 18:31
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 19:08
 Operator : YP/AJ
 Sample : PO110425ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 02:03:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 02:02:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.348	3.649	350.6E6	216.0E6	49.817	48.358
2)	SA Decachlor...	9.932	8.640	217.2E6	202.0E6	49.145	51.742
Target Compounds							
3)	L1 AR-1016-1	5.487	4.728	119.9E6	92647170	487.281	486.962
4)	L1 AR-1016-2	5.508	4.747	193.0E6	132.4E6	485.938	488.627
5)	L1 AR-1016-3	5.570	4.921	110.7E6	72980433	496.242	493.200
6)	L1 AR-1016-4	5.666	4.963	92089690	57444659	487.569	489.233
7)	L1 AR-1016-5	5.957	5.176	84866211	74982175	476.249	490.999
31)	L7 AR-1260-1	7.071	6.205	144.7E6	141.9E6	473.335	479.180
32)	L7 AR-1260-2	7.325	6.393	200.2E6	209.0E6	505.658	486.103
33)	L7 AR-1260-3	7.682	6.546	167.5E6	170.3E6	504.732	484.991
34)	L7 AR-1260-4	7.904	7.017	152.7E6	150.0E6	500.629	495.013
35)	L7 AR-1260-5	8.210	7.257	350.1E6	404.4E6	490.903	492.224



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 19:08
Operator : YP/AJ
Sample : PO110425ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
ICVPO110425

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:02:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 20:21
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425AR1242

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/05/2025
 Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:57:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:57:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	369.9E6	227.0E6	50.980	48.485
2) SA Decachlor...	9.932	8.639	237.1E6	216.0E6	50.883	50.180
Target Compounds						
16) L4 AR-1242-1	5.486	4.727	91685645	72678150	492.296m	497.041
17) L4 AR-1242-2	5.508	4.747	146.5E6	102.4E6	490.193m	494.324
18) L4 AR-1242-3	5.570	4.921	84401159	56402697	492.512m	503.697
19) L4 AR-1242-4	5.666	5.005	70483519	53471210	492.260m	499.119
20) L4 AR-1242-5	6.393	5.525	75170815	70192007	533.288	491.136

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 20:21
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1242

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:57:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:57:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 20:58
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:49:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:49:02 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	364.4E6	233.5E6	49.370	48.616
2) SA Decachlor...	9.931	8.640	231.0E6	213.4E6	50.545	50.607
Target Compounds						
21) L5 AR-1248-1	5.487	4.727	67941811	53066852	485.834	485.472
22) L5 AR-1248-2	5.758	4.963	96189175	72201389	494.605	484.897
23) L5 AR-1248-3	5.958	5.005	104.8E6	75367544	497.247	484.966
24) L5 AR-1248-4	6.354	5.175	125.7E6	89630325	497.795	489.109
25) L5 AR-1248-5	6.393	5.567	118.0E6	89347145	500.480	484.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 20:58
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:49:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114902.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 21:35
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 05:47:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 05:47:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	365.7E6	233.3E6	48.598	47.717
2) SA Decachlor...	9.931	8.640	237.7E6	218.5E6	49.015	49.554
Target Compounds						
26) L6 AR-1254-1	6.331	5.526	122.7E6	141.2E6	484.213	470.374
27) L6 AR-1254-2	6.546	5.673	176.1E6	124.6E6	475.284	462.124
28) L6 AR-1254-3	6.906	6.075	193.7E6	200.1E6	503.340	473.848
29) L6 AR-1254-4	7.188	6.302	159.9E6	143.9E6	486.428	484.000
30) L6 AR-1254-5	7.603	6.719	147.9E6	190.6E6	490.023	479.436


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 21:35
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:47:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:47:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
 Data File : PO114903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 22:12
 Operator : YP/AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO110425AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 01:39:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 01:38:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.647	376.8E6	240.7E6	50.252	49.688
2) SA Decachlor...	9.929	8.640	416.0E6	381.0E6	50.328	49.259
Target Compounds						
41) L9 AR-1268-1	8.509	7.541	437.5E6	476.9E6	502.163	488.581
42) L9 AR-1268-2	8.599	7.605	374.7E6	393.6E6	501.949	494.742
43) L9 AR-1268-3	8.819	7.814	331.8E6	336.4E6	502.049	497.748
44) L9 AR-1268-4	9.214	8.096	138.2E6	136.4E6	492.596	495.687
45) L9 AR-1268-5	9.608	8.385	968.0E6	889.7E6	499.171	498.578



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 22:12
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:39:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:38:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Contract: REMI02

SDG NO.: Q3552

Calibration Date(s): **10/21/2025** **10/21/2025**

Calibration Times: 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		RT 1000 =	<u>PP075888.D</u>	RT 750 =	<u>PP075889.D</u>
RT 500 =	PP075890.D	RT 250 =	PP075891.D	RT 050 =	PP075892.D

[illegible]

[illegible]

Calibration Times: 09:06 17:01

LAB FILE ID: RT 1000 = PP075888.D RT 750 = PP075889.D
RT 500 = PP075890.D RT 250 = PP075891.D RT 050 = PP075892.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.79	8.78	8.79	8.79	8.79	8.79	8.69	8.89
Tetrachloro-m-xylene	3.78	3.78	3.79	3.79	3.79	3.78	3.68	3.88



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: REMI02
SDG NO.: Q3552

Calibration Date(s): 10/21/2025 10/21/2025
Calibration Times: 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP075888.D</u>		CF 750 = <u>PP075889.D</u>				
CF 500 = <u>PP075890.D</u>		CF 250 = <u>PP075891.D</u>		CF 050 = <u>PP075892.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	53893845	57848439	60228768	63056856	60602640	59126110	6
Aroclor-1016-2	(2)	81387326	86593457	91038396	94931280	88270120	88444116	6
Aroclor-1016-3	(3)	50025937	53493744	56764402	59544472	56926600	55351031	7
Aroclor-1016-4	(4)	41815325	44617539	46982858	48544784	45240520	45440205	6
Aroclor-1016-5	(5)	39075321	41377448	43150568	44782124	48642160	43405524	8
Aroclor-1260-1	(1)	64655691	67908073	70553690	74100180	66565440	68756615	5
Aroclor-1260-2	(2)	80744063	85234851	88594752	93987136	99325680	89577296	8
Aroclor-1260-3	(3)	64964687	67552583	69003280	72715708	72196540	69286560	5
Aroclor-1260-4	(4)	65448766	70711301	72542976	74553116	59600740	68571380	9
Aroclor-1260-5	(5)	139384380	146230631	152452738	155039368	146650940	147951611	4
Decachlorobiphenyl		1012132950	1071302787	1124785220	1137889120	1036523400	1076526695	5
Tetrachloro-m-xylene		1370219590	1453116893	1490571040	1483020960	1348537200	1429093137	5
Aroclor-1242-1	(1)	44019933	46315467	47674416	50255084	51268840	47906748	6
Aroclor-1242-2	(2)	64942800	67715737	70979556	73767248	76553360	70791740	7
Aroclor-1242-3	(3)	40140584	42044764	44589646	47135032	54298900	45641785	12
Aroclor-1242-4	(4)	33543618	35162207	36420600	36967860	39702420	36359341	6
Aroclor-1242-5	(5)	36405140	38662289	41888622	39296052	48467620	40943945	11
Decachlorobiphenyl		1136280900	1209279307	1231278960	1289202560	1357807000	1244769745	7
Tetrachloro-m-xylene		1466601570	1512648573	1545825660	1595207720	1573043800	1538665465	3
Aroclor-1248-1	(1)	33471805	35785521	36963520	39121904	34010240	35870598	6
Aroclor-1248-2	(2)	44816839	47584043	49402910	52441868	49476340	48744400	6
Aroclor-1248-3	(3)	49477903	52446567	61105936	57335280	50958020	54264741	9
Aroclor-1248-4	(4)	60435029	64167197	67505182	72174088	67539280	66364155	7
Aroclor-1248-5	(5)	58134534	61713151	65347636	71297512	67810300	64860627	8
Decachlorobiphenyl		1140637530	1190681093	1243623480	1306886720	1250686400	1226503045	5
Tetrachloro-m-xylene		1468701040	1529744213	1558324940	1626496320	1450854600	1526824223	5
Aroclor-1254-1	(1)	66261960	71898345	78270020	82464920	100572060	79893461	16
Aroclor-1254-2	(2)	87851253	92570580	96669370	102809560	120020480	99984249	12
Aroclor-1254-3	(3)	92035771	96900691	99787992	105988040	125061300	103954759	12
Aroclor-1254-4	(4)	73136868	76689867	79390286	83967352	104884540	83613783	15
Aroclor-1254-5	(5)	85440842	89720104	92252192	95013656	95427040	91570767	5
Decachlorobiphenyl		1194970230	1245235800	1296058720	1332579960	1514859800	1316740902	9
Tetrachloro-m-xylene		1499217980	1559170693	1583264260	1614158480	1720822800	1595326843	5
Aroclor-1268-1	(1)	186733090	195066952	205606148	210022232	235253460	206536376	9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	172486766	179135464	188184104	190078376	211424800	188261902	8
Aroclor-1268-3	(3)	145245892	150281380	156894280	158090072	165408840	155184093	5
Aroclor-1268-4	(4)	71036137	74805685	76392484	76627212	94373740	78647052	12
Aroclor-1268-5	(5)	446420076	458917541	475352492	479939416	541304600	480386825	8
Decachlorobiphenyl		1988394870	2075434813	2173906840	2211711440	2356033600	2161096313	6
Tetrachloro-m-xylene		1515083040	1549265533	1603737900	1591085800	1659449000	1583724255	3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: REMI02
SDG NO.: Q3552

Calibration Date(s): 10/21/2025 10/21/2025
Calibration Times: 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP075888.D</u>		CF 750 = <u>PP075889.D</u>			
CF 500 = <u>PP075890.D</u>		CF 250 = <u>PP075891.D</u>		CF 050 = <u>PP075892.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	708597432	755130533	750168220	768814916	718149060	740172032 3
Aroclor-1016-2	(2)	344688013	362255933	349033076	368003724	285681000	341932349 10
Aroclor-1016-3	(3)	184756021	198805969	201689770	204443556	237017160	205342495 9
Aroclor-1016-4	(4)	185429480	195612065	193130110	203643812	156612400	186885573 10
Aroclor-1016-5	(5)	202087874	223186380	221264342	249936892	243439240	227982946 8
Aroclor-1260-1	(1)	519115007	547143451	551523748	550034516	570658900	547695124 3
Aroclor-1260-2	(2)	816933203	847057133	836099344	841169992	727408040	813733542 6
Aroclor-1260-3	(3)	571186815	591919587	603061842	595667512	451948620	562756875 11
Aroclor-1260-4	(4)	562542796	557220415	556309062	533352684	387232680	519331527 14
Aroclor-1260-5	(5)	1443277342	1495374137	1448146362	1379476524	1011499060	1355554685 15
Decachlorobiphenyl		7970697090	8448130240	8416417320	8375691800	7903993200	8222985930 3
Tetrachloro-m-xylene		6687145430	7018651333	6804923780	6758234840	5822487600	6618288597 7
Aroclor-1242-1	(1)	571147464	588341980	597723718	569760388	546790280	574752766 3
Aroclor-1242-2	(2)	271622570	270496133	289123856	291348184	264965740	277511297 4
Aroclor-1242-3	(3)	148417684	150685240	162169734	157714788	168071120	157411713 5
Aroclor-1242-4	(4)	195275059	191450107	210824584	191270784	136740280	185112163 15
Aroclor-1242-5	(5)	229588702	236959084	244202752	234962408	205474700	230237529 6
Decachlorobiphenyl		8905080650	8993413907	9094156860	9113318400	9309338200	9083061603 2
Tetrachloro-m-xylene		7556133750	7417192413	7366207320	7201170040	6123421000	7132824905 8
Aroclor-1248-1	(1)	362937087	375551145	389160858	390611340	359825360	375617158 4
Aroclor-1248-2	(2)	226906631	241863020	243838414	257801160	259130280	245907901 5
Aroclor-1248-3	(3)	258884180	267447340	286041186	300772772	321318420	286892780 9
Aroclor-1248-4	(4)	264944699	288582168	281543548	284810560	293675780	282711351 4
Aroclor-1248-5	(5)	531974990	530976477	538260328	580030476	417051840	519658822 12
Decachlorobiphenyl		8995204870	9175718960	9269974040	9050625160	7344080200	8767120646 9
Tetrachloro-m-xylene		7343926730	7496283907	7432157100	7125361600	5416154400	6962776747 13
Aroclor-1254-1	(1)	570889099	597066183	605466528	586480320	460097860	563999998 11
Aroclor-1254-2	(2)	484331632	496825463	519240264	512723972	637282480	530080762 12
Aroclor-1254-3	(3)	866771846	893592493	892235862	873540968	729313040	851090842 8
Aroclor-1254-4	(4)	656603428	687178148	673132272	636725912	495267720	629781496 12
Aroclor-1254-5	(5)	714919063	793859644	718111150	679323440	693198580	719882375 6
Decachlorobiphenyl		9294247170	9652211800	9737198260	9430246600	10010120400	9624804846 3
Tetrachloro-m-xylene		7467434200	7759548133	7675903540	7263970760	7005463400	7434464007 4
Aroclor-1268-1	(1)	1831610079	1813456776	1858863308	1732000476	1817331660	1810652460 3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1800363506	1772458256	1795415328	1662166904	1645167940	1735114387	4
Aroclor-1268-3	(3)	1369797429	1358701644	1402746042	1358715304	1345500120	1367092108	2
Aroclor-1268-4	(4)	542764588	536784571	540251440	501746924	465672060	517443917	6
Aroclor-1268-5	(5)	3855476951	3900359425	4013489614	3707995788	3381331260	3771730608	6
Decachlorobiphenyl		16718332370	16659738347	17037476180	16197346520	15850998600	16492778403	3
Tetrachloro-m-xylene		7499400880	7531972947	7553431460	7073074800	6855756800	7302727377	4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: REMI02

Lab Code: ACE SDG NO.: Q3552

Instrument ID: ECD_P Date(s) Analyzed: 10/21/2025 10/21/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.84	4.74	4.94	20176600
		2	4.93	4.83	5.03	15061400
		3	5.01	4.91	5.11	46915600
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.01	4.91	5.11	39314800
		2	5.53	5.43	5.63	20627600
		3	5.82	5.72	5.92	38639600
		4	5.98	5.88	6.08	20172600
		5	6.07	5.97	6.17	15662900
Aroclor-1262	500	1	8.23	8.13	8.33	84026000
		2	8.56	8.46	8.66	163669000
		3	8.88	8.78	8.98	121787000
		4	8.97	8.87	9.07	92111600
		5	9.64	9.54	9.74	67824600



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
Instrument ID: ECD_P **Date(s) Analyzed:** 10/21/2025 10/21/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.99	3.89	4.09	90009000
		2	4.08	3.98	4.18	63177800
		3	4.15	4.05	4.25	212036000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.15	4.05	4.25	168044000
		2	4.88	4.78	4.98	297516000
		3	5.06	4.96	5.16	81766800
		4	5.14	5.04	5.24	84943400
		5	5.31	5.21	5.41	100651000
Aroclor-1262	500	1	6.89	6.79	6.99	986952000
		2	7.15	7.05	7.25	703208000
		3	7.67	7.57	7.77	606960000
		4	7.74	7.64	7.84	1101340000
		5	8.23	8.13	8.33	461492000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:06
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:20:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	137.0E6	668.7E6	95.793	99.127
2) SA Decachlor...	10.422	8.784	101.2E6	797.1E6	94.728	97.280
Target Compounds						
3) L1 AR-1016-1	5.799	4.880	53893845	708.6E6	944.490	971.503
4) L1 AR-1016-2	5.820	4.938	81387326	344.7E6	944.028	993.737
5) L1 AR-1016-3	5.883	5.058	50025937	184.8E6	936.900	956.181
6) L1 AR-1016-4	5.980	5.099	41815325	185.4E6	941.806	979.658
7) L1 AR-1016-5	6.272	5.313	39075321	202.1E6	950.438	954.703
31) L7 AR-1260-1	7.391	6.341	64655691	519.1E6	956.379	969.730
32) L7 AR-1260-2	7.643	6.527	80744063	816.9E6	953.639	988.405
33) L7 AR-1260-3	8.001	6.681	64964687	571.2E6	969.854	972.855
34) L7 AR-1260-4	8.229	7.150	65448766	562.5E6	948.590	1005.572
35) L7 AR-1260-5	8.555	7.390	139.4E6	1443.3E6	955.220	998.316



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:06
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075889.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:22
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:23:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.781	109.0E6	526.4E6	75.790	76.994
2) SA Decachlor...	10.419	8.782	80347709	633.6E6	75.133	76.538
Target Compounds						
3) L1 AR-1016-1	5.795	4.877	43386329	566.3E6	756.866	767.445
4) L1 AR-1016-2	5.817	4.935	64945093	271.7E6	752.204	771.869
5) L1 AR-1016-3	5.879	5.055	40120308	149.1E6	750.922	764.309
6) L1 AR-1016-4	5.977	5.096	33463154	146.7E6	752.456	766.543
7) L1 AR-1016-5	6.270	5.309	31033086	167.4E6	753.210	776.704
31) L7 AR-1260-1	7.387	6.338	50931055	410.4E6	752.240	760.963
32) L7 AR-1260-2	7.640	6.525	63926138	635.3E6	753.332	762.324
33) L7 AR-1260-3	7.999	6.678	50664437	443.9E6	754.232	754.073
34) L7 AR-1260-4	8.226	7.147	53033476	417.9E6	762.329	748.026
35) L7 AR-1260-5	8.552	7.387	109.7E6	1121.5E6	751.069	766.981



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:22
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:23:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075890.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:38
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:17:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.781	74528552	340.2E6	50.000	50.000
2) SA Decachlor...	10.419	8.782	56239261	420.8E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.795	4.878	30114384	375.1E6	500.000	500.000
4) L1 AR-1016-2	5.817	4.936	45519198	174.5E6	500.000	500.000
5) L1 AR-1016-3	5.880	5.054	28382201	100.8E6	500.000	500.000
6) L1 AR-1016-4	5.977	5.096	23491429	96565055	500.000	500.000
7) L1 AR-1016-5	6.269	5.310	21575284	110.6E6	500.000	500.000
31) L7 AR-1260-1	7.387	6.338	35276845	275.8E6	500.000	500.000
32) L7 AR-1260-2	7.640	6.525	44297376	418.0E6	500.000	500.000
33) L7 AR-1260-3	7.999	6.678	34501640	301.5E6	500.000	500.000
34) L7 AR-1260-4	8.227	7.147	36271488	278.2E6	500.000	500.000
35) L7 AR-1260-5	8.552	7.387	76226369	724.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:38
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:17:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:54
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:27:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.784	37075524	169.0E6	25.583	24.784
2) SA Decachlor...	10.422	8.784	28447228	209.4E6	26.182	25.220
Target Compounds						
3) L1 AR-1016-1	5.798	4.879	15764214	192.2E6	268.295	257.757
4) L1 AR-1016-2	5.819	4.938	23732820	92000931	268.205	258.433
5) L1 AR-1016-3	5.882	5.057	14886118	51110889	270.868	258.889
6) L1 AR-1016-4	5.979	5.098	12136196	50910953	266.787	261.815
7) L1 AR-1016-5	6.272	5.311	11195531	62484223	265.950	278.799
31) L7 AR-1260-1	7.390	6.340	18525045	137.5E6	267.300	253.727
32) L7 AR-1260-2	7.643	6.526	23496784	210.3E6	269.643	251.752
33) L7 AR-1260-3	8.001	6.680	18178927	148.9E6	265.157	252.205
34) L7 AR-1260-4	8.229	7.150	18638279	133.3E6	263.200	241.399
35) L7 AR-1260-5	8.554	7.389	38759842	344.9E6	261.402	239.232



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:54
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:27:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075892.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:11
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:33:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.781	6742686	29112438	4.718	4.278m
2) SA Decachlor...	10.422	8.783	5182617	39519966	4.814	4.806
Target Compounds						
3) L1 AR-1016-1	5.798	4.880	3030132	35907453	51.249	48.512
4) L1 AR-1016-2	5.819	4.943	4413506	14284050	49.902	41.774
5) L1 AR-1016-3	5.880	5.052	2846330	11850858	50.609m	58.316m
6) L1 AR-1016-4	5.978	5.094	2262026	7830620	48.653m	41.996m
7) L1 AR-1016-5	6.272	5.308	2432108	12171962	56.032	53.433m
31) L7 AR-1260-1	7.390	6.339	3328272	28532945	48.407	52.096
32) L7 AR-1260-2	7.643	6.528	4966284	36370402	55.441	44.696
33) L7 AR-1260-3	8.001	6.681	3609827	22597431	52.100	40.155
34) L7 AR-1260-4	8.229	7.150	2980037	19361634	43.459	37.282
35) L7 AR-1260-5	8.555	7.388	7332547	50574953	49.560	37.309

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:33:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075893.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:27
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:37:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:37:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.643	3.782	75268417	347.9E6	50.000	50.000
2) SA Decachlor...	10.419	8.781	55765598	413.6E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.843	3.993	10088314	45004530	500.000	500.000
9) L2 AR-1221-2	4.929	4.076	7530715	31588869	500.000	500.000
10) L2 AR-1221-3	5.005	4.154	23457783	106.0E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:37:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:43
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:45:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:45:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.783	76762692	367.2E6	50.000	50.000
2) SA Decachlor...	10.418	8.783	57573853	421.9E6	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	5.006	4.154	19657396	84022049	500.000	500.000
12) L3 AR-1232-2	5.532	4.877	10313847	148.8E6	500.000	500.000
13) L3 AR-1232-3	5.818	5.056	19319843	40883369	500.000	500.000
14) L3 AR-1232-4	5.978	5.136	10086255	42471740	500.000	480.688m
15) L3 AR-1232-5	6.068	5.311	7831441	50325540	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:45:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:45:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075895.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:59
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:57:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.784	146.7E6	755.6E6	97.370	101.273
2)	SA Decachlor...	10.422	8.784	113.6E6	890.5E6	95.988	98.950
Target Compounds							
16)	L4 AR-1242-1	5.799	4.880	44019933	571.1E6	960.145	977.263
17)	L4 AR-1242-2	5.820	4.938	64942800	271.6E6	955.587	968.789
18)	L4 AR-1242-3	5.882	5.058	40140584	148.4E6	947.491	955.722
19)	L4 AR-1242-4	5.980	5.140	33543618	195.3E6	958.879	946.968m
20)	L4 AR-1242-5	6.710	5.660	36405140	229.6E6	929.963	969.155

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:59
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:57:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 11:16
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:03:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.782	113.4E6	556.3E6	75.213	74.705
2) SA Decachlor...	10.420	8.782	90695948	674.5E6	76.069	74.966
Target Compounds						
16) L4 AR-1242-1	5.797	4.877	34736600	441.3E6	755.090	753.335
17) L4 AR-1242-2	5.818	4.936	50786803	202.9E6	748.192	732.177
18) L4 AR-1242-3	5.881	5.056	31533573	113.0E6	746.210	735.014
19) L4 AR-1242-4	5.978	5.138	26371655	143.6E6	752.570	715.813m
20) L4 AR-1242-5	6.708	5.658	28996717	177.7E6	743.785	750.134

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:16
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075897.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 11:48
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:54:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.784	77291283	368.3E6	50.000	50.000
2) SA Decachlor...	10.422	8.785	61563948	454.7E6	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.799	4.880	23837208	298.9E6	498.805	500.000
17) L4 AR-1242-2	5.821	4.938	35489778	144.6E6	500.000	500.000
18) L4 AR-1242-3	5.884	5.059	22294823	81084867	500.000	500.000
19) L4 AR-1242-4	5.981	5.141	18210300	105.4E6	500.000	515.181m
20) L4 AR-1242-5	6.710	5.660	20944311	122.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:48
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:54:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075898.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:04
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:06:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.645	3.783	39880193	180.0E6	26.064	24.377
2)	SA Decachlor...	10.420	8.782	32230064	227.8E6	26.494	25.240
Target Compounds							
16)	L4 AR-1242-1	5.797	4.880	12563771	142.4E6	266.938	244.850
17)	L4 AR-1242-2	5.819	4.936	18441812	72837046	265.919	259.532
18)	L4 AR-1242-3	5.881	5.057	11783758	39428697	271.031	254.795
19)	L4 AR-1242-4	5.979	5.139	9241965	47817696	260.164	239.194m
20)	L4 AR-1242-5	6.707	5.659	9824013	58740602	248.605m	248.450

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:04
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:06:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:21
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:10:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.783	7865219	30617105	5.112	4.301m
2)	SA Decachlor...	10.422	8.783	6789035	46546691	5.454	5.125
Target Compounds							
16)	L4 AR-1242-1	5.798	4.882	2563442	27339514	53.509	47.567
17)	L4 AR-1242-2	5.820	4.939	3827668	13248287	54.069	47.740
18)	L4 AR-1242-3	5.883	5.054	2714945	8403556	59.484	53.386
19)	L4 AR-1242-4	5.979	5.138	1985121	6837014	54.597	36.934 #
20)	L4 AR-1242-5	6.708	5.660	2423381	10273735	59.725m	44.622 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:21
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:10:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:41
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:17:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.780	146.9E6	734.4E6	97.039	99.403
2) SA Decachlor...	10.426	8.783	114.1E6	899.5E6	95.681	98.496
Target Compounds						
21) L5 AR-1248-1	5.799	4.876	33471805	362.9E6	950.427	965.133
22) L5 AR-1248-2	6.071	5.095	44816839	226.9E6	951.326	964.032
23) L5 AR-1248-3	6.274	5.136	49477903	258.9E6	894.849	947.454m
24) L5 AR-1248-4	6.672	5.309	60435029	264.9E6	944.739	969.626
25) L5 AR-1248-5	6.711	5.698	58134534	532.0E6	941.586	994.127

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:41
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:17:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:57
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:20:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.644	3.784	114.7E6	562.2E6	75.534	75.729
2)	SA Decachlor...	10.422	8.784	89301082	688.2E6	74.939	75.236
Target Compounds							
21)	L5 AR-1248-1	5.797	4.879	26839141	281.7E6	758.019	749.338
22)	L5 AR-1248-2	6.068	5.099	35688032	181.4E6	755.016	763.662
23)	L5 AR-1248-3	6.271	5.139	39334925	200.6E6	723.821	728.506m
24)	L5 AR-1248-4	6.670	5.311	48125398	216.4E6	751.539	777.551
25)	L5 AR-1248-5	6.709	5.701	46284863	398.2E6	749.774	746.121

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:57
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:20:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075902.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:13
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:13:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.645	3.784	77916247	371.6E6	50.000	50.000
2)	SA Decachlor...	10.424	8.784	62181174	463.5E6	50.000	50.000
Target Compounds							
21)	L5 AR-1248-1	5.798	4.877	18481760	194.6E6	500.000	500.000
22)	L5 AR-1248-2	6.069	5.098	24701455	121.9E6	500.000	500.000
23)	L5 AR-1248-3	6.273	5.138	30552968	143.0E6	500.000	491.112m
24)	L5 AR-1248-4	6.671	5.312	33752591	140.8E6	500.000	500.000
25)	L5 AR-1248-5	6.710	5.701	32673818	269.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:13
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:13:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:30
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:24:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.783	40662408	178.1E6	26.305	24.238
2) SA Decachlor...	10.423	8.784	32672168	226.3E6	26.770	24.802
Target Compounds						
21) L5 AR-1248-1	5.799	4.878	9780476	97652835	269.170	257.276
22) L5 AR-1248-2	6.070	5.099	13110467	64450290	269.977	265.662
23) L5 AR-1248-3	6.273	5.138	14333820	75193193	260.182	273.023m
24) L5 AR-1248-4	6.671	5.312	18043522	71202640	273.096	254.322
25) L5 AR-1248-5	6.710	5.701	17824378	145.0E6	277.971	265.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:30
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:24:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075904.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:46
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:29:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.780	7254273	27080772	4.751	3.914m
2) SA Decachlor...	10.422	8.783	6253432	36720401	5.099	4.188
Target Compounds						
21) L5 AR-1248-1	5.798	4.873	1700512	17991268	47.407	47.898
22) L5 AR-1248-2	6.069	5.101	2473817	12956514	50.751	52.688
23) L5 AR-1248-3	6.273	5.138	2547901	16065921	46.953	56.000
24) L5 AR-1248-4	6.669	5.313	3376964	14683789	49.218m	51.704m
25) L5 AR-1248-5	6.708	5.730	3390515	20852592	50.206m	38.783m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075904.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:46
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:29:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:02
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 15:57:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.783	149.9E6	746.7E6	97.273	98.623
2)	SA Decachlor...	10.424	8.784	119.5E6	929.4E6	95.942	97.673
Target Compounds							
26)	L6 AR-1254-1	6.646	5.662	66261960	570.9E6	872.670m	970.606
27)	L6 AR-1254-2	6.864	5.809	87851253	484.3E6	952.211	965.216
28)	L6 AR-1254-3	7.226	6.211	92035771	866.8E6	959.587	985.524
29)	L6 AR-1254-4	7.509	6.437	73136868	656.6E6	959.001	987.570
30)	L6 AR-1254-5	7.925	6.854	85440842	714.9E6	961.668	997.772

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:02
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:57:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:18
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:03:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.644	3.783	116.9E6	582.0E6	75.579	76.230
2)	SA Decachlor...	10.423	8.784	93392685	723.9E6	74.989	75.714
Target Compounds							
26)	L6 AR-1254-1	6.645	5.661	53923759	447.8E6	731.053m	757.518
27)	L6 AR-1254-2	6.863	5.808	69427935	372.6E6	751.680	745.041
28)	L6 AR-1254-3	7.225	6.210	72675518	670.2E6	755.137	757.967
29)	L6 AR-1254-4	7.507	6.436	57517400	515.4E6	752.790	766.592
30)	L6 AR-1254-5	7.923	6.851	67290078	595.4E6	754.900	785.342m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:18
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:03:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:35
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 15:24:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.647	3.785	79163213	383.8E6	50.000	50.000
2)	SA Decachlor...	10.424	8.785	64802936	486.9E6	50.000	50.000
Target Compounds							
26)	L6 AR-1254-1	6.648	5.663	39135010	302.7E6	480.703m	500.000
27)	L6 AR-1254-2	6.866	5.810	48334685	259.6E6	500.000	500.000
28)	L6 AR-1254-3	7.228	6.212	49893996	446.1E6	500.000	500.000
29)	L6 AR-1254-4	7.510	6.438	39695143	336.6E6	500.000	500.000
30)	L6 AR-1254-5	7.926	6.854	46126096	359.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:35
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:24:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075908.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:51
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:07:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	40353962	181.6E6	25.803	24.079
2) SA Decachlor...	10.422	8.784	33314499	235.8E6	26.290	24.742
Target Compounds						
26) L6 AR-1254-1	6.647	5.662	20616230	146.6E6	278.580m	248.519
27) L6 AR-1254-2	6.864	5.809	25702390	128.2E6	270.622	254.691
28) L6 AR-1254-3	7.226	6.210	26497010	218.4E6	268.520	247.733
29) L6 AR-1254-4	7.508	6.436	20991838	159.2E6	268.108	239.944
30) L6 AR-1254-5	7.925	6.853	23753414	169.8E6	262.160	233.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075908.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:51
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:07:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075909.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:07
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:12:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.780	8604114	35027317	5.393	4.744m
2) SA Decachlor...	10.419	8.782	7574299	50050602	5.752	5.200
Target Compounds						
26) L6 AR-1254-1	6.644	5.659	5028603	23004893	64.906	40.789 #
27) L6 AR-1254-2	6.860	5.807	6001024	31864124	60.020	60.390
28) L6 AR-1254-3	7.222	6.206	6253065	36465652	60.152	42.846 #
29) L6 AR-1254-4	7.504	6.433	5244227	24763386	62.720	39.321 #
30) L6 AR-1254-5	7.921	6.850	4771352	34659929	52.106	48.147

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075909.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:07
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:12:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075910.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:40
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:21:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:20:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.785	76058633	352.0E6	50.000	50.000
2) SA Decachlor...	10.425	8.786	61011705	448.1E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.232	6.893	42013007	493.5E6	500.000	500.000
37) L8 AR-1262-2	8.557	7.150	81834582	351.6E6	500.000	500.000
38) L8 AR-1262-3	8.883	7.672	60893486	303.5E6	500.000	500.000
39) L8 AR-1262-4	8.971	7.736	46055772	550.7E6	500.000	500.000
40) L8 AR-1262-5	9.642	8.231	33912301	230.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075910.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:40
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:21:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:20:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:56
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:02:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	151.5E6	749.9E6	97.157	99.641
2) SA Decachlor...	10.424	8.785	198.8E6	1671.8E6	95.543	99.055
Target Compounds						
41) L9 AR-1268-1	8.878	7.673	186.7E6	1831.6E6	951.896	992.615
42) L9 AR-1268-2	8.973	7.737	172.5E6	1800.4E6	956.477	1001.376
43) L9 AR-1268-3	9.213	7.944	145.2E6	1369.8E6	961.447	988.116
44) L9 AR-1268-4	9.640	8.230	71036137	542.8E6	963.668	1002.321
45) L9 AR-1268-5	10.071	8.526	446.4E6	3855.5E6	968.612	979.920



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:56
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075912.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:12
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:06:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.783	116.2E6	564.9E6	74.674	75.037
2) SA Decachlor...	10.423	8.784	155.7E6	1249.5E6	74.863	74.351
Target Compounds						
41) L9 AR-1268-1	8.877	7.671	146.3E6	1360.1E6	747.184	741.339
42) L9 AR-1268-2	8.973	7.736	134.4E6	1329.3E6	746.666	742.894
43) L9 AR-1268-3	9.211	7.943	112.7E6	1019.0E6	747.385	739.990
44) L9 AR-1268-4	9.640	8.229	56104264	402.6E6	757.366	745.626
45) L9 AR-1268-5	10.070	8.525	344.2E6	2925.3E6	747.861	745.651



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075912.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:12
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:06:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075913.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:29
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:59:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.785	80186895	377.7E6	50.000	50.000
2) SA Decachlor...	10.424	8.785	108.7E6	851.9E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.878	7.673	102.8E6	929.4E6	500.000	500.000
42) L9 AR-1268-2	8.973	7.737	94092052	897.7E6	500.000	500.000
43) L9 AR-1268-3	9.213	7.943	78447140	701.4E6	500.000	500.000
44) L9 AR-1268-4	9.641	8.229	38196242	270.1E6	500.000	500.000
45) L9 AR-1268-5	10.070	8.526	237.7E6	2006.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075913.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:29
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:59:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075914.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:45
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:11:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.648	3.785	39777145	176.8E6	25.420	23.849
2) SA Decachlor...	10.427	8.786	55292786	404.9E6	26.176	24.316
Target Compounds						
41) L9 AR-1268-1	8.880	7.673	52505558	433.0E6	263.374	239.361
42) L9 AR-1268-2	8.975	7.737	47519594	415.5E6	260.422	236.426
43) L9 AR-1268-3	9.214	7.943	39522518	339.7E6	258.947	247.491
44) L9 AR-1268-4	9.642	8.231	19156803	125.4E6	256.397	236.500
45) L9 AR-1268-5	10.073	8.527	120.0E6	927.0E6	257.945	239.576



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075914.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:45
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:11:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075915.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 17:01
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:15:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.785	8297245	34278784	5.239	4.694
2) SA Decachlor...	10.424	8.786	11780168	79254993	5.451	4.805
Target Compounds						
41) L9 AR-1268-1	8.878	7.671	11762673	90866583	56.952	50.184
42) L9 AR-1268-2	8.973	7.736	10571240	82258397	56.152	47.408
43) L9 AR-1268-3	9.212	7.944	8270442	67275006	53.294	49.210
44) L9 AR-1268-4	9.641	8.228	4718687	23283603	59.998	44.997 #
45) L9 AR-1268-5	10.071	8.525	27065230	169.1E6	56.340	44.825



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075915.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 17:01
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:15:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075916.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 17:34
 Operator : YP\AJ
 Sample : PP102125ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:26:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm


Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	73895870	340.9E6	51.708	51.507
2) SA Decachlor...	10.423	8.784	58947888	432.2E6	54.757	52.559
Target Compounds						
3) L1 AR-1016-1	5.799	4.879	30579076	390.9E6	517.184	528.086
4) L1 AR-1016-2	5.820	4.937	46283521	183.5E6	523.308	536.724
5) L1 AR-1016-3	5.883	5.057	29014100	102.8E6	524.184	500.470
6) L1 AR-1016-4	5.980	5.097	23843232	102.7E6	524.717	549.750
7) L1 AR-1016-5	6.273	5.312	22384300	107.6E6	515.702	472.171
31) L7 AR-1260-1	7.391	6.340	38313477	286.2E6	557.233	522.496
32) L7 AR-1260-2	7.644	6.526	47189702	452.3E6	526.804	555.781
33) L7 AR-1260-3	8.002	6.680	37186118	328.9E6	536.700	584.366
34) L7 AR-1260-4	8.230	7.149	40498377	308.8E6	590.602	594.624
35) L7 AR-1260-5	8.556	7.389	79566551	796.8E6	537.788	587.781

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Instrument :
ECD_P
ClientSampleId :
ICVPP102125

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075917.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 18:06
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1242

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 08:16:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	79486082	365.2E6	55.620	55.173
2) SA Decachlor...	10.423	8.785	64567530	484.9E6	59.978m	58.969
Target Compounds						
16) L4 AR-1242-1	5.799	4.879	24532478	301.7E6	512.088	524.891
17) L4 AR-1242-2	5.820	4.937	37086619	138.8E6	523.883	499.981
18) L4 AR-1242-3	5.883	5.057	22990427	78773054	503.714	500.427
19) L4 AR-1242-4	5.980	5.139	18909181	99357590	520.064	536.743m
20) L4 AR-1242-5	6.710	5.660	21676788	123.9E6	529.426	538.030

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:06
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1242

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:16:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075918.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 18:39
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1248

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 08:17:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.784	77354608	375.4E6	54.128	56.717
2) SA Decachlor...	10.425	8.785	62090346	474.3E6	57.677	57.676
Target Compounds						
21) L5 AR-1248-1	5.799	4.878	18454423	193.8E6	514.472	515.973
22) L5 AR-1248-2	6.070	5.099	24718037	124.2E6	507.095	505.124
23) L5 AR-1248-3	6.273	5.138	27349933	132.7E6	504.009	462.616m
24) L5 AR-1248-4	6.671	5.312	33802442	144.9E6	509.348	512.423
25) L5 AR-1248-5	6.711	5.701	33065578	279.8E6	509.794	538.456

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:39
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1248

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:17:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075919.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 08:49
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1254

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 09:02:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.654	3.784	73513288	361.3E6	51.441	54.591
2) SA Decachlor...	10.439	8.789	61994754	468.2E6	57.588	56.937
Target Compounds						
26) L6 AR-1254-1	6.657	5.662	33305659	337.3E6	416.876	597.997m#
27) L6 AR-1254-2	6.873	5.811	45460472	248.5E6	454.676	468.721
28) L6 AR-1254-3	7.235	6.214	47667437	420.5E6	458.540	494.093
29) L6 AR-1254-4	7.517	6.439	37597676	337.3E6	449.659	535.605
30) L6 AR-1254-5	7.934	6.855	44185805	348.6E6	482.532	484.302

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 08:49
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 09:02:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075920.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 09:07
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 15:40:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:38:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.648	3.782	80645176	382.7E6	50.921	52.402
2) SA Decachlor...	10.430	8.784	112.8E6	913.8E6	52.199	55.404
Target Compounds						
41) L9 AR-1268-1	8.881	7.671	106.5E6	989.5E6	515.825	546.466
42) L9 AR-1268-2	8.976	7.736	98184801	960.8E6	521.533	553.714
43) L9 AR-1268-3	9.216	7.942	81566200	749.4E6	525.609	548.157
44) L9 AR-1268-4	9.645	8.228	39672536	280.8E6	504.438	542.576
45) L9 AR-1268-5	10.076	8.525	246.0E6	2111.9E6	512.030	559.930



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 09:07
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 15:40:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:38:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 14:00

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.57	5.47	5.67	0.00
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-2	(2)	7.33	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.68	7.58	7.78	0.00
Aroclor-1260-4	(4)	7.90	7.90	7.80	8.00	0.00
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 14:00

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.21	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.00
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114936.D **Time Analyzed:** 14:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.487	5.387	5.587	505.510	500.000	1.1
Aroclor-1016-2	5.509	5.408	5.608	500.610	500.000	0.1
Aroclor-1016-3	5.570	5.470	5.670	519.930	500.000	4.0
Aroclor-1016-4	5.666	5.566	5.766	511.720	500.000	2.3
Aroclor-1016-5	5.958	5.858	6.058	486.570	500.000	-2.7
Aroclor-1260-1	7.072	6.971	7.171	485.660	500.000	-2.9
Aroclor-1260-2	7.325	7.225	7.425	513.500	500.000	2.7
Aroclor-1260-3	7.683	7.582	7.782	513.020	500.000	2.6
Aroclor-1260-4	7.904	7.803	8.003	499.360	500.000	-0.1
Aroclor-1260-5	8.210	8.110	8.310	502.550	500.000	0.5
Decachlorobiphenyl	9.930	9.830	10.030	49.940	50.000	-0.1
Tetrachloro-m-xylene	4.348	4.248	4.448	51.460	50.000	2.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114936.D **Time Analyzed:** 14:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	491.610	500.000	-1.7
Aroclor-1016-2	4.745	4.647	4.847	494.270	500.000	-1.1
Aroclor-1016-3	4.920	4.821	5.021	500.530	500.000	0.1
Aroclor-1016-4	4.962	4.864	5.064	496.850	500.000	-0.6
Aroclor-1016-5	5.174	5.076	5.276	502.840	500.000	0.6
Aroclor-1260-1	6.205	6.106	6.306	483.630	500.000	-3.3
Aroclor-1260-2	6.392	6.294	6.494	488.750	500.000	-2.3
Aroclor-1260-3	6.545	6.446	6.646	489.890	500.000	-2.0
Aroclor-1260-4	7.016	6.917	7.117	498.910	500.000	-0.2
Aroclor-1260-5	7.257	7.158	7.358	493.840	500.000	-1.2
Decachlorobiphenyl	8.639	8.541	8.741	52.450	50.000	4.9
Tetrachloro-m-xylene	3.648	3.548	3.748	49.210	50.000	-1.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
 Data File : PO114936.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 14:00
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:34:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	362.2E6	219.8E6	51.458	49.214
2) SA Decachlor...	9.930	8.639	220.8E6	204.7E6	49.942	52.446
Target Compounds						
3) L1 AR-1016-1	5.487	4.726	124.4E6	93530825	505.507	491.606
4) L1 AR-1016-2	5.509	4.745	198.8E6	133.9E6	500.613	494.266
5) L1 AR-1016-3	5.570	4.920	115.9E6	74065629	519.926	500.534
6) L1 AR-1016-4	5.666	4.962	96651891	58339098	511.724	496.851
7) L1 AR-1016-5	5.958	5.174	86704886	76791197	486.567	502.845
31) L7 AR-1260-1	7.072	6.205	148.4E6	143.2E6	485.660	483.632
32) L7 AR-1260-2	7.325	6.392	203.3E6	210.2E6	513.504	488.753
33) L7 AR-1260-3	7.683	6.545	170.2E6	172.0E6	513.017	489.887
34) L7 AR-1260-4	7.904	7.016	152.3E6	151.2E6	499.356	498.913
35) L7 AR-1260-5	8.210	7.257	358.4E6	405.8E6	502.550	493.841



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110625\
Data File : PO114936.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 14:00
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:34:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 19:28

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.57	5.47	5.67	0.00
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-2	(2)	7.32	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.68	7.58	7.78	0.00
Aroclor-1260-4	(4)	7.90	7.90	7.80	8.00	0.00
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 19:28

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.20	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.01
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114951.D **Time Analyzed:** 19:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.486	5.387	5.587	512.950	500.000	2.6
Aroclor-1016-2	5.508	5.408	5.608	504.850	500.000	1.0
Aroclor-1016-3	5.569	5.470	5.670	531.690	500.000	6.3
Aroclor-1016-4	5.665	5.566	5.766	520.950	500.000	4.2
Aroclor-1016-5	5.957	5.858	6.058	497.090	500.000	-0.6
Aroclor-1260-1	7.070	6.971	7.171	491.430	500.000	-1.7
Aroclor-1260-2	7.324	7.225	7.425	525.920	500.000	5.2
Aroclor-1260-3	7.682	7.582	7.782	524.330	500.000	4.9
Aroclor-1260-4	7.903	7.803	8.003	511.630	500.000	2.3
Aroclor-1260-5	8.210	8.110	8.310	510.830	500.000	2.2
Decachlorobiphenyl	9.928	9.830	10.030	50.820	50.000	1.6
Tetrachloro-m-xylene	4.347	4.248	4.448	52.160	50.000	4.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114951.D **Time Analyzed:** 19:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	493.530	500.000	-1.3
Aroclor-1016-2	4.745	4.647	4.847	495.870	500.000	-0.8
Aroclor-1016-3	4.920	4.821	5.021	502.760	500.000	0.6
Aroclor-1016-4	4.961	4.864	5.064	498.680	500.000	-0.3
Aroclor-1016-5	5.174	5.076	5.276	501.150	500.000	0.2
Aroclor-1260-1	6.204	6.106	6.306	481.910	500.000	-3.6
Aroclor-1260-2	6.392	6.294	6.494	484.950	500.000	-3.0
Aroclor-1260-3	6.544	6.446	6.646	488.220	500.000	-2.4
Aroclor-1260-4	7.015	6.917	7.117	496.560	500.000	-0.7
Aroclor-1260-5	7.256	7.158	7.358	494.130	500.000	-1.2
Decachlorobiphenyl	8.639	8.541	8.741	52.800	50.000	5.6
Tetrachloro-m-xylene	3.647	3.548	3.748	49.810	50.000	-0.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
 Data File : PO114951.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 19:28
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:39:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.347	3.647	367.1E6	222.5E6	52.163	49.811
2)	SA Decachlor...	9.928	8.639	224.6E6	206.1E6	50.822	52.796
Target Compounds							
3)	L1 AR-1016-1	5.486	4.726	126.2E6	93896122	512.954	493.526
4)	L1 AR-1016-2	5.508	4.745	200.5E6	134.3E6	504.851	495.868
5)	L1 AR-1016-3	5.569	4.920	118.6E6	74395554	531.689	502.764
6)	L1 AR-1016-4	5.665	4.961	98395068	58553486	520.953	498.676
7)	L1 AR-1016-5	5.957	5.174	88580836	76532697	497.095	501.152
31)	L7 AR-1260-1	7.070	6.204	150.2E6	142.7E6	491.430	481.915
32)	L7 AR-1260-2	7.324	6.392	208.2E6	208.5E6	525.921	484.947
33)	L7 AR-1260-3	7.682	6.544	174.0E6	171.4E6	524.327	488.223
34)	L7 AR-1260-4	7.903	7.015	156.1E6	150.5E6	511.626	496.559
35)	L7 AR-1260-5	8.210	7.256	364.3E6	406.0E6	510.827	494.125



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110625\
Data File : PO114951.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 19:28
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:39:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 14:02

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 14:02

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.77	8.78	8.68	8.88	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076352.D **Time Analyzed:** 14:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.793	5.695	5.895	500.170	500.000	0.0
Aroclor-1016-2	5.814	5.717	5.917	500.320	500.000	0.1
Aroclor-1016-3	5.877	5.780	5.980	505.820	500.000	1.2
Aroclor-1016-4	5.974	5.877	6.077	509.170	500.000	1.8
Aroclor-1016-5	6.266	6.169	6.369	501.660	500.000	0.3
Aroclor-1260-1	7.384	7.287	7.487	529.390	500.000	5.9
Aroclor-1260-2	7.637	7.540	7.740	518.180	500.000	3.6
Aroclor-1260-3	7.995	7.899	8.099	536.210	500.000	7.2
Aroclor-1260-4	8.223	8.127	8.327	546.660	500.000	9.3
Aroclor-1260-5	8.548	8.452	8.652	507.390	500.000	1.5
Decachlorobiphenyl	10.411	10.319	10.519	47.980	50.000	-4.0
Tetrachloro-m-xylene	4.641	4.542	4.742	49.360	50.000	-1.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076352.D **Time Analyzed:** 14:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.872	4.778	4.978	487.590	500.000	-2.5
Aroclor-1016-2	4.931	4.836	5.036	466.830	500.000	-6.6
Aroclor-1016-3	5.050	4.954	5.154	458.550	500.000	-8.3
Aroclor-1016-4	5.091	4.996	5.196	523.780	500.000	4.8
Aroclor-1016-5	5.306	5.210	5.410	466.790	500.000	-6.6
Aroclor-1260-1	6.332	6.238	6.438	534.840	500.000	7.0
Aroclor-1260-2	6.519	6.425	6.625	531.060	500.000	6.2
Aroclor-1260-3	6.673	6.578	6.778	529.430	500.000	5.9
Aroclor-1260-4	7.142	7.047	7.247	594.100	500.000	18.8
Aroclor-1260-5	7.382	7.287	7.487	538.080	500.000	7.6
Decachlorobiphenyl	8.774	8.682	8.882	44.670	50.000	-10.7
Tetrachloro-m-xylene	3.778	3.681	3.881	51.150	50.000	2.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
 Data File : PP076352.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 14:02
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/07/2025
 Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:24:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.641	3.778	70541349	338.5E6	49.361	51.149
2) SA Decachlor...	10.411	8.774	51655376	367.3E6	47.983	44.671
Target Compounds						
3) L1 AR-1016-1	5.793	4.872	29573173	360.9E6	500.171	487.594m
4) L1 AR-1016-2	5.814	4.931	44250135	159.6E6	500.317	466.827
5) L1 AR-1016-3	5.877	5.050	27997934	94159259	505.825	458.547
6) L1 AR-1016-4	5.974	5.091	23136693	97887748	509.168	523.784
7) L1 AR-1016-5	6.266	5.306	21774792	106.4E6	501.659	466.794
31) L7 AR-1260-1	7.384	6.332	36399384	292.9E6	529.395	534.837
32) L7 AR-1260-2	7.637	6.519	46416853	432.1E6	518.177	531.063
33) L7 AR-1260-3	7.995	6.673	37152372	297.9E6	536.213	529.427
34) L7 AR-1260-4	8.223	7.142	37485132	308.5E6	546.659	594.098
35) L7 AR-1260-5	8.548	7.382	75068789	729.4E6	507.387	538.078

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076352.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 14:02
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:24:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 18:54

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.87	5.88	5.78	5.98	0.01
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.26	6.27	6.17	6.37	0.01
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.63	7.64	7.54	7.74	0.01
Aroclor-1260-3	(3)	7.99	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.54	8.55	8.45	8.65	0.01
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.40	10.42	10.32	10.52	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Continuing Calib Date: 11/06/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 18:54

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.30	5.31	5.21	5.41	0.01
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.77	8.78	8.68	8.88	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076367.D **Time Analyzed:** 18:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.786	5.695	5.895	532.580	500.000	6.5
Aroclor-1016-2	5.807	5.717	5.917	445.900	500.000	-10.8
Aroclor-1016-3	5.870	5.780	5.980	467.410	500.000	-6.5
Aroclor-1016-4	5.967	5.877	6.077	478.000	500.000	-4.4
Aroclor-1016-5	6.259	6.169	6.369	478.990	500.000	-4.2
Aroclor-1260-1	7.378	7.287	7.487	525.640	500.000	5.1
Aroclor-1260-2	7.631	7.540	7.740	512.000	500.000	2.4
Aroclor-1260-3	7.989	7.899	8.099	524.980	500.000	5.0
Aroclor-1260-4	8.217	8.127	8.327	518.980	500.000	3.8
Aroclor-1260-5	8.542	8.452	8.652	500.170	500.000	0.0
Decachlorobiphenyl	10.404	10.319	10.519	47.610	50.000	-4.8
Tetrachloro-m-xylene	4.635	4.542	4.742	49.100	50.000	-1.8



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3552
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076367.D **Time Analyzed:** 18:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.869	4.778	4.978	482.510	500.000	-3.5
Aroclor-1016-2	4.928	4.836	5.036	463.880	500.000	-7.2
Aroclor-1016-3	5.048	4.954	5.154	430.090	500.000	-14.0
Aroclor-1016-4	5.090	4.996	5.196	484.090	500.000	-3.2
Aroclor-1016-5	5.302	5.210	5.410	452.950	500.000	-9.4
Aroclor-1260-1	6.329	6.238	6.438	539.390	500.000	7.9
Aroclor-1260-2	6.516	6.425	6.625	537.630	500.000	7.5
Aroclor-1260-3	6.668	6.578	6.778	531.440	500.000	6.3
Aroclor-1260-4	7.137	7.047	7.247	583.090	500.000	16.6
Aroclor-1260-5	7.378	7.287	7.487	543.820	500.000	8.8
Decachlorobiphenyl	8.770	8.682	8.882	45.790	50.000	-8.4
Tetrachloro-m-xylene	3.777	3.681	3.881	49.890	50.000	-0.2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
 Data File : PP076367.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 18:54
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/07/2025
 Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:29:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.635	3.777	70165098	330.2E6	49.098	49.888
2) SA Decachlor...	10.404	8.770	51258412	376.5E6	47.615	45.792
Target Compounds						
3) L1 AR-1016-1	5.786	4.869	31489324	357.1E6	532.579m	482.510m
4) L1 AR-1016-2	5.807	4.928	39437051	158.6E6	445.898m	463.880
5) L1 AR-1016-3	5.870	5.048	25871472	88316650	467.407m	430.094
6) L1 AR-1016-4	5.967	5.090	21720490	90469312	478.002m	484.089
7) L1 AR-1016-5	6.259	5.302	20790693	103.3E6	478.987m	452.948
31) L7 AR-1260-1	7.378	6.329	36141010	295.4E6	525.637	539.386
32) L7 AR-1260-2	7.631	6.516	45863827	437.5E6	512.003	537.634
33) L7 AR-1260-3	7.989	6.668	36374334	299.1E6	524.984	531.442m
34) L7 AR-1260-4	8.217	7.137	35587416	302.8E6	518.984	583.087
35) L7 AR-1260-5	8.542	7.378	74001174	737.2E6	500.171	543.815

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076367.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 18:54
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/07/2025
Supervised By :Yogesh Patel 11/07/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:29:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: Remington & Vernick Engineers	SDG No.: Q3552
Project: Edison Landfill	Instrument ID: ECD_O
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 11/04/2025 11/04/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	11/04/2025	09:21	PO114870.D	9.93	4.35
AR1660ICC1000	AR1660ICC1000	11/04/2025	09:39	PO114871.D	9.93	4.35
AR1660ICC750	AR1660ICC750	11/04/2025	09:58	PO114872.D	9.93	4.35
AR1660ICC500	AR1660ICC500	11/04/2025	10:16	PO114873.D	9.93	4.35
AR1660ICC250	AR1660ICC250	11/04/2025	10:34	PO114874.D	9.93	4.35
AR1660ICC050	AR1660ICC050	11/04/2025	10:53	PO114875.D	9.93	4.35
AR1221ICC500	AR1221ICC500	11/04/2025	11:11	PO114876.D	9.93	4.35
AR1232ICC500	AR1232ICC500	11/04/2025	11:30	PO114877.D	9.93	4.35
AR1242ICC1000	AR1242ICC1000	11/04/2025	11:48	PO114878.D	9.93	4.35
AR1242ICC750	AR1242ICC750	11/04/2025	12:06	PO114879.D	9.93	4.35
AR1242ICC500	AR1242ICC500	11/04/2025	12:25	PO114880.D	9.93	4.35
AR1242ICC250	AR1242ICC250	11/04/2025	12:42	PO114881.D	9.93	4.35
AR1242ICC050	AR1242ICC050	11/04/2025	13:00	PO114882.D	9.93	4.35
AR1248ICC1000	AR1248ICC1000	11/04/2025	13:19	PO114883.D	9.93	4.35
AR1248ICC750	AR1248ICC750	11/04/2025	13:37	PO114884.D	9.93	4.35
AR1248ICC500	AR1248ICC500	11/04/2025	13:56	PO114885.D	9.93	4.35
AR1248ICC250	AR1248ICC250	11/04/2025	14:14	PO114886.D	9.93	4.35
AR1248ICC050	AR1248ICC050	11/04/2025	14:33	PO114887.D	9.93	4.35
AR1254ICC1000	AR1254ICC1000	11/04/2025	14:51	PO114888.D	9.93	4.35
AR1254ICC750	AR1254ICC750	11/04/2025	15:09	PO114889.D	9.93	4.35
AR1254ICC500	AR1254ICC500	11/04/2025	15:28	PO114890.D	9.93	4.35
AR1254ICC250	AR1254ICC250	11/04/2025	16:04	PO114891.D	9.93	4.35
AR1254ICC050	AR1254ICC050	11/04/2025	16:23	PO114892.D	9.93	4.35
AR1262ICC500	AR1262ICC500	11/04/2025	17:00	PO114893.D	9.95	4.37
AR1268ICC1000	AR1268ICC1000	11/04/2025	17:18	PO114894.D	9.93	4.35
AR1268ICC750	AR1268ICC750	11/04/2025	17:36	PO114895.D	9.93	4.35
AR1268ICC500	AR1268ICC500	11/04/2025	17:55	PO114896.D	9.93	4.35
AR1268ICC250	AR1268ICC250	11/04/2025	18:13	PO114897.D	9.93	4.35
AR1268ICC050	AR1268ICC050	11/04/2025	18:31	PO114898.D	9.93	4.35
AR1660CCC500	AR1660CCC500	11/06/2025	14:00	PO114936.D	9.93	4.35
IBLK	IBLK	11/06/2025	15:13	PO114940.D	9.93	4.35
PB170425BL	PB170425BL	11/06/2025	15:32	PO114941.D	9.93	4.35
PB170425BS	PB170425BS	11/06/2025	15:50	PO114942.D	9.93	4.35
PB170425BSD	PB170425BSD	11/06/2025	16:08	PO114943.D	9.93	4.34
AR1660CCC500	AR1660CCC500	11/06/2025	19:28	PO114951.D	9.93	4.35
IBLK	IBLK	11/06/2025	21:17	PO114955.D	9.93	4.35
IBLK	IBLK	10/21/2025	08:49	PP075887.D	10.42	4.64
AR1660ICC1000	AR1660ICC1000	10/21/2025	09:06	PP075888.D	10.42	4.65
AR1660ICC750	AR1660ICC750	10/21/2025	09:22	PP075889.D	10.42	4.64
AR1660ICC500	AR1660ICC500	10/21/2025	09:38	PP075890.D	10.42	4.64
AR1660ICC250	AR1660ICC250	10/21/2025	09:54	PP075891.D	10.42	4.65
AR1660ICC050	AR1660ICC050	10/21/2025	10:11	PP075892.D	10.42	4.65

Analytical Sequence

AR1221ICC500	AR1221ICC500	10/21/2025	10:27	PP075893.D	10.42	4.64
AR12321ICC500	AR12321ICC500	10/21/2025	10:43	PP075894.D	10.42	4.64
AR12421ICC1000	AR12421ICC1000	10/21/2025	10:59	PP075895.D	10.42	4.65
AR12421ICC750	AR12421ICC750	10/21/2025	11:16	PP075896.D	10.42	4.64
AR12421ICC500	AR12421ICC500	10/21/2025	11:48	PP075897.D	10.42	4.65
AR12421ICC250	AR12421ICC250	10/21/2025	12:04	PP075898.D	10.42	4.65
AR12421ICC050	AR12421ICC050	10/21/2025	12:21	PP075899.D	10.42	4.65
AR12481ICC1000	AR12481ICC1000	10/21/2025	12:41	PP075900.D	10.43	4.65
AR12481ICC750	AR12481ICC750	10/21/2025	12:57	PP075901.D	10.42	4.64
AR12481ICC500	AR12481ICC500	10/21/2025	13:13	PP075902.D	10.42	4.65
AR12481ICC250	AR12481ICC250	10/21/2025	13:30	PP075903.D	10.42	4.65
AR12481ICC050	AR12481ICC050	10/21/2025	13:46	PP075904.D	10.42	4.65
AR12541ICC1000	AR12541ICC1000	10/21/2025	14:02	PP075905.D	10.42	4.65
AR12541ICC750	AR12541ICC750	10/21/2025	14:18	PP075906.D	10.42	4.64
AR12541ICC500	AR12541ICC500	10/21/2025	14:35	PP075907.D	10.42	4.65
AR12541ICC250	AR12541ICC250	10/21/2025	14:51	PP075908.D	10.42	4.65
AR12541ICC050	AR12541ICC050	10/21/2025	15:07	PP075909.D	10.42	4.64
AR12621ICC500	AR12621ICC500	10/21/2025	15:40	PP075910.D	10.43	4.65
AR12681ICC1000	AR12681ICC1000	10/21/2025	15:56	PP075911.D	10.42	4.65
AR12681ICC750	AR12681ICC750	10/21/2025	16:12	PP075912.D	10.42	4.65
AR12681ICC500	AR12681ICC500	10/21/2025	16:29	PP075913.D	10.42	4.65
AR12681ICC250	AR12681ICC250	10/21/2025	16:45	PP075914.D	10.43	4.65
AR12681ICC050	AR12681ICC050	10/21/2025	17:01	PP075915.D	10.42	4.65
AR1660CCC500	AR1660CCC500	11/06/2025	14:02	PP076352.D	10.41	4.64
LBLK	LBLK	11/06/2025	15:07	PP076356.D	10.41	4.64
MW-4	Q3552-01	11/06/2025	16:44	PP076362.D	10.41	4.64
MW-8	Q3552-02	11/06/2025	17:00	PP076363.D	10.41	4.64
MW-9	Q3552-03	11/06/2025	17:17	PP076364.D	10.41	4.64
MW-10	Q3552-04	11/06/2025	17:33	PP076365.D	10.41	4.64
MW-3	Q3552-05	11/06/2025	17:49	PP076366.D	10.41	4.64
AR1660CCC500	AR1660CCC500	11/06/2025	18:54	PP076367.D	10.40	4.64
LBLK	LBLK	11/06/2025	20:48	PP076371.D	10.41	4.64

Analytical Sequence

Client: Remington & Vernick Engineers	SDG No.: Q3552
Project: Edison Landfill	Instrument ID: ECD_O
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 11/04/2025 11/04/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	11/04/2025	09:21	PO114870.D	8.64	3.65
AR1660ICC1000	AR1660ICC1000	11/04/2025	09:39	PO114871.D	8.64	3.65
AR1660ICC750	AR1660ICC750	11/04/2025	09:58	PO114872.D	8.64	3.65
AR1660ICC500	AR1660ICC500	11/04/2025	10:16	PO114873.D	8.64	3.65
AR1660ICC250	AR1660ICC250	11/04/2025	10:34	PO114874.D	8.64	3.65
AR1660ICC050	AR1660ICC050	11/04/2025	10:53	PO114875.D	8.64	3.65
AR1221ICC500	AR1221ICC500	11/04/2025	11:11	PO114876.D	8.64	3.65
AR1232ICC500	AR1232ICC500	11/04/2025	11:30	PO114877.D	8.64	3.65
AR1242ICC1000	AR1242ICC1000	11/04/2025	11:48	PO114878.D	8.64	3.65
AR1242ICC750	AR1242ICC750	11/04/2025	12:06	PO114879.D	8.64	3.65
AR1242ICC500	AR1242ICC500	11/04/2025	12:25	PO114880.D	8.64	3.65
AR1242ICC250	AR1242ICC250	11/04/2025	12:42	PO114881.D	8.64	3.65
AR1242ICC050	AR1242ICC050	11/04/2025	13:00	PO114882.D	8.64	3.65
AR1248ICC1000	AR1248ICC1000	11/04/2025	13:19	PO114883.D	8.64	3.65
AR1248ICC750	AR1248ICC750	11/04/2025	13:37	PO114884.D	8.64	3.65
AR1248ICC500	AR1248ICC500	11/04/2025	13:56	PO114885.D	8.64	3.65
AR1248ICC250	AR1248ICC250	11/04/2025	14:14	PO114886.D	8.64	3.65
AR1248ICC050	AR1248ICC050	11/04/2025	14:33	PO114887.D	8.64	3.65
AR1254ICC1000	AR1254ICC1000	11/04/2025	14:51	PO114888.D	8.64	3.65
AR1254ICC750	AR1254ICC750	11/04/2025	15:09	PO114889.D	8.64	3.65
AR1254ICC500	AR1254ICC500	11/04/2025	15:28	PO114890.D	8.64	3.65
AR1254ICC250	AR1254ICC250	11/04/2025	16:04	PO114891.D	8.64	3.65
AR1254ICC050	AR1254ICC050	11/04/2025	16:23	PO114892.D	8.64	3.65
AR1262ICC500	AR1262ICC500	11/04/2025	17:00	PO114893.D	8.64	3.65
AR1268ICC1000	AR1268ICC1000	11/04/2025	17:18	PO114894.D	8.64	3.65
AR1268ICC750	AR1268ICC750	11/04/2025	17:36	PO114895.D	8.64	3.65
AR1268ICC500	AR1268ICC500	11/04/2025	17:55	PO114896.D	8.64	3.65
AR1268ICC250	AR1268ICC250	11/04/2025	18:13	PO114897.D	8.64	3.65
AR1268ICC050	AR1268ICC050	11/04/2025	18:31	PO114898.D	8.64	3.65
AR1660CCC500	AR1660CCC500	11/06/2025	14:00	PO114936.D	8.64	3.65
IBLK	IBLK	11/06/2025	15:13	PO114940.D	8.64	3.65
PB170425BL	PB170425BL	11/06/2025	15:32	PO114941.D	8.64	3.65
PB170425BS	PB170425BS	11/06/2025	15:50	PO114942.D	8.64	3.65
PB170425BSD	PB170425BSD	11/06/2025	16:08	PO114943.D	8.64	3.65
AR1660CCC500	AR1660CCC500	11/06/2025	19:28	PO114951.D	8.64	3.65
IBLK	IBLK	11/06/2025	21:17	PO114955.D	8.64	3.65
IBLK	IBLK	10/21/2025	08:49	PP075887.D	8.78	3.78
AR1660ICC1000	AR1660ICC1000	10/21/2025	09:06	PP075888.D	8.78	3.78
AR1660ICC750	AR1660ICC750	10/21/2025	09:22	PP075889.D	8.78	3.78
AR1660ICC500	AR1660ICC500	10/21/2025	09:38	PP075890.D	8.78	3.78
AR1660ICC250	AR1660ICC250	10/21/2025	09:54	PP075891.D	8.78	3.78
AR1660ICC050	AR1660ICC050	10/21/2025	10:11	PP075892.D	8.78	3.78

Analytical Sequence

AR1221ICC500	AR1221ICC500	10/21/2025	10:27	PP075893.D	8.78	3.78
AR1232ICC500	AR1232ICC500	10/21/2025	10:43	PP075894.D	8.78	3.78
AR1242ICC1000	AR1242ICC1000	10/21/2025	10:59	PP075895.D	8.78	3.78
AR1242ICC750	AR1242ICC750	10/21/2025	11:16	PP075896.D	8.78	3.78
AR1242ICC500	AR1242ICC500	10/21/2025	11:48	PP075897.D	8.79	3.78
AR1242ICC250	AR1242ICC250	10/21/2025	12:04	PP075898.D	8.78	3.78
AR1242ICC050	AR1242ICC050	10/21/2025	12:21	PP075899.D	8.78	3.78
AR1248ICC1000	AR1248ICC1000	10/21/2025	12:41	PP075900.D	8.78	3.78
AR1248ICC750	AR1248ICC750	10/21/2025	12:57	PP075901.D	8.78	3.78
AR1248ICC500	AR1248ICC500	10/21/2025	13:13	PP075902.D	8.78	3.78
AR1248ICC250	AR1248ICC250	10/21/2025	13:30	PP075903.D	8.78	3.78
AR1248ICC050	AR1248ICC050	10/21/2025	13:46	PP075904.D	8.78	3.78
AR1254ICC1000	AR1254ICC1000	10/21/2025	14:02	PP075905.D	8.78	3.78
AR1254ICC750	AR1254ICC750	10/21/2025	14:18	PP075906.D	8.78	3.78
AR1254ICC500	AR1254ICC500	10/21/2025	14:35	PP075907.D	8.79	3.79
AR1254ICC250	AR1254ICC250	10/21/2025	14:51	PP075908.D	8.78	3.78
AR1254ICC050	AR1254ICC050	10/21/2025	15:07	PP075909.D	8.78	3.78
AR1262ICC500	AR1262ICC500	10/21/2025	15:40	PP075910.D	8.79	3.79
AR1268ICC1000	AR1268ICC1000	10/21/2025	15:56	PP075911.D	8.79	3.78
AR1268ICC750	AR1268ICC750	10/21/2025	16:12	PP075912.D	8.78	3.78
AR1268ICC500	AR1268ICC500	10/21/2025	16:29	PP075913.D	8.79	3.79
AR1268ICC250	AR1268ICC250	10/21/2025	16:45	PP075914.D	8.79	3.79
AR1268ICC050	AR1268ICC050	10/21/2025	17:01	PP075915.D	8.79	3.79
AR1660CCC500	AR1660CCC500	11/06/2025	14:02	PP076352.D	8.77	3.78
LBLK	LBLK	11/06/2025	15:07	PP076356.D	8.78	3.78
MW-4	Q3552-01	11/06/2025	16:44	PP076362.D	8.77	3.77
MW-8	Q3552-02	11/06/2025	17:00	PP076363.D	8.77	3.78
MW-9	Q3552-03	11/06/2025	17:17	PP076364.D	8.77	3.77
MW-10	Q3552-04	11/06/2025	17:33	PP076365.D	8.77	3.78
MW-3	Q3552-05	11/06/2025	17:49	PP076366.D	8.77	3.78
AR1660CCC500	AR1660CCC500	11/06/2025	18:54	PP076367.D	8.77	3.78
LBLK	LBLK	11/06/2025	20:48	PP076371.D	8.77	3.78



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170425BS

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Lab Sample ID: PB170425BS

Date(s) Analyzed: 11/06/2025 11/06/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file P0114942.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.486	5.436	5.536	2.25	2.20	4.65
COLUMN 1	2	5.508	5.458	5.558	2.21		
	3	5.57	5.52	5.62	2.30		
	4	5.665	5.615	5.715	2.26		
	5	5.957	5.907	6.007	2.18		
	1	4.727	4.677	4.777	2.15		
COLUMN 2	2	4.746	4.696	4.796	2.18	2.10	
	3	4.92	4.87	4.97	2.16		
	4	4.963	4.913	5.013	2.12		
	5	5.174	5.124	5.224	2.09		
Aroclor-1260	1	7.071	7.021	7.121	2.22	2.10	0
COLUMN 1	2	7.324	7.274	7.374	2.38		
	3	7.682	7.632	7.732	1.97		
	4	7.904	7.854	7.954	2.07		
	5	8.21	8.16	8.26	1.99		
	1	6.205	6.155	6.255	2.20		
COLUMN 2	2	6.392	6.342	6.442	2.21	2.10	
	3	6.544	6.494	6.594	2.18		
	4	7.015	6.965	7.065	1.96		
	5	7.256	7.206	7.306	1.97		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170425BSD

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

Lab Sample ID: PB170425BSD

Date(s) Analyzed: 11/06/2025 11/06/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO114943.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.483	5.433	5.533	2.27	2.20	4.65
COLUMN 1	2	5.505	5.455	5.555	2.21		
	3	5.567	5.517	5.617	2.31		
	4	5.662	5.612	5.712	2.25		
	5	5.954	5.904	6.004	2.17		
	1	4.725	4.675	4.775	2.13		
COLUMN 2	2	4.744	4.694	4.794	2.13	2.10	
	3	4.918	4.868	4.968	2.18		
	4	4.96	4.91	5.01	2.14		
	5	5.172	5.122	5.222	2.10		
Aroclor-1260	1	7.068	7.018	7.118	2.25	2.10	0
COLUMN 1	2	7.322	7.272	7.372	2.30		
	3	7.679	7.629	7.729	1.97		
	4	7.901	7.851	7.951	2.07		
	5	8.208	8.158	8.258	1.98		
	1	6.203	6.153	6.253	2.20		
COLUMN 2	2	6.39	6.34	6.44	2.19	2.10	
	3	6.543	6.493	6.593	2.20		
	4	7.014	6.964	7.064	1.96		
	5	7.255	7.205	7.305	1.97		



QC SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Remington & Vernick Engineers
Project: Edison Landfill
Client Sample ID: PB170425BL
Lab Sample ID: PB170425BL
Analytical Method: 8082A
Sample Wt/Vol: 1000 mL Final Vol: 5000 uL
Prep Method: 3510C Prep Date: 11/06/25

Date Collected:
Date Received:
SDG No.: Q3552
Matrix: WATER
% Solid: 0
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.049	U	1	0.049	0.25	ug/L	11/06/25 15:32	PB170425
11104-28-2	Aroclor-1221	0.065	U	1	0.065	0.25	ug/L	11/06/25 15:32	PB170425
11141-16-5	Aroclor-1232	0.048	U	1	0.048	0.25	ug/L	11/06/25 15:32	PB170425
53469-21-9	Aroclor-1242	0.060	U	1	0.060	0.25	ug/L	11/06/25 15:32	PB170425
12672-29-6	Aroclor-1248	0.036	U	1	0.036	0.25	ug/L	11/06/25 15:32	PB170425
11097-69-1	Aroclor-1254	0.047	U	1	0.047	0.25	ug/L	11/06/25 15:32	PB170425
37324-23-5	Aroclor-1262	0.070	U	1	0.070	0.25	ug/L	11/06/25 15:32	PB170425
11100-14-4	Aroclor-1268	0.055	U	1	0.055	0.25	ug/L	11/06/25 15:32	PB170425
11096-82-5	Aroclor-1260	0.041	U	1	0.041	0.25	ug/L	11/06/25 15:32	PB170425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	20.6			30 (30) - 150 (173)	103%	SPK: 20		
2051-24-3	Decachlorobiphenyl	23.0			30 (10) - 150 (173)	115%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
Data File : PO114941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 15:32
Operator : YP/AJ
Sample : PB170425BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170425BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:36:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	143.9E6	91775633	20.453	20.546
2) SA Decachlor...	9.931	8.639	93823936	89846853	21.226	23.015

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
Data File : PO114941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 15:32
Operator : YP/AJ
Sample : PB170425BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170425BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:36:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: 0.000 min
Response: 143949782
Conc: 20.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB170425BL




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 91775633
Conc: 20.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.931 min
Delta R.T.: 0.000 min
Response: 93823936
Conc: 21.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.639 min
Delta R.T.: -0.002 min
Response: 89846853
Conc: 23.02 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	11/04/25
Project:	Edison Landfill		Date Received:	11/04/25
Client Sample ID:	PIBLK-PO114870.D		SDG No.:	Q3552
Lab Sample ID:	I.BLK-PO114870.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/04/25	po110425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/04/25	po110425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/04/25	po110425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/04/25	po110425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/04/25	po110425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/04/25	po110425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/04/25	po110425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/04/25	po110425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/04/25	po110425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	15.5			70 (60) - 130 (140)	77%	SPK: 20		
2051-24-3	Decachlorobiphenyl	16.8			70 (60) - 130 (140)	84%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:16:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:14:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.649	108.8E6	72481407	15.462	16.227
2) SA Decachlor...	9.930	8.641	74180497	67480996	16.782	17.286

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110425\
Data File : PO114870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:16:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:14:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: 0.000 min
Response: 108823881
Conc: 15.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: 0.000 min
Response: 72481407
Conc: 16.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.930 min
Delta R.T.: 0.000 min
Response: 74180497
Conc: 16.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.641 min
Delta R.T.: 0.000 min
Response: 67480996
Conc: 17.29 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	11/06/25
Project:	Edison Landfill		Date Received:	11/06/25
Client Sample ID:	PIBLK-PO114940.D		SDG No.:	Q3552
Lab Sample ID:	I.BLK-PO114940.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/06/25	po110625
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/06/25	po110625
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/06/25	po110625
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/06/25	po110625
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/06/25	po110625
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/06/25	po110625
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/06/25	po110625
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/06/25	po110625
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/06/25	po110625
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.7			70 (60) - 130 (140)	94%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.3			70 (60) - 130 (140)	96%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
 Data File : PO114940.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 15:13
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:35:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.648	131.6E6	84284003	18.700	18.869
2) SA Decachlor...	9.930	8.638	85107499	81330330	19.254	20.834

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
Data File : PO114940.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 15:13
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:35:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: 0.000 min
Response: 131615447
Conc: 18.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 84284003
Conc: 18.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.930 min
Delta R.T.: 0.000 min
Response: 85107499
Conc: 19.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.638 min
Delta R.T.: -0.003 min
Response: 81330330
Conc: 20.83 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	11/06/25
Project:	Edison Landfill		Date Received:	11/06/25
Client Sample ID:	PIBLK-PO114955.D		SDG No.:	Q3552
Lab Sample ID:	I.BLK-PO114955.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/06/25	po110625
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/06/25	po110625
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/06/25	po110625
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/06/25	po110625
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/06/25	po110625
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/06/25	po110625
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/06/25	po110625
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/06/25	po110625
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/06/25	po110625
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.1			70 (60) - 130 (140)	96%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.6			70 (60) - 130 (140)	98%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
 Data File : PO114955.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 21:17
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:40:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.647	134.7E6	85826961	19.136	19.215
2) SA Decachlor...	9.930	8.638	86567664	81660686	19.584	20.918

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO110625\
Data File : PO114955.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 21:17
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:40:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.347 min
Delta R.T.: 0.000 min
Response: 134678533
Conc: 19.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: 0.000 min
Response: 85826961
Conc: 19.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.930 min
Delta R.T.: 0.000 min
Response: 86567664
Conc: 19.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.638 min
Delta R.T.: -0.003 min
Response: 81660686
Conc: 20.92 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	10/21/25
Project:	Edison Landfill		Date Received:	10/21/25
Client Sample ID:	PIBLK-PP075887.D		SDG No.:	Q3552
Lab Sample ID:	I.BLK-PP075887.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/21/25	PP102125
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/21/25	PP102125
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/21/25	PP102125
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/21/25	PP102125
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/21/25	PP102125
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/21/25	PP102125
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/21/25	PP102125
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/21/25	PP102125
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/21/25	PP102125
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.8			70 (60) - 130 (140)	94%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.3			70 (60) - 130 (140)	96%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075887.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 08:49
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:26:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.783	28503905	124.7E6	19.945	18.843
2) SA Decachlor...	10.419	8.784	21974278	158.6E6	20.412m	19.281

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 08:49
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:26:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP075887.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 4.644 min
 Delta R.T.: 0.002 min
 Response: 28503905
 Conc: 19.95 ng/ml

Instrument :

ECD_P

ClientSampleId :


I.BLK

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025


Time Response_ Signal: PP075887.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.783 min
 Delta R.T.: 0.002 min
 Response: 124709992
 Conc: 18.84 ng/ml


Time Response_ Signal: PP075887.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.419 min
 Delta R.T.: 0.000 min
 Response: 21974278
 Conc: 20.41 ng/ml m

Time Response_ Signal: PP075887.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.784 min
 Delta R.T.: 0.002 min
 Response: 158550080
 Conc: 19.28 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	11/06/25
Project:	Edison Landfill		Date Received:	11/06/25
Client Sample ID:	PIBLK-PP076356.D		SDG No.:	Q3552
Lab Sample ID:	I.BLK-PP076356.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/06/25	pp110625
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/06/25	pp110625
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/06/25	pp110625
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/06/25	pp110625
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/06/25	pp110625
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/06/25	pp110625
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/06/25	pp110625
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/06/25	pp110625
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/06/25	pp110625
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.0			70 (60) - 130 (140)	90%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.1			70 (60) - 130 (140)	85%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
 Data File : PP076356.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 15:07
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:25:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.639	3.779	28452589	119.3E6	19.910	18.020
2)	SA Decachlor...	10.409	8.775	21608845	140.3E6	20.073	17.056

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076356.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 15:07
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:25:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 28452589
Conc: 19.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.002 min
Response: 119264703
Conc: 18.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.409 min
Delta R.T.: -0.010 min
Response: 21608845
Conc: 20.07 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.775 min
Delta R.T.: -0.007 min
Response: 140254484
Conc: 17.06 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Remington & Vernick Engineers		Date Collected:	11/06/25
Project:	Edison Landfill		Date Received:	11/06/25
Client Sample ID:	PIBLK-PP076371.D		SDG No.:	Q3552
Lab Sample ID:	I.BLK-PP076371.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/06/25	pp110625
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/06/25	pp110625
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/06/25	pp110625
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/06/25	pp110625
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/06/25	pp110625
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/06/25	pp110625
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/06/25	pp110625
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/06/25	pp110625
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/06/25	pp110625
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.2			70 (60) - 130 (140)	91%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.7			70 (60) - 130 (140)	88%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076371.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 20:48
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:30:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.637	3.776	29124182	120.3E6	20.379	18.172
2)	SA Decachlor...	10.406	8.769	21070454	145.2E6	19.573	17.660

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110625\
Data File : PP076371.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 20:48
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:30:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 29124182
Conc: 20.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.776 min
Delta R.T.: -0.006 min
Response: 120268977
Conc: 18.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.406 min
Delta R.T.: -0.013 min
Response: 21070454
Conc: 19.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.769 min
Delta R.T.: -0.013 min
Response: 145217347
Conc: 17.66 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Remington & Vernick Engineers
Project: Edison Landfill
Client Sample ID: PB170425BS
Lab Sample ID: PB170425BS
Analytical Method: 8082A
Sample Wt/Vol: 1000 mL Final Vol: 5000 uL
Prep Method: 3510C Prep Date: 11/06/25

Date Collected:
Date Received:
SDG No.: Q3552
Matrix: WATER
% Solid: 0
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	2.20		1	0.049	0.25	ug/L	11/06/25 15:50	PB170425
11104-28-2	Aroclor-1221	0.065	U	1	0.065	0.25	ug/L	11/06/25 15:50	PB170425
11141-16-5	Aroclor-1232	0.048	U	1	0.048	0.25	ug/L	11/06/25 15:50	PB170425
53469-21-9	Aroclor-1242	0.060	U	1	0.060	0.25	ug/L	11/06/25 15:50	PB170425
12672-29-6	Aroclor-1248	0.036	U	1	0.036	0.25	ug/L	11/06/25 15:50	PB170425
11097-69-1	Aroclor-1254	0.047	U	1	0.047	0.25	ug/L	11/06/25 15:50	PB170425
37324-23-5	Aroclor-1262	0.070	U	1	0.070	0.25	ug/L	11/06/25 15:50	PB170425
11100-14-4	Aroclor-1268	0.055	U	1	0.055	0.25	ug/L	11/06/25 15:50	PB170425
11096-82-5	Aroclor-1260	2.10		1	0.041	0.25	ug/L	11/06/25 15:50	PB170425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.3			30 (30) - 150 (173)	106%	SPK: 20		
2051-24-3	Decachlorobiphenyl	23.5			30 (10) - 150 (173)	117%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
 Data File : PO114942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 15:50
 Operator : YP/AJ
 Sample : PB170425BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB170425BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:36:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.347	3.647	149.6E6	88358215	21.250	19.781
2) SA Decachlor...	9.930	8.639	96922052	91607653	21.927	23.466
Target Compounds						
3) L1 AR-1016-1	5.486	4.727	110.6E6	81659454	449.710	429.209
4) L1 AR-1016-2	5.508	4.746	175.7E6	117.9E6	442.551	435.096
5) L1 AR-1016-3	5.570	4.920	102.5E6	63954410	459.589	432.203
6) L1 AR-1016-4	5.665	4.963	85541096	49725582	452.898	423.493
7) L1 AR-1016-5	5.957	5.174	77536715	63826844	435.118	417.952
31) L7 AR-1260-1	7.071	6.205	135.7E6	130.5E6	444.121	440.844
32) L7 AR-1260-2	7.324	6.392	188.1E6	189.7E6	475.173	441.117
33) L7 AR-1260-3	7.682	6.544	130.4E6	153.4E6	393.058	436.959
34) L7 AR-1260-4	7.904	7.015	126.6E6	118.7E6	414.909	391.648
35) L7 AR-1260-5	8.210	7.256	284.6E6	323.9E6	398.994	394.187


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
Data File : PO114942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 15:50
Operator : YP/AJ
Sample : PB170425BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170425BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:36:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: Remington & Vernick Engineers
Project: Edison Landfill
Client Sample ID: PB170425BSD
Lab Sample ID: PB170425BSD
Analytical Method: 8082A
Sample Wt/Vol: 1000 mL Final Vol: 5000 uL
Prep Method: 3510C Prep Date: 11/06/25

Date Collected:
Date Received:
SDG No.: Q3552
Matrix: WATER
% Solid: 0
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	2.20		1	0.049	0.25	ug/L	11/06/25 16:08	PB170425
11104-28-2	Aroclor-1221	0.065	U	1	0.065	0.25	ug/L	11/06/25 16:08	PB170425
11141-16-5	Aroclor-1232	0.048	U	1	0.048	0.25	ug/L	11/06/25 16:08	PB170425
53469-21-9	Aroclor-1242	0.060	U	1	0.060	0.25	ug/L	11/06/25 16:08	PB170425
12672-29-6	Aroclor-1248	0.036	U	1	0.036	0.25	ug/L	11/06/25 16:08	PB170425
11097-69-1	Aroclor-1254	0.047	U	1	0.047	0.25	ug/L	11/06/25 16:08	PB170425
37324-23-5	Aroclor-1262	0.070	U	1	0.070	0.25	ug/L	11/06/25 16:08	PB170425
11100-14-4	Aroclor-1268	0.055	U	1	0.055	0.25	ug/L	11/06/25 16:08	PB170425
11096-82-5	Aroclor-1260	2.10		1	0.041	0.25	ug/L	11/06/25 16:08	PB170425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.6			30 (30) - 150 (173)	108%	SPK: 20		
2051-24-3	Decachlorobiphenyl	23.7			30 (10) - 150 (173)	119%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
 Data File : PO114943.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2025 16:08
 Operator : YP/AJ
 Sample : PB170425BSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB170425BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:36:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.344	3.646	152.0E6	90431578	21.590	20.245
2) SA Decachlor...	9.926	8.637	98857066	92639578	22.365	23.730
Target Compounds						
3) L1 AR-1016-1	5.483	4.725	111.6E6	80915333	453.746	425.298
4) L1 AR-1016-2	5.505	4.744	175.2E6	115.3E6	441.239	425.710
5) L1 AR-1016-3	5.567	4.918	102.9E6	64381435	461.577	435.089
6) L1 AR-1016-4	5.662	4.960	85128922	50175003	450.715	427.320
7) L1 AR-1016-5	5.954	5.172	77180123	64056423	433.117	419.455
31) L7 AR-1260-1	7.068	6.203	137.6E6	130.3E6	450.248	440.197
32) L7 AR-1260-2	7.322	6.390	182.1E6	188.4E6	459.972	438.004
33) L7 AR-1260-3	7.679	6.543	131.0E6	154.8E6	394.726	440.899
34) L7 AR-1260-4	7.901	7.014	126.1E6	118.7E6	413.441	391.772
35) L7 AR-1260-5	8.208	7.255	282.7E6	323.1E6	396.426	393.277


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110625\
Data File : PO114943.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2025 16:08
Operator : YP/AJ
Sample : PB170425BSD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170425BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:36:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	po110425	Instrument	ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PO114874.D	AR-1260-1 #2	rahul	11/5/2025 8:39:44 AM	yogesh	11/5/2025 10:22:06	Peak Integrated by Software
AR1660ICC250	PO114874.D	AR-1260-2 #2	rahul	11/5/2025 8:39:44 AM	yogesh	11/5/2025 10:22:06	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1016-3 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1016-5 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-1	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-1 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-2 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-3	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-4	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1242ICC050	PO114882.D	AR-1242-5 #2	rahul	11/5/2025 8:39:51 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	AR-1254-1	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	AR-1254-2	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software

Manual Integration Report

Sequence:

po110425

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254ICC050	PO114892.D	AR-1254-4	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	Decachlorobiphenyl	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1268ICC250	PO114897.D	AR-1268-1 #2	rahul	11/5/2025 8:40:09 AM	yogesh	11/5/2025 10:22:13	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-1	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-2	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-3	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-4	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

po110625

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PO114954.D	AR-1254-2	rahul	11/7/2025 10:38:10 AM	yogesh	11/7/2025 10:48:48	Peak Integrated by Software

Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PP075887.D	Decachlorobiphenyl	yogesh	10/22/2025 12:24:41 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-3	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-3 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-4	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-4 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-5 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1232ICC500	PP075894.D	AR-1232-4 #2	yogesh	10/22/2025 12:24:44 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC100 0	PP075895.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:45 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC750	PP075896.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:46 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC500	PP075897.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:48 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC250	PP075898.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:50 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC250	PP075898.D	AR-1242-5	yogesh	10/22/2025 12:24:50 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PP075899.D	AR-1242-5	yogesh	10/22/2025 12:24:52 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC050	PP075899.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:24:52 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC100 0	PP075900.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:53 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC750	PP075901.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:55 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC500	PP075902.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:57 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC250	PP075903.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:58 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-4	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-4 #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-5	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-5 #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC100 0	PP075905.D	AR-1254-1	yogesh	10/22/2025 12:25:02 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC750	PP075906.D	AR-1254-1	yogesh	10/22/2025 12:25:03 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:

PP102125

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254ICC750	PP075906.D	AR-1254-5 #2	yogesh	10/22/2025 12:25:03 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC500	PP075907.D	AR-1254-1	yogesh	10/22/2025 12:25:05 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC250	PP075908.D	AR-1254-1	yogesh	10/22/2025 12:25:07 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC050	PP075909.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:09 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICV500	PP075917.D	AR-1242-4 #2	yogesh	10/24/2025 4:12:30 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICV500	PP075917.D	Decachlorobiphenyl	yogesh	10/24/2025 4:12:30 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICV500	PP075918.D	AR-1248-3 #2	yogesh	10/24/2025 4:12:35 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICV500	PP075919.D	AR-1254-1 #2	yogesh	10/24/2025 4:12:33 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:	Pp110625	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP076343.D	AR-1016-1 #2	rahul	11/7/2025 10:41:56 AM	yogesh	11/7/2025 10:52:15	Peak Integrated by Software
AR1242CCC500	PP076344.D	AR-1242-4 #2	rahul	11/7/2025 10:42:04 AM	yogesh	11/7/2025 10:52:18	Peak Integrated by Software
AR1242CCC500	PP076344.D	AR-1242-5 #2	rahul	11/7/2025 10:42:04 AM	yogesh	11/7/2025 10:52:18	Peak Integrated by Software
AR1248CCC500	PP076345.D	AR-1248-3 #2	rahul	11/7/2025 10:42:09 AM	yogesh	11/7/2025 10:52:21	Peak Integrated by Software
AR1248CCC500	PP076345.D	AR-1248-5 #2	rahul	11/7/2025 10:42:09 AM	yogesh	11/7/2025 10:52:21	Peak Integrated by Software
AR1254CCC500	PP076346.D	AR-1254-1	rahul	11/7/2025 10:42:12 AM	yogesh	11/7/2025 10:52:24	Peak Integrated by Software
AR1254CCC500	PP076346.D	Tetrachloro-m-xylene	rahul	11/7/2025 10:42:12 AM	yogesh	11/7/2025 10:52:24	Peak Integrated by Software
AR1660CCC500	PP076352.D	AR-1016-1 #2	rahul	11/7/2025 10:42:23 AM	yogesh	11/7/2025 10:52:35	Peak Integrated by Software
AR1242CCC500	PP076353.D	AR-1242-1 #2	rahul	11/7/2025 10:42:26 AM	yogesh	11/7/2025 10:52:37	Peak Integrated by Software
AR1242CCC500	PP076353.D	AR-1242-4 #2	rahul	11/7/2025 10:42:26 AM	yogesh	11/7/2025 10:52:37	Peak Integrated by Software
AR1248CCC500	PP076354.D	AR-1248-1	rahul	11/7/2025 10:42:29 AM	yogesh	11/7/2025 10:52:38	Peak Integrated by Software
AR1248CCC500	PP076354.D	AR-1248-3 #2	rahul	11/7/2025 10:42:29 AM	yogesh	11/7/2025 10:52:38	Peak Integrated by Software
AR1254CCC500	PP076355.D	AR-1254-1	rahul	11/7/2025 10:42:33 AM	yogesh	11/7/2025 10:52:40	Peak Integrated by Software

Manual Integration Report

Sequence:	Pp110625	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PP076355.D	Tetrachloro-m-xylene	rahul	11/7/2025 10:42:33 AM	yogesh	11/7/2025 10:52:40	Peak Integrated by Software
Q3552-01	PP076362.D	Tetrachloro-m-xylene	rahul	11/7/2025 10:42:49 AM	yogesh	11/7/2025 10:52:54	Peak Integrated by Software
Q3552-01	PP076362.D	Tetrachloro-m-xylene #2	rahul	11/7/2025 10:42:49 AM	yogesh	11/7/2025 10:52:54	Peak Integrated by Software
Q3552-03	PP076364.D	Tetrachloro-m-xylene	rahul	11/7/2025 10:42:51 AM	yogesh	11/7/2025 10:53:20	Peak Integrated by Software
Q3552-03	PP076364.D	Tetrachloro-m-xylene #2	rahul	11/7/2025 10:42:51 AM	yogesh	11/7/2025 10:53:20	Peak Integrated by Software
Q3552-04	PP076365.D	Tetrachloro-m-xylene	rahul	11/7/2025 10:42:56 AM	yogesh	11/7/2025 10:53:22	Peak Integrated by Software
Q3552-04	PP076365.D	Tetrachloro-m-xylene #2	rahul	11/7/2025 10:42:56 AM	yogesh	11/7/2025 10:53:22	Peak Integrated by Software
Q3552-05	PP076366.D	Tetrachloro-m-xylene	rahul	11/7/2025 10:42:58 AM	yogesh	11/7/2025 10:53:24	Peak Integrated by Software
Q3552-05	PP076366.D	Tetrachloro-m-xylene #2	rahul	11/7/2025 10:42:58 AM	yogesh	11/7/2025 10:53:24	Peak Integrated by Software
AR1660CCC500	PP076367.D	AR-1016-1	rahul	11/7/2025 10:43:04 AM	yogesh	11/7/2025 10:53:26	Peak Integrated by Software
AR1660CCC500	PP076367.D	AR-1016-1 #2	rahul	11/7/2025 10:43:04 AM	yogesh	11/7/2025 10:53:26	Peak Integrated by Software
AR1660CCC500	PP076367.D	AR-1016-2	rahul	11/7/2025 10:43:04 AM	yogesh	11/7/2025 10:53:26	Peak Integrated by Software
AR1660CCC500	PP076367.D	AR-1016-3	rahul	11/7/2025 10:43:04 AM	yogesh	11/7/2025 10:53:26	Peak Integrated by Software

Manual Integration Report

Sequence:	Pp110625	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP076367.D	AR-1016-4	rahul	11/7/2025 10:43:04 AM	yogesh	11/7/2025 10:53:26	Peak Integrated by Software
AR1660CCC500	PP076367.D	AR-1016-5	rahul	11/7/2025 10:43:04 AM	yogesh	11/7/2025 10:53:26	Peak Integrated by Software
AR1660CCC500	PP076367.D	AR-1260-3 #2	rahul	11/7/2025 10:43:04 AM	yogesh	11/7/2025 10:53:26	Peak Integrated by Software
AR1242CCC500	PP076368.D	AR-1242-1	rahul	11/7/2025 10:43:07 AM	yogesh	11/7/2025 10:53:28	Peak Integrated by Software
AR1242CCC500	PP076368.D	AR-1242-1 #2	rahul	11/7/2025 10:43:07 AM	yogesh	11/7/2025 10:53:28	Peak Integrated by Software
AR1242CCC500	PP076368.D	AR-1242-2	rahul	11/7/2025 10:43:07 AM	yogesh	11/7/2025 10:53:28	Peak Integrated by Software
AR1242CCC500	PP076368.D	AR-1242-3	rahul	11/7/2025 10:43:07 AM	yogesh	11/7/2025 10:53:28	Peak Integrated by Software
AR1242CCC500	PP076368.D	AR-1242-4	rahul	11/7/2025 10:43:07 AM	yogesh	11/7/2025 10:53:28	Peak Integrated by Software
AR1242CCC500	PP076368.D	AR-1242-4 #2	rahul	11/7/2025 10:43:07 AM	yogesh	11/7/2025 10:53:28	Peak Integrated by Software
AR1242CCC500	PP076368.D	AR-1242-5	rahul	11/7/2025 10:43:07 AM	yogesh	11/7/2025 10:53:28	Peak Integrated by Software
AR1248CCC500	PP076369.D	AR-1248-1	rahul	11/7/2025 10:43:10 AM	yogesh	11/7/2025 10:53:33	Peak Integrated by Software
AR1248CCC500	PP076369.D	AR-1248-2	rahul	11/7/2025 10:43:10 AM	yogesh	11/7/2025 10:53:33	Peak Integrated by Software
AR1248CCC500	PP076369.D	AR-1248-3	rahul	11/7/2025 10:43:10 AM	yogesh	11/7/2025 10:53:33	Peak Integrated by Software

Manual Integration Report

Sequence:

Pp110625

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248CCC500	PP076369.D	AR-1248-3 #2	rahul	11/7/2025 10:43:10 AM	yogesh	11/7/2025 10:53:33	Peak Integrated by Software
AR1248CCC500	PP076369.D	AR-1248-4	rahul	11/7/2025 10:43:10 AM	yogesh	11/7/2025 10:53:33	Peak Integrated by Software
AR1248CCC500	PP076369.D	AR-1248-5	rahul	11/7/2025 10:43:10 AM	yogesh	11/7/2025 10:53:33	Peak Integrated by Software
AR1254CCC500	PP076370.D	AR-1254-1	rahul	11/7/2025 10:43:13 AM	yogesh	11/7/2025 10:53:31	Peak Integrated by Software
AR1254CCC500	PP076370.D	AR-1254-3 #2	rahul	11/7/2025 10:43:13 AM	yogesh	11/7/2025 10:53:31	Peak Integrated by Software

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO114869.D	04 Nov 2025 09:03	YP/AJ	Ok
2	I.BLK	PO114870.D	04 Nov 2025 09:21	YP/AJ	Ok
3	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39	YP/AJ	Ok
4	AR1660ICC750	PO114872.D	04 Nov 2025 09:58	YP/AJ	Ok
5	AR1660ICC500	PO114873.D	04 Nov 2025 10:16	YP/AJ	Ok
6	AR1660ICC250	PO114874.D	04 Nov 2025 10:34	YP/AJ	Ok,M
7	AR1660ICC050	PO114875.D	04 Nov 2025 10:53	YP/AJ	Ok,M
8	AR1221ICC500	PO114876.D	04 Nov 2025 11:11	YP/AJ	Ok
9	AR1232ICC500	PO114877.D	04 Nov 2025 11:30	YP/AJ	Ok
10	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48	YP/AJ	Ok
11	AR1242ICC750	PO114879.D	04 Nov 2025 12:06	YP/AJ	Ok
12	AR1242ICC500	PO114880.D	04 Nov 2025 12:25	YP/AJ	Ok
13	AR1242ICC250	PO114881.D	04 Nov 2025 12:42	YP/AJ	Ok
14	AR1242ICC050	PO114882.D	04 Nov 2025 13:00	YP/AJ	Ok,M
15	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19	YP/AJ	Ok
16	AR1248ICC750	PO114884.D	04 Nov 2025 13:37	YP/AJ	Ok
17	AR1248ICC500	PO114885.D	04 Nov 2025 13:56	YP/AJ	Ok
18	AR1248ICC250	PO114886.D	04 Nov 2025 14:14	YP/AJ	Ok
19	AR1248ICC050	PO114887.D	04 Nov 2025 14:33	YP/AJ	Ok
20	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51	YP/AJ	Ok
21	AR1254ICC750	PO114889.D	04 Nov 2025 15:09	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO114890.D	04 Nov 2025 15:28	YP/AJ	Ok
23	AR1254ICC250	PO114891.D	04 Nov 2025 16:04	YP/AJ	Ok
24	AR1254ICC050	PO114892.D	04 Nov 2025 16:23	YP/AJ	Ok,M
25	AR1262ICC500	PO114893.D	04 Nov 2025 17:00	YP/AJ	Ok
26	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18	YP/AJ	Ok
27	AR1268ICC750	PO114895.D	04 Nov 2025 17:36	YP/AJ	Ok
28	AR1268ICC500	PO114896.D	04 Nov 2025 17:55	YP/AJ	Ok
29	AR1268ICC250	PO114897.D	04 Nov 2025 18:13	YP/AJ	Ok,M
30	AR1268ICC050	PO114898.D	04 Nov 2025 18:31	YP/AJ	Ok
31	PO110425ICV500	PO114899.D	04 Nov 2025 19:08	YP/AJ	Ok
32	AR1242ICV500	PO114900.D	04 Nov 2025 20:21	YP/AJ	Ok,M
33	AR1248ICV500	PO114901.D	04 Nov 2025 20:58	YP/AJ	Ok
34	AR1254ICV500	PO114902.D	04 Nov 2025 21:35	YP/AJ	Ok
35	AR1268ICV500	PO114903.D	04 Nov 2025 22:12	YP/AJ	Ok
36	DDT ANALOG	PO114904.D	04 Nov 2025 22:48	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110625

Review By	yogesh	Review On	11/6/2025 10:15:27 AM
Supervise By	rahul	Supervise On	11/7/2025 10:38:22 AM
SubDirectory	PO110625	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO114926.D	06 Nov 2025 09:02	YP/AJ	Ok
2	AR1660CCC500	PO114927.D	06 Nov 2025 09:20	YP/AJ	Ok
3	AR1242CCC500	PO114928.D	06 Nov 2025 09:39	YP/AJ	Ok
4	AR1248CCC500	PO114929.D	06 Nov 2025 09:58	YP/AJ	Ok
5	AR1254CCC500	PO114930.D	06 Nov 2025 10:16	YP/AJ	Ok
6	I.BLK	PO114931.D	06 Nov 2025 10:35	YP/AJ	Ok
7	Q3536-01	PO114932.D	06 Nov 2025 10:53	YP/AJ	Ok,M
8	PB170419BL	PO114933.D	06 Nov 2025 12:10	YP/AJ	Ok
9	PB170419BS	PO114934.D	06 Nov 2025 12:28	YP/AJ	Ok
10	Q3544-01	PO114935.D	06 Nov 2025 12:46	YP/AJ	Ok
11	AR1660CCC500	PO114936.D	06 Nov 2025 14:00	YP/AJ	Ok
12	AR1242CCC500	PO114937.D	06 Nov 2025 14:18	YP/AJ	Ok
13	AR1248CCC500	PO114938.D	06 Nov 2025 14:37	YP/AJ	Ok
14	AR1254CCC500	PO114939.D	06 Nov 2025 14:55	YP/AJ	Ok
15	I.BLK	PO114940.D	06 Nov 2025 15:13	YP/AJ	Ok
16	PB170425BL	PO114941.D	06 Nov 2025 15:32	YP/AJ	Ok
17	PB170425BS	PO114942.D	06 Nov 2025 15:50	YP/AJ	Ok
18	PB170425BSD	PO114943.D	06 Nov 2025 16:08	YP/AJ	Ok
19	Q3534-01	PO114944.D	06 Nov 2025 16:26	YP/AJ	Ok
20	Q3534-02	PO114945.D	06 Nov 2025 16:44	YP/AJ	Ok
21	Q3534-03	PO114946.D	06 Nov 2025 17:03	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110625

Review By	yogesh	Review On	11/6/2025 10:15:27 AM
Supervise By	rahul	Supervise On	11/7/2025 10:38:22 AM
SubDirectory	PO110625	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3534-04	PO114947.D	06 Nov 2025 17:20	YP/AJ	Ok
23	Q3534-05	PO114948.D	06 Nov 2025 17:39	YP/AJ	Ok
24	Q3534-06	PO114949.D	06 Nov 2025 17:57	YP/AJ	Ok
25	Q3534-07	PO114950.D	06 Nov 2025 18:15	YP/AJ	Not Ok
26	AR1660CCC500	PO114951.D	06 Nov 2025 19:28	YP/AJ	Ok
27	AR1242CCC500	PO114952.D	06 Nov 2025 20:22	YP/AJ	Ok
28	AR1248CCC500	PO114953.D	06 Nov 2025 20:41	YP/AJ	Ok
29	AR1254CCC500	PO114954.D	06 Nov 2025 20:59	YP/AJ	Ok,M
30	I.BLK	PO114955.D	06 Nov 2025 21:17	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075886.D	21 Oct 2025 08:33	YP\AJ	Ok
2	I.BLK	PP075887.D	21 Oct 2025 08:49	YP\AJ	Ok,M
3	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06	YP\AJ	Ok
4	AR1660ICC750	PP075889.D	21 Oct 2025 09:22	YP\AJ	Ok
5	AR1660ICC500	PP075890.D	21 Oct 2025 09:38	YP\AJ	Ok
6	AR1660ICC250	PP075891.D	21 Oct 2025 09:54	YP\AJ	Ok
7	AR1660ICC050	PP075892.D	21 Oct 2025 10:11	YP\AJ	Ok,M
8	AR1221ICC500	PP075893.D	21 Oct 2025 10:27	YP\AJ	Ok
9	AR1232ICC500	PP075894.D	21 Oct 2025 10:43	YP\AJ	Ok,M
10	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59	YP\AJ	Ok,M
11	AR1242ICC750	PP075896.D	21 Oct 2025 11:16	YP\AJ	Ok,M
12	AR1242ICC500	PP075897.D	21 Oct 2025 11:48	YP\AJ	Ok,M
13	AR1242ICC250	PP075898.D	21 Oct 2025 12:04	YP\AJ	Ok,M
14	AR1242ICC050	PP075899.D	21 Oct 2025 12:21	YP\AJ	Ok,M
15	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41	YP\AJ	Ok,M
16	AR1248ICC750	PP075901.D	21 Oct 2025 12:57	YP\AJ	Ok,M
17	AR1248ICC500	PP075902.D	21 Oct 2025 13:13	YP\AJ	Ok,M
18	AR1248ICC250	PP075903.D	21 Oct 2025 13:30	YP\AJ	Ok,M
19	AR1248ICC050	PP075904.D	21 Oct 2025 13:46	YP\AJ	Ok,M
20	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02	YP\AJ	Ok,M
21	AR1254ICC750	PP075906.D	21 Oct 2025 14:18	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP075907.D	21 Oct 2025 14:35	YP\AJ	Ok,M
23	AR1254ICC250	PP075908.D	21 Oct 2025 14:51	YP\AJ	Ok,M
24	AR1254ICC050	PP075909.D	21 Oct 2025 15:07	YP\AJ	Ok,M
25	AR1262ICC500	PP075910.D	21 Oct 2025 15:40	YP\AJ	Ok
26	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56	YP\AJ	Ok
27	AR1268ICC750	PP075912.D	21 Oct 2025 16:12	YP\AJ	Ok
28	AR1268ICC500	PP075913.D	21 Oct 2025 16:29	YP\AJ	Ok
29	AR1268ICC250	PP075914.D	21 Oct 2025 16:45	YP\AJ	Ok
30	AR1268ICC050	PP075915.D	21 Oct 2025 17:01	YP\AJ	Ok
31	PP102125ICV500	PP075916.D	21 Oct 2025 17:34	YP\AJ	Ok
32	AR1242ICV500	PP075917.D	21 Oct 2025 18:06	YP\AJ	Ok,M
33	AR1248ICV500	PP075918.D	21 Oct 2025 18:39	YP\AJ	Ok,M
34	AR1254ICV500	PP075919.D	22 Oct 2025 08:49	YP\AJ	Ok,M
35	AR1268ICV500	PP075920.D	22 Oct 2025 09:07	YP\AJ	Ok
36	DDT ANALOG	PP075921.D	22 Oct 2025 09:25	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110625

Review By	rahul	Review On	11/7/2025 10:43:29 AM
Supervise By	yogesh	Supervise On	11/7/2025 10:54:03 AM
SubDirectory	PP110625	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP076342.D	06 Nov 2025 08:56	YP\AJ	Ok
2	AR1660CCC500	PP076343.D	06 Nov 2025 09:30	YP\AJ	Ok,M
3	AR1242CCC500	PP076344.D	06 Nov 2025 09:51	YP\AJ	Ok,M
4	AR1248CCC500	PP076345.D	06 Nov 2025 10:07	YP\AJ	Ok,M
5	AR1254CCC500	PP076346.D	06 Nov 2025 10:23	YP\AJ	Ok,M
6	I.BLK	PP076347.D	06 Nov 2025 10:40	YP\AJ	Ok
7	Q3546-01	PP076348.D	06 Nov 2025 12:08	YP\AJ	Ok,M
8	Q3546-01MS	PP076349.D	06 Nov 2025 12:24	YP\AJ	Ok,M
9	Q3546-01MSD	PP076350.D	06 Nov 2025 12:40	YP\AJ	Ok,M
10	Q3549-02	PP076351.D	06 Nov 2025 12:56	YP\AJ	Ok
11	AR1660CCC500	PP076352.D	06 Nov 2025 14:02	YP\AJ	Ok,M
12	AR1242CCC500	PP076353.D	06 Nov 2025 14:18	YP\AJ	Ok,M
13	AR1248CCC500	PP076354.D	06 Nov 2025 14:34	YP\AJ	Ok,M
14	AR1254CCC500	PP076355.D	06 Nov 2025 14:50	YP\AJ	Not Ok
15	I.BLK	PP076356.D	06 Nov 2025 15:07	YP\AJ	Ok
16	Q3534-09	PP076357.D	06 Nov 2025 15:23	YP\AJ	Ok,M
17	Q3534-10	PP076358.D	06 Nov 2025 15:39	YP\AJ	Ok,M
18	Q3537-03	PP076359.D	06 Nov 2025 15:55	YP\AJ	Ok,M
19	Q3541-01	PP076360.D	06 Nov 2025 16:12	YP\AJ	Ok,M
20	Q3542-01	PP076361.D	06 Nov 2025 16:28	YP\AJ	Ok,M
21	Q3552-01	PP076362.D	06 Nov 2025 16:44	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110625

Review By	rahul	Review On	11/7/2025 10:43:29 AM
Supervise By	yogesh	Supervise On	11/7/2025 10:54:03 AM
SubDirectory	PP110625	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3552-02	PP076363.D	06 Nov 2025 17:00	YP\AJ	Ok
23	Q3552-03	PP076364.D	06 Nov 2025 17:17	YP\AJ	Ok,M
24	Q3552-04	PP076365.D	06 Nov 2025 17:33	YP\AJ	Ok,M
25	Q3552-05	PP076366.D	06 Nov 2025 17:49	YP\AJ	Ok,M
26	AR1660CCC500	PP076367.D	06 Nov 2025 18:54	YP\AJ	Ok,M
27	AR1242CCC500	PP076368.D	06 Nov 2025 19:43	YP\AJ	Ok,M
28	AR1248CCC500	PP076369.D	06 Nov 2025 19:59	YP\AJ	Ok,M
29	AR1254CCC500	PP076370.D	06 Nov 2025 20:15	YP\AJ	Ok,M
30	I.BLK	PP076371.D	06 Nov 2025 20:48	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114869.D	04 Nov 2025 09:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO114870.D	04 Nov 2025 09:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO114872.D	04 Nov 2025 09:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO114873.D	04 Nov 2025 10:16		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO114874.D	04 Nov 2025 10:34		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO114875.D	04 Nov 2025 10:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO114876.D	04 Nov 2025 11:11		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO114877.D	04 Nov 2025 11:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO114879.D	04 Nov 2025 12:06		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO114880.D	04 Nov 2025 12:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO114881.D	04 Nov 2025 12:42		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO114882.D	04 Nov 2025 13:00		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO114884.D	04 Nov 2025 13:37		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO114885.D	04 Nov 2025 13:56		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO114886.D	04 Nov 2025 14:14		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO114887.D	04 Nov 2025 14:33		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO114889.D	04 Nov 2025 15:09		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO114890.D	04 Nov 2025 15:28		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO114891.D	04 Nov 2025 16:04		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO114892.D	04 Nov 2025 16:23		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO114893.D	04 Nov 2025 17:00		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO114895.D	04 Nov 2025 17:36		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO114896.D	04 Nov 2025 17:55		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO114897.D	04 Nov 2025 18:13		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO114898.D	04 Nov 2025 18:31		YP/AJ	Ok
31	PO110425ICV500	ICVPO110425	PO114899.D	04 Nov 2025 19:08		YP/AJ	Ok
32	AR1242ICV500	ICVPO110425AR1242	PO114900.D	04 Nov 2025 20:21		YP/AJ	Ok,M
33	AR1248ICV500	ICVPO110425AR1248	PO114901.D	04 Nov 2025 20:58		YP/AJ	Ok
34	AR1254ICV500	ICVPO110425AR1254	PO114902.D	04 Nov 2025 21:35		YP/AJ	Ok
35	AR1268ICV500	ICVPO110425AR1268	PO114903.D	04 Nov 2025 22:12		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO114904.D	04 Nov 2025 22:48		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110625

Review By	yogesh	Review On	11/6/2025 10:15:27 AM
Supervise By	rahul	Supervise On	11/7/2025 10:38:22 AM
SubDirectory	PO110625	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114926.D	06 Nov 2025 09:02		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO114927.D	06 Nov 2025 09:20		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO114928.D	06 Nov 2025 09:39		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO114929.D	06 Nov 2025 09:58		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO114930.D	06 Nov 2025 10:16		YP/AJ	Ok
6	I.BLK	I.BLK	PO114931.D	06 Nov 2025 10:35		YP/AJ	Ok
7	Q3536-01	FILTER-COMP	PO114932.D	06 Nov 2025 10:53	TCMX high in 2nd column , AR1242 Hit	YP/AJ	Ok,M
8	PB170419BL	PB170419BL	PO114933.D	06 Nov 2025 12:10		YP/AJ	Ok
9	PB170419BS	PB170419BS	PO114934.D	06 Nov 2025 12:28		YP/AJ	Ok
10	Q3544-01	ONYX-1	PO114935.D	06 Nov 2025 12:46		YP/AJ	Ok
11	AR1660CCC500	AR1660CCC500	PO114936.D	06 Nov 2025 14:00		YP/AJ	Ok
12	AR1242CCC500	AR1242CCC500	PO114937.D	06 Nov 2025 14:18		YP/AJ	Ok
13	AR1248CCC500	AR1248CCC500	PO114938.D	06 Nov 2025 14:37		YP/AJ	Ok
14	AR1254CCC500	AR1254CCC500	PO114939.D	06 Nov 2025 14:55		YP/AJ	Ok
15	I.BLK	I.BLK	PO114940.D	06 Nov 2025 15:13		YP/AJ	Ok
16	PB170425BL	PB170425BL	PO114941.D	06 Nov 2025 15:32		YP/AJ	Ok
17	PB170425BS	PB170425BS	PO114942.D	06 Nov 2025 15:50		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110625

Review By	yogesh	Review On	11/6/2025 10:15:27 AM
Supervise By	rahul	Supervise On	11/7/2025 10:38:22 AM
SubDirectory	PO110625	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	PB170425BSD	PB170425BSD	PO114943.D	06 Nov 2025 16:08		YP/AJ	Ok
19	Q3534-01	MW-6	PO114944.D	06 Nov 2025 16:26		YP/AJ	Ok
20	Q3534-02	MW-1	PO114945.D	06 Nov 2025 16:44		YP/AJ	Ok
21	Q3534-03	MW-11	PO114946.D	06 Nov 2025 17:03		YP/AJ	Ok
22	Q3534-04	MW-5	PO114947.D	06 Nov 2025 17:20		YP/AJ	Ok
23	Q3534-05	MW-EB-1	PO114948.D	06 Nov 2025 17:39		YP/AJ	Ok
24	Q3534-06	MW-EB-2	PO114949.D	06 Nov 2025 17:57		YP/AJ	Ok
25	Q3534-07	MW-EB-3	PO114950.D	06 Nov 2025 18:15	need cleanup , confirm corrected typo	YP/AJ	Not Ok
26	AR1660CCC500	AR1660CCC500	PO114951.D	06 Nov 2025 19:28		YP/AJ	Ok
27	AR1242CCC500	AR1242CCC500	PO114952.D	06 Nov 2025 20:22		YP/AJ	Ok
28	AR1248CCC500	AR1248CCC500	PO114953.D	06 Nov 2025 20:41		YP/AJ	Ok
29	AR1254CCC500	AR1254CCC500	PO114954.D	06 Nov 2025 20:59		YP/AJ	Ok,M
30	I.BLK	I.BLK	PO114955.D	06 Nov 2025 21:17		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075886.D	21 Oct 2025 08:33		YP\AJ	Ok
2	I.BLK	I.BLK	PP075887.D	21 Oct 2025 08:49		YP\AJ	Ok,M
3	AR1660ICC1000	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075889.D	21 Oct 2025 09:22		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075890.D	21 Oct 2025 09:38		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075891.D	21 Oct 2025 09:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP075892.D	21 Oct 2025 10:11		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075893.D	21 Oct 2025 10:27		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075894.D	21 Oct 2025 10:43		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075896.D	21 Oct 2025 11:16		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075897.D	21 Oct 2025 11:48		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075898.D	21 Oct 2025 12:04		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075899.D	21 Oct 2025 12:21		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41		YP\AJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075901.D	21 Oct 2025 12:57		YP\AJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075902.D	21 Oct 2025 13:13		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075903.D	21 Oct 2025 13:30		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP075904.D	21 Oct 2025 13:46		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075906.D	21 Oct 2025 14:18		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075907.D	21 Oct 2025 14:35		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075908.D	21 Oct 2025 14:51		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075909.D	21 Oct 2025 15:07		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075910.D	21 Oct 2025 15:40		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075912.D	21 Oct 2025 16:12		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075913.D	21 Oct 2025 16:29		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075914.D	21 Oct 2025 16:45		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075915.D	21 Oct 2025 17:01		YPIAJ	Ok
31	PP102125ICV500	ICVPP102125	PP075916.D	21 Oct 2025 17:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP102125AR1242	PP075917.D	21 Oct 2025 18:06		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP102125AR1248	PP075918.D	21 Oct 2025 18:39		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP102125AR1254	PP075919.D	22 Oct 2025 08:49		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP102125AR1268	PP075920.D	22 Oct 2025 09:07		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PP075921.D	22 Oct 2025 09:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110625

Review By	rahul	Review On	11/7/2025 10:43:29 AM
Supervise By	yogesh	Supervise On	11/7/2025 10:54:03 AM
SubDirectory	PP110625	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076342.D	06 Nov 2025 08:56		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076343.D	06 Nov 2025 09:30		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP076344.D	06 Nov 2025 09:51		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076345.D	06 Nov 2025 10:07		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076346.D	06 Nov 2025 10:23		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP076347.D	06 Nov 2025 10:40		YP\AJ	Ok
7	Q3546-01	TP-COMP	PP076348.D	06 Nov 2025 12:08		YP\AJ	Ok,M
8	Q3546-01MS	TP-COMPMS	PP076349.D	06 Nov 2025 12:24		YP\AJ	Ok,M
9	Q3546-01MSD	TP-COMPMSD	PP076350.D	06 Nov 2025 12:40		YP\AJ	Ok,M
10	Q3549-02	RT4922	PP076351.D	06 Nov 2025 12:56		YP\AJ	Ok
11	AR1660CCC500	AR1660CCC500	PP076352.D	06 Nov 2025 14:02		YP\AJ	Ok,M
12	AR1242CCC500	AR1242CCC500	PP076353.D	06 Nov 2025 14:18		YP\AJ	Ok,M
13	AR1248CCC500	AR1248CCC500	PP076354.D	06 Nov 2025 14:34		YP\AJ	Ok,M
14	AR1254CCC500	AR1254CCC500	PP076355.D	06 Nov 2025 14:50	CCC fail on lower side	YP\AJ	Not Ok
15	I.BLK	I.BLK	PP076356.D	06 Nov 2025 15:07		YP\AJ	Ok
16	Q3534-09	FB-1	PP076357.D	06 Nov 2025 15:23		YP\AJ	Ok,M
17	Q3534-10	EB-1	PP076358.D	06 Nov 2025 15:39		YP\AJ	Ok,M
18	Q3537-03	MW-17-PURGE-WATE	PP076359.D	06 Nov 2025 15:55		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110625

Review By	rahul	Review On	11/7/2025 10:43:29 AM
Supervise By	yogesh	Supervise On	11/7/2025 10:54:03 AM
SubDirectory	PP110625	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3541-01	N48969	PP076360.D	06 Nov 2025 16:12		YPIAJ	Ok,M
20	Q3542-01	AUD-25-182-194	PP076361.D	06 Nov 2025 16:28		YPIAJ	Ok,M
21	Q3552-01	MW-4	PP076362.D	06 Nov 2025 16:44		YPIAJ	Ok,M
22	Q3552-02	MW-8	PP076363.D	06 Nov 2025 17:00		YPIAJ	Ok
23	Q3552-03	MW-9	PP076364.D	06 Nov 2025 17:17		YPIAJ	Ok,M
24	Q3552-04	MW-10	PP076365.D	06 Nov 2025 17:33		YPIAJ	Ok,M
25	Q3552-05	MW-3	PP076366.D	06 Nov 2025 17:49		YPIAJ	Ok,M
26	AR1660CCC500	AR1660CCC500	PP076367.D	06 Nov 2025 18:54		YPIAJ	Ok,M
27	AR1242CCC500	AR1242CCC500	PP076368.D	06 Nov 2025 19:43		YPIAJ	Ok,M
28	AR1248CCC500	AR1248CCC500	PP076369.D	06 Nov 2025 19:59		YPIAJ	Ok,M
29	AR1254CCC500	AR1254CCC500	PP076370.D	06 Nov 2025 20:15		YPIAJ	Ok,M
30	I.BLK	I.BLK	PP076371.D	06 Nov 2025 20:48		YPIAJ	Ok

M : Manual Integration

SOP ID:	M3510C,3580A-Extraction PCB-15		
Clean Up SOP #:	Acid Cleanup	Extraction Start Date :	11/06/2025
Matrix :	Water	Extraction Start Time :	09:15
Weigh By:	N/A	Extraction By:	RS
Balance check:	N/A	Extraction End Date :	11/06/2025
Balance ID:	N/A	Filter By:	RS
pH Strip Lot#:	E3953	Extraction End Time :	13:55
		pH Meter ID:	N/A
		Concentration By:	EH
		Hood ID:	4,6,7
		Supervisor By :	RUPESH

Extraction Method:
☒ Separatory Funnel
 ☐ Continuous Liquid/Liquid
 ☐ Sonication
 ☐ Waste Dilution
 ☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3980
Baked Na2SO4	N/A	EP2655
Hexane	N/A	E3981
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot# 03-40BTS721. Q3534 & Q3552 all samples split the volume.

KD Bath ID:	WATER BATH-1	Envap ID:	NEVAP-02
KD Bath Temperature:	60 °C	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/6/25	RS (Ext-Lab)	J-P. Pestil PCB
14:00	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction PCB-15

Concentration Date: 11/06/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170425BL	ABLK425	PCB	1000	6	RUPESH	ritesh	5			SEP-1
PB170425BS	ALCS425	PCB	1000	6	RUPESH	ritesh	5			2
PB170425BS D	ALCSD425	PCB	1000	6	RUPESH	ritesh	5			3
Q3534-01	MW-6	PCB	450	6	RUPESH	ritesh	5	O		4
Q3534-02	MW-1	PCB	470	6	RUPESH	ritesh	5	O		5
Q3534-03	MW-11	PCB	490	6	RUPESH	ritesh	5	O		6
Q3534-04	MW-5	PCB	485	6	RUPESH	ritesh	5	O		7
Q3534-05	MW-EB-1	PCB	475	6	RUPESH	ritesh	5	O		8
Q3534-06	MW-EB-2	PCB	485	6	RUPESH	ritesh	5	O		9
Q3534-07	MW-EB-3	PCB	480	6	RUPESH	ritesh	5	O		10
Q3534-09	FB-1	PCB	495	6	RUPESH	ritesh	5	H		11
Q3534-10	EB-1	PCB	495	6	RUPESH	ritesh	5	H		12
Q3537-03	MW-17-PURGE-WATER	PCB	990	6	RUPESH	ritesh	10	P		13
Q3541-01	N48969	PCB	980	6	RUPESH	ritesh	10	P		14
Q3542-01	AUD-25-182-194	PCB	500	6	RUPESH	ritesh	10	C		15
Q3552-01	MW-4	PCB	500	6	RUPESH	ritesh	5	L		16
Q3552-02	MW-8	PCB	500	6	RUPESH	ritesh	5	L		SEP-1
Q3552-03	MW-9	PCB	500	6	RUPESH	ritesh	5	L		2
Q3552-04	MW-10	PCB	500	6	RUPESH	ritesh	5	L		3
Q3552-05	MW-3	PCB	485	6	RUPESH	ritesh	5	L		4

170425
9:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3542

WorkList ID : 192911

Department : Extraction

Date : 11-06-2025 08:30:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3534-01	MW-6	Water	PCB	Cool 4 deg C	REMI02	D31	11/03/2025	8082A
Q3534-02	MW-1	Water	PCB	Cool 4 deg C	REMI02	D31	11/03/2025	8082A
Q3534-03	MW-11	Water	PCB	Cool 4 deg C	REMI02	D31	11/03/2025	8082A
Q3534-04	MW-5	Water	PCB	Cool 4 deg C	REMI02	D31	11/04/2025	8082A
Q3534-05	MW-EB-1	Water	PCB	Cool 4 deg C	REMI02	D31	11/03/2025	8082A
Q3534-06	MW-EB-2	Water	PCB	Cool 4 deg C	REMI02	D31	11/04/2025	8082A
Q3534-07	MW-EB-3	Water	PCB	Cool 4 deg C	REMI02	D31	11/03/2025	8082A
Q3534-09	FB-1	Water	PCB	Cool 4 deg C	REMI02	D31	11/04/2025	8082A
Q3534-10	EB-1	Water	PCB	Cool 4 deg C	REMI02	D31	11/04/2025	8082A
Q3537-03	MW-17-PURGE-WATER	Water	PCB	Cool 4 deg C	PSEG03	D41	11/04/2025	8082A
Q3541-01	N48969	Water	PCB	Cool 4 deg C	PSEG03	D41	11/04/2025	8082A
Q3542-01	AUD-25-182-194	Water	PCB	Cool 4 deg C	PSEG03	D41	11/04/2025	8082A
Q3552-01	MW-4	Water	PCB	Cool 4 deg C	REMI02	D41	11/04/2025	8082A
Q3552-02	MW-8	Water	PCB	Cool 4 deg C	REMI02	D41	11/04/2025	8082A
Q3552-03	MW-9	Water	PCB	Cool 4 deg C	REMI02	D41	11/04/2025	8082A
Q3552-04	MW-10	Water	PCB	Cool 4 deg C	REMI02	D41	11/05/2025	8082A
Q3552-05	MW-3	Water	PCB	Cool 4 deg C	REMI02	D41	11/04/2025	8082A

Date/Time 11/6/25 9:10
Raw Sample Received by: RS (Ext-96)
Raw Sample Relinquished by: RS (Ext-96)

Date/Time 11/6/25 10:00
Raw Sample Received by: RS (Ext-96)
Raw Sample Relinquished by: RS (Ext-96)

Prep Standard - Chemical Standard Summary

Order ID : Q3552

Test : PCB

Prepbatch ID : PB170425,

Sequence ID/Qc Batch ID: po110625,pp110625,

Standard ID :

EP2610,EP2655,PP24329,PP24330,PP24381,PP24663,PP24805,PP24806,PP24807,PP24808,PP24809,PP24810,PP24811,PP24812,PP24814,PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862,PP24930,

Chemical ID :

E3804,E3875,E3944,E3962,E3963,E3972,E3980,E3981,M6157,P 11522,P11592,P12700,P12703,P12952,P12953,P12957,P13356,P13373,P13375,P13590,P13592,P13695,P13697,P13706,P13786,P13789,P13833,P13879,P13881,P13885,W3112,W3177,W3234,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2655	10/24/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/24/2025
<u>FROM</u> 4000.00000gram of E3875 = Final Quantity: 4000.000 gram								

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel 07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24805	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13789 + 9.00000ml of E3962 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24806	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13375 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24807	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13879 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24808	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13706 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24809	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13695 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24810	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P11592 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24811	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13881 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24812	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13885 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24814	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12952 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24815	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 98.50000ml of E3962 + 0.50000ml of PP24805 + 1.00000ml of PP24814 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24816	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13373 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24817	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P12700 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24818	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13590 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24819	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13592 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2193	AR1254 100 PPM STOCK SOLUTION	PP24820	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13833 + 9.00000ml of E3963 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24821	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P12703 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24822	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P11522 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24823	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24824	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24825	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24826	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24824 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24827	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24806 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24828	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24806 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24829	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24806 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24830	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24828 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24831	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24807 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24832	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24807 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24833	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24807 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24834	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24832 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24835	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24808 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24836	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24808 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24837	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24808 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24838	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24836 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24839	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24809 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24840	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24809 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24841	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24809 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24842	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24840 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24843	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24810 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24844	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24810 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24845	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24810 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24846	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24844 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24847	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24811 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24848	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24811 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24849	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24811 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24850	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24848 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24851	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24812 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24852	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24812 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24853	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24812 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24854	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24852 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24855	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24815 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3790	AR1221 500 PPB ICV(AGILENT)	PP24856	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24816 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24857	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24817 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24858	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24818 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24859	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24819 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24860	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24861	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24821 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24862	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24822 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24930	09/18/2025	03/18/2026	Abdul Mirza	None	None	Yogesh Patel
								09/19/2025

FROM 0.50000ml of P12953 + 99.50000ml of E3972 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	02/18/2026	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	07/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	08/05/2025 / RUPESH	07/30/2025 / RUPESH	E3962

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	02/15/2026	08/15/2025 / RUPESH	08/07/2025 / RUPESH	E3963

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	10/10/2025 / RUPESH	10/10/2025 / RUPESH	E3980

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/30/2026	10/27/2025 / RUPESH	10/22/2025 / RUPESH	E3981

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0175403	02/18/2026	08/18/2025 / yogesh	03/18/2022 / Abdul	P11592

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	91867 / Aroclor 1232 100 ug/mL	020823	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12700

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12703

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	02/18/2026	08/18/2025 / yogesh	12/20/2023 / Yogesh	P12952

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	03/18/2026	09/18/2025 / Abdul	12/20/2023 / Yogesh	P12953

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0207511	02/18/2026	08/18/2025 / yogesh	05/03/2024 / Abdul	P13375

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13590

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13592

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	A0214733	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13695

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	A0210232	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13706

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	02/18/2026	08/18/2025 / yogesh	11/19/2024 / Ankita	P13789

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	a0217391	02/18/2026	08/18/2025 / yogesh	12/09/2024 / Ankita	P13833

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13879

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0217041	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13881

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0217264	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13885

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/28/2025 / jignesh	07/25/2025 / jignesh	W3234

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/30/25

E3962

Jamie Croak
Director Quality Operations, Bioscience Production

Acetone.
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 8/6/25

E 3963

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 25C1262005
Manufactured Date: 2025-01-15
Expiration Date: 2026-04-16
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

REC. on 10/10/25
RJ

E3980

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™



Material No.: 9262-03

Batch No.: 25C0362006

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	4
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Received on 10/22/25

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3981

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



MG157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantor**™



Material No.: 9673-33

Batch No.: 2411262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 Lot No.: A0175403
Description : Aroclor® 1254 Standard
Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : November 30, 2027 Storage: 25°C nominal
Handling: This product contains PCBs. Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Aroclor 1254 CAS # 11097-69-1 (Lot 124-191-B) Purity ----%	1,000.7 µg/mL	+/- 5.9437 µg/mL Gravimetric +/- 31.7284 µg/mL Unstressed +/- 41.4406 µg/mL Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P11588
↓
P11592 / (S)

AR
04/30/2022

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

300°C

Det. Type:

ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Cathleen Soltis

Cathleen Soltis - Mix Technician

Date Mixed: 15-Aug-2021

Balance: 1128360905

Alexis Shelow

Alexis Shelow - Operations Tech I

Date Passed: 17-Aug-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P11588
↓
P11592 / (5)

AR
04/30/22



CERTIFIED WEIGHT REPORT

Part Number: **91867**
 Lot Number: **020823**
 Description: **WP 037 - Aroclor 1232**
PCB Technical Mixture
 Expiration Date: **020833**
 Recommended Storage: **Ambient (20 °C)**
 Nominal Concentration ($\mu\text{g/mL}$): **100**
 NIST Test ID#: **6UTB**

Solvent(
Aceton


Weight(s) shown below were combined and diluted to (mL): **100.0**
 5E-05 Balance Uncertainty
 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc ($\mu\text{g/mL}$)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
 Column ID SPB-608 30 meter X 0.53mm X 5 μm film thickness
 Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
 Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
 Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
 Rate = 8°C/min, Total run time = 35 min
 Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
 Standard injection = 1.5 μL , Range=3





CERTIFIED WEIGHT REPORT

Part Number: X9166

Lot Number: 060523

Description: Aroclor 1262

Solvent(s): Methanol

E

Appendix IX Compound

Expiration Date: 060533

Recommended Storage: Ambient (20 °C)

Nominal Concentration (µg/mL): 100

NIST Test ID#: 6UTB


Volume(s) shown below were combined and diluted to (mL): 20.0

5E-05 Balance Uncertainty

0.002 Flask Uncertainty

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Co
1. Aroclor 1262	70444	080322	0.10	2.00	0.017	

Method GC7MSD-7.M: Column:(30m X 0.25mm ID X 0.25µm film thickn min.), Rate = 8°C/min., Injector B= 200°C, Detector B = 290°C.



- The certified value is the concentration calculated from gravimetric and volumetric me
- Standards are prepared gravimetrically using balances that are calibrated with weight
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appro
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty


Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL.):
200.0

Balance Uncertainty
0.010

Flask Uncertainty
0.010

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL.):
200.0


Balance Uncertainty
0.010

Flask Uncertainty
0.010

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty (%)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Volume(s) shown below were combined and diluted to (mL): 20.0
Note: Aroclor 1254 is a mix of isomers.

Solvent(s): Iso-octane
Lot#: 82227

Formulated By: Anthony Mahoney	121823	DATE
Reviewed By: Pedro L. Rentas	121823	DATE


P12956 / Y.P.
12/18/23
P12957

Compound	Part Number	Lot Number	Dilution Factor	Initial Volume (mL)	Initial Uncertainty	Pipette (mL)	Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information		
											(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA) LD50
1. Aroclor 1254	79100	121823	0.10	2.00	0.017	0.017	1003.3	100.1	11097-69-1	1.8	0.5mg/m3 (skin)		or-rat 1295mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 280 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = EDAQ Channel 1
Standard injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0207511

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	13192500	----%	1,002.7 µg/mL	+/- 55.6424

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13374
↓
P 13375
/

[Signature]
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 06-Feb-2024

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC


Date Passed: 08-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13374 } ②
↓
P 13375 }


05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 95T215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1

ISO 17025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32010 **Lot No.:** A0214733
Description : Aroclor® 1248 Standard
Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : November 30, 2030 **Storage:** 25°C nominal
Handling: This product contains PCBs. **Ship:** Ambient

P13695 } Y.P.
P13696 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	15242800	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Rebecca Gingerich - Operations Tech II

Date Mixed: 02-Aug-2024

Balance Serial # C322230531

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32009 **Lot No.:** A0210232

Description : Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13706 } y.p.
P13707 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	----%	1,000.7 µg/mL	+/- 55.5295

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024

Balance Serial # B442140311


Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 17-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent:

Acetone

CAS # 67-64-1

Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
↓
P13834
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024

Balance Serial #

C322230531

Brittany Federhik

Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32409 **Lot No.:** A0217041

Description : Aroclor® 1262 Standard

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	-----%	1,000.0 µg/mL	+/- 55.4925

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13881
AJ
01128125

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Rebecca Gingerich - Operations Tech II


Jennifer Pollino - Operations Tech III - ARM QC

Date Mixed: 25-Sep-2024 Balance Serial # C322230531

Date Passed: 30-Sep-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410

Lot No.: A0217264

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : January 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	12353400	----%	1,009.8 µg/mL	+/- 56.0364

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13884

+

AJ
01/28/25

P13886

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 01-Oct-2024

Balance Serial #

1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 02-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantor**™



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

W3234 JB
OP4H1. 07/28/2025

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:
 COMPANY: Remington Vernick Engineers
 ADDRESS: 2059 Springdale Road
 CITY: Cherry Hill STATE: NJ ZIP: 08003
 ATTENTION: Kyle Carison @ NE.com
 PHONE: 609-682-3049 FAX:

PROJECT NAME: Edison Landfill
 PROJECT NO.: 1205T007 LOCATION: Edison NJ
 PROJECT MANAGER: Leslie Hartzell
 e-mail: Leslie.Hartzell@NE.com
 PHONE: 610-663-4665 FAX:

BILL TO: APink@ONE.com PO#: 1205T007
 ADDRESS: 2059 Springdale Road
 CITY: Cherry Hill STATE: NJ ZIP: 08003
 ATTENTION: Leslie Hartzell PHONE: 610-663-4665

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) _____ DAYS*
 HARDCOPY (DATA PACKAGE): STAT _____ DAYS*
 EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
 + Raw Data ☐ Other
☒ EDD FORMAT: NJ HAT

TEL: 610-663-4665
 2. VOCs
 3. TML/Heavy Metals
 4. Cyanide
 5. COD
 6. TKN
 7. Oil & Grease
 8. Hexachrome
 9.

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES								COMMENTS
			COMP	GRAB	DATE	TIME		E	A	B	D	C	C	C	F	
1.	MW-4	GW		P	11/4/25	13:40	15	7	2	1	1	1	1	1	1	
2.	MW-8	GW		P	11/4/25	13:45	15	7	2	1	1	1	1	1	1	
3.	MW-9	GW		P	11/4/25	16:00	15	7	2	1	1	1	1	1	1	
4.	MW-10	GW		P	11/5/25	12:00	15	7	2	1	1	1	1	1	1	
5.	MW-3	GW		P	11/4/25	16:15	15	7	2	1	1	1	1	1	1	
6.	TB-1	GW		P	11/5/25	8:00	2	2	2	~~XXXXXXXXXX~~						trip blank
7.																
8.																
9.																
10.																

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. Anna Nayer	DATE/TIME: 11/5/25 (13:30)	RECEIVED BY: JT	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>3.8 + 0 = 3.8</u>
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY:	Comments:
RELINQUISHED BY SAMPLER: 3. JT	DATE/TIME: 11/5/25	RECEIVED BY:	

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3552	REMI02	Order Date : 11/5/2025 2:23:00 PM	Project Mgr :
Client Name : Remington & Vernick Engi		Project Name : Edison Landfill	Report Type : NJ Reduced
Client Contact : Kyle Carlson		Receive DateTime : 11/5/2025 ~~12:00:00 AM~~ 2:45 PM	EDD Type : Excel NJ
Invoice Name : Remington & Vernick Engi		Purchase Order :	Hard Copy Date :
Invoice Contact : Kyle Carlson			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3552-01	MW-4	Water	11/04/2025	13:40	VOC-TCLVOA-10	TCL+30/TAL	8260-Low	10 Bus. Days	
Q3552-02	MW-8	Water	11/04/2025	13:45	VOC-TCLVOA-10	TCL+30/TAL	8260-Low	10 Bus. Days	
Q3552-03	MW-9	Water	11/04/2025	16:00	VOC-TCLVOA-10	TCL+30/TAL	8260-Low	10 Bus. Days	
Q3552-04	MW-10	Water	11/05/2025	12:00	VOC-TCLVOA-10	TCL+30/TAL	8260-Low	10 Bus. Days	
Q3552-05	MW-3	Water	11/04/2025	16:15	VOC-TCLVOA-10	TCL+30/TAL	8260-Low	10 Bus. Days	
Q3552-06	TB	Water	11/05/2025	08:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3552	REMI02	Order Date : 11/5/2025 2:23:00 PM	Project Mgr :
Client Name : Remington & Vernick Engi		Project Name : Edison Landfill	Report Type : NJ Reduced
Client Contact : Kyle Carlson		Receive DateTime : 11/5/2025 ~~12:00:00 AM~~ 2:45 PM	EDD Type : Excel NJ
Invoice Name : Remington & Vernick Engi		Purchase Order :	Hard Copy Date :
Invoice Contact : Kyle Carlson			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

*Stored in VOA
FRZ # 04*

Relinquished By : *al*
Date / Time : 11/5/25 15/5

Received By : *Wm Duda*
Date / Time : 11/5/25 15:20

Storage Area : VOA Refridgerator Room