

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3552

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD091016.D Date Analyzed: 11/07/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:56

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.063	8.960	9.160	22.710	20.000	13.6
Tetrachloro-m-xylene	3.544	3.490	3.590	22.750	20.000	13.8
alpha-BHC	3.992	3.940	4.040	9.790	10.000	-2.1
beta-BHC	4.511	4.460	4.560	11.070	10.000	10.7
gamma-BHC (Lindane)	4.323	4.270	4.370	9.990	10.000	-0.1
Endrin	6.566	6.500	6.640	51.180	50.000	2.4
4,4'-DDT	7.014	6.940	7.080	108.650	100.000	8.7
Methoxychlor	7.487	7.420	7.560	255.870	250.000	2.3

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD091016.D Date Analyzed: 11/07/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:56

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.058	7.960	8.160	25.950	20.000	29.8
Tetrachloro-m-xylene	2.873	2.820	2.920	24.210	20.000	21.1
alpha-BHC	3.385	3.330	3.440	11.810	10.000	18.1
beta-BHC	4.017	3.970	4.070	12.100	10.000	21.0
gamma-BHC (Lindane)	3.721	3.670	3.770	11.830	10.000	18.3
Endrin	5.778	5.710	5.850	52.540	50.000	5.1
4,4'-DDT	6.171	6.100	6.240	110.850	100.000	10.9
Methoxychlor	6.741	6.670	6.810	231.280	250.000	-7.5