

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3552

Analytical Method: 8081B

Client: Remington & Vernick Engineers

Datafile : PD091023.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170426BSD (Column 1)	alpha-BHC	0.5	0.43	ug/L	87	0			40 (85)	140 (130)	20 (20)
	beta-BHC	0.5	0.42	ug/L	83	1			40 (83)	140 (126)	20 (20)
	delta-BHC	0.5	0.43	ug/L	87	1			40 (69)	140 (141)	20 (20)
	gamma-BHC (Lindane)	0.5	0.42	ug/L	85	1			40 (82)	140 (129)	20 (20)
	Heptachlor	0.5	0.52	ug/L	105	1			40 (79)	140 (127)	20 (20)
	Aldrin	0.5	0.42	ug/L	85	2			40 (79)	140 (126)	20 (20)
	Heptachlor epoxide	0.5	0.45	ug/L	89	2			40 (81)	140 (124)	20 (20)
	Endosulfan I	0.5	0.43	ug/L	87	1			40 (85)	140 (122)	20 (20)
	Dieldrin	0.5	0.44	ug/L	87	1			40 (83)	140 (125)	20 (20)
	4,4'-DDE	0.5	0.42	ug/L	84	1			40 (80)	140 (127)	20 (20)
	Endrin	0.5	0.43	ug/L	87	1			40 (81)	140 (128)	20 (20)
	Endosulfan II	0.5	0.44	ug/L	89	0			40 (82)	140 (123)	20 (20)
	4,4'-DDD	0.5	0.46	ug/L	92	0			40 (77)	140 (131)	20 (20)
	Endosulfan sulfate	0.5	0.44	ug/L	87	1			40 (76)	140 (129)	20 (20)
	4,4'-DDT	0.5	0.47	ug/L	95	0			40 (80)	140 (133)	20 (20)
	Methoxychlor	0.5	0.49	ug/L	97	1			40 (78)	140 (108)	20 (20)
	Endrin ketone	0.5	0.46	ug/L	92	2			40 (80)	140 (131)	20 (20)
	Endrin aldehyde	0.5	0.44	ug/L	88	1			40 (82)	140 (127)	20 (20)
	alpha-Chlordane	0.5	0.42	ug/L	83	0			40 (82)	140 (125)	20 (20)
	gamma-Chlordane	0.5	0.43	ug/L	86	1			40 (82)	140 (125)	20 (20)
	alpha-BHC	0.5	0.45	ug/L	89	2			40 (85)	140 (130)	20 (20)
PB170426BSD (Column 2)	beta-BHC	0.5	0.43	ug/L	87	2			40 (83)	140 (126)	20 (20)
	delta-BHC	0.5	0.44	ug/L	87	4			40 (69)	140 (141)	20 (20)
	gamma-BHC (Lindane)	0.5	0.44	ug/L	88	3			40 (82)	140 (129)	20 (20)
	Heptachlor	0.5	0.46	ug/L	92	3			40 (79)	140 (127)	20 (20)
	Aldrin	0.5	0.44	ug/L	88	3			40 (79)	140 (126)	20 (20)
	Heptachlor epoxide	0.5	0.43	ug/L	86	3			40 (81)	140 (124)	20 (20)
	Endosulfan I	0.5	0.40	ug/L	80	4			40 (85)	140 (122)	20 (20)
	Dieldrin	0.5	0.44	ug/L	88	4			40 (83)	140 (125)	20 (20)
	4,4'-DDE	0.5	0.45	ug/L	89	1			40 (80)	140 (127)	20 (20)
	Endrin	0.5	0.46	ug/L	93	4			40 (81)	140 (128)	20 (20)
	Endosulfan II	0.5	0.44	ug/L	89	4			40 (82)	140 (123)	20 (20)
	4,4'-DDD	0.5	0.45	ug/L	91	4			40 (77)	140 (131)	20 (20)
	Endosulfan sulfate	0.5	0.45	ug/L	90	4			40 (76)	140 (129)	20 (20)
	4,4'-DDT	0.5	0.47	ug/L	93	3			40 (80)	140 (133)	20 (20)
	Methoxychlor	0.5	0.46	ug/L	93	4			40 (78)	140 (108)	20 (20)
	Endrin ketone	0.5	0.56	ug/L	111	3			40 (80)	140 (131)	20 (20)
	Endrin aldehyde	0.5	0.46	ug/L	92	4			40 (82)	140 (127)	20 (20)
	alpha-Chlordane	0.5	0.43	ug/L	87	3			40 (82)	140 (125)	20 (20)
	gamma-Chlordane	0.5	0.43	ug/L	87	4			40 (82)	140 (125)	20 (20)