

Surrogate Summary

SDG No.: Q3552

Client: Remington & Vernick Engineers

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	19.4	97		30 (41)	150 (134)
		Decachlorobiphen	2	20	20.3	102		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	20.0	100		30 (41)	150 (134)
I.BLK-PD090991.D	PIBLK-PD090991.D	Decachlorobiphen	1	20	20.9	105		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	20.5	102		30 (41)	150 (134)
		Decachlorobiphen	2	20	23.1	116		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	21.3	106		30 (41)	150 (134)
PB170426BL	PB170426BL	Decachlorobiphen	1	20	17.3	86		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	16.8	84		30 (41)	150 (134)
		Decachlorobiphen	2	20	18.9	94		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	17.6	88		30 (41)	150 (134)
PB170426BS	PB170426BS	Decachlorobiphen	1	20	24.7	124		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	24.4	122		30 (41)	150 (134)
		Decachlorobiphen	2	20	28.2	141		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	25.8	129		30 (41)	150 (134)
PB170426BSD	PB170426BSD	Decachlorobiphen	1	20	25.0	125		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	23.6	118		30 (41)	150 (134)
		Decachlorobiphen	2	20	27.5	137		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	25.3	127		30 (41)	150 (134)
Q3552-01	MW-4	Decachlorobiphen	1	20	11.5	58		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	26.4	132		30 (41)	150 (134)
		Decachlorobiphen	2	20	30.3	151	*	30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	12.1	61		30 (41)	150 (134)
Q3552-02	MW-8	Decachlorobiphen	1	20	19.8	99		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	19.7	99		30 (41)	150 (134)
		Decachlorobiphen	2	20	21.4	107		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	19.5	98		30 (41)	150 (134)
Q3552-03	MW-9	Decachlorobiphen	1	20	12.6	63		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	18.7	94		30 (41)	150 (134)
		Decachlorobiphen	2	20	10.3	51		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	15.4	77		30 (41)	150 (134)
Q3552-04	MW-10	Decachlorobiphen	1	20	18.6	93		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	17.8	89		30 (41)	150 (134)
		Decachlorobiphen	2	20	18.8	94		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	19.5	97		30 (41)	150 (134)
Q3552-05	MW-3	Decachlorobiphen	1	20	14.4	72		30 (31)	150 (149)
		Tetrachloro-m-xyl	1	20	16.9	85		30 (41)	150 (134)
		Decachlorobiphen	2	20	20.4	102		30 (31)	150 (149)
		Tetrachloro-m-xyl	2	20	15.2	76		30 (41)	150 (134)
I.BLK-PD091012.D	PIBLK-PD091012.D	Decachlorobiphen	1	20	20.9	104		30 (31)	150 (149)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q3552

Client: Remington & Vernick Engineers

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD091012.D	PIBLK-PD091012.D	Tetrachloro-m-xyI	1	20	20.3	101		30 (41)	150 (134)
		Decachlorobiphen	2	20	23.1	116		30 (31)	150 (149)
		Tetrachloro-m-xyI	2	20	21.4	107		30 (41)	150 (134)