

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

|                 |                   |                        |                              |
|-----------------|-------------------|------------------------|------------------------------|
| Lab Name:       | <u>Alliance</u>   | Contract:              | <u>REMI02</u>                |
| Lab Code:       | <u>ACE</u>        | SDG No.:               | <u>Q3552</u>                 |
| Instrument ID:  | <u>BNA_N</u>      | Calibration Date/Time: | <u>11/07/2025 09:39</u>      |
| Lab File ID:    | <u>BN038147.D</u> | Init. Calib. Date(s):  | <u>10/28/2025 10/28/2025</u> |
| EPA Sample No.: | <u>SSTDCCC0.4</u> | Init. Calib. Time(s):  | <u>13:00 16:38</u>           |
| GC Column:      | <u>ZB-GR</u>      | ID:                    | <u>0.25</u> (mm)             |

| COMPOUND                | RRF   | RRF0.4 | MIN<br>RRF | %D    | MAX%D |
|-------------------------|-------|--------|------------|-------|-------|
| 2-Methylnaphthalene-d10 | 0.571 | 0.545  |            | -4.6  | 20.0  |
| Fluoranthene-d10        | 1.008 | 0.947  |            | -6.1  | 20.0  |
| 2-Fluorophenol          | 0.943 | 0.847  |            | -10.2 | 20.0  |
| Phenol-d6               | 1.148 | 1.036  |            | -9.8  | 20.0  |
| Nitrobenzene-d5         | 0.323 | 0.310  |            | -4.0  | 20.0  |
| 2-Fluorobiphenyl        | 1.602 | 1.557  |            | -2.8  | 20.0  |
| 2,4,6-Tribromophenol    | 0.165 | 0.133  |            | -19.4 | 20.0  |
| Terphenyl-d14           | 0.869 | 0.848  |            | -2.4  | 20.0  |
| 1,4-Dioxane             | 0.416 | 0.407  |            | -2.2  | 20.0  |

All other compounds must meet a minimum RRF of 0.010.