

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102825

Review By	rahul	Review On	10/28/2025 3:41:07 PM
Supervise By	mohammad	Supervise On	10/30/2025 5:29:45 AM
SubDirectory	BN102825	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038110.D	28 Oct 2025 12:05		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN038111.D	28 Oct 2025 13:00		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN038112.D	28 Oct 2025 13:36		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN038113.D	28 Oct 2025 14:13	compound #20 kept on LR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN038114.D	28 Oct 2025 14:49		RC/JU	Ok,M
6	SSTDICC1.6	SSTDICC1.6	BN038115.D	28 Oct 2025 15:25		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN038116.D	28 Oct 2025 16:02		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN038117.D	28 Oct 2025 16:38	Compound #20 removed from 5PPM	RC/JU	Ok
9	SSTDICV0.4	ICVBN102825	BN038118.D	28 Oct 2025 17:51		RC/JU	Ok
10	PB170167BL	PB170167BL	BN038119.D	28 Oct 2025 19:04	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN110725

Review By	rahul	Review On	11/7/2025 2:34:23 PM
Supervise By	Jagrut	Supervise On	11/10/2025 12:20:28 PM
SubDirectory	BN110725	HP Acquire Method	BNA_N
		HP Processing Method	BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038146.D	07 Nov 2025 09:00		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038147.D	07 Nov 2025 09:39		RC/JU	Ok
3	PB170427BL	PB170427BL	BN038148.D	07 Nov 2025 10:15	Analyzed for contamination check	RC/JU	Not Ok
4	Q3534-01	MW-6	BN038149.D	07 Nov 2025 10:52	Internal standard failed,,Need 10X Dilution	RC/JU	Dilution
5	Q3534-02	MW-1	BN038150.D	07 Nov 2025 11:27	Need 30X Dilution	RC/JU	Dilution
6	Q3534-03	MW-11	BN038151.D	07 Nov 2025 12:04	Need 2X Dilution	RC/JU	Dilution
7	PB170427BS	PB170427BS	BN038152.D	07 Nov 2025 13:44		RC/JU	Ok
8	PB170427BSD	PB170427BSD	BN038153.D	07 Nov 2025 14:20		RC/JU	Ok
9	Q3552-03	MW-9	BN038154.D	07 Nov 2025 14:56	Internal standard failed,Need 10X Dilution	RC/JU	Dilution
10	Q3534-05	MW-EB-1	BN038155.D	07 Nov 2025 15:33	Internal standard failed,Need 2X Dilution	RC/JU	Dilution
11	Q3552-05	MW-3	BN038156.D	07 Nov 2025 16:10	Internal standard failed,Need 10X Dilution	RC/JU	Dilution
12	Q3534-04	MW-5	BN038157.D	07 Nov 2025 16:46		RC/JU	Ok
13	Q3534-06	MW-EB-2	BN038158.D	07 Nov 2025 17:21		RC/JU	Ok
14	Q3534-07	MW-EB-3	BN038159.D	07 Nov 2025 17:58		RC/JU	Ok
15	Q3552-04	MW-10	BN038160.D	07 Nov 2025 18:34		RC/JU	Ok
16	Q3552-01	MW-4	BN038161.D	07 Nov 2025 19:10	Need 40X Dilution	RC/JU	Dilution
17	Q3552-02	MW-8	BN038162.D	07 Nov 2025 19:46	Need 5X Dilution	RC/JU	Dilution

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN110725

Review By	rahul	Review On	11/7/2025 2:34:23 PM		
Supervise By	Jagrut	Supervise On	11/10/2025 12:20:28 PM		
SubDirectory	BN110725	HP Acquire Method	BNA_N	HP Processing Method	BN102825
STD. NAME	STD REF.#				
Tune/Reschk	SP6875				
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904				
CCC	SP6902				
Internal Standard/PEM	SP6830,1ul/100ul sample				
ICV/I.BLK	SP6906				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	SSTDCCC0.4	SSTDCCC0.4EC	BN038163.D	07 Nov 2025 20:22		RC/JU	Ok,M
19	DFTPP	DFTPP	BN038164.D	07 Nov 2025 21:38		RC/JU	Ok
20	SSTDCCC0.4	SSTDCCC0.4	BN038165.D	07 Nov 2025 22:18		RC/JU	Ok
21	PB170427BL	PB170427BL	BN038166.D	07 Nov 2025 22:54		RC/JU	Ok
22	Q3534-09	FB-1	BN038167.D	07 Nov 2025 23:31		RC/JU	Ok
23	Q3534-10	EB-1	BN038168.D	08 Nov 2025 00:07		RC/JU	Ok,M
24	Q3534-01DL	MW-6DL	BN038169.D	08 Nov 2025 00:44		RC/JU	Ok,M
25	Q3534-02DL	MW-1DL	BN038170.D	08 Nov 2025 01:20		RC/JU	Ok
26	Q3534-03DL	MW-11DL	BN038171.D	08 Nov 2025 01:57		RC/JU	Ok
27	SSTDCCC0.4	SSTDCCC0.4EC	BN038172.D	08 Nov 2025 02:34		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN111025

Review By	rahul	Review On	11/11/2025 8:29:55 AM
Supervise By	Jagrut	Supervise On	11/11/2025 10:33:05 AM
SubDirectory	BN111025	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038173.D	10 Nov 2025 09:30		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038174.D	10 Nov 2025 10:14		RC/JU	Ok
3	PB170427BL	PB170427BL	BN038175.D	10 Nov 2025 10:50		RC/JU	Ok
4	Q3534-05DL	MW-EB-1DL	BN038176.D	10 Nov 2025 11:26		RC/JU	Ok
5	Q3552-01DL	MW-4DL	BN038177.D	10 Nov 2025 12:02		RC/JU	Ok
6	Q3552-05DL	MW-3DL	BN038178.D	10 Nov 2025 12:39		RC/JU	Ok
7	Q3552-03DL	MW-9DL	BN038179.D	10 Nov 2025 13:16		RC/JU	Ok
8	Q3552-02DL	MW-8DL	BN038180.D	10 Nov 2025 13:52		RC/JU	Ok
9	SSTDCCC0.4	SSTDCCC0.4EC	BN038181.D	10 Nov 2025 15:01		RC/JU	Ok

M : Manual Integration