

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3552

Analytical Method: 8270-Modified

Client: Remington & Vernick Engineers

DataFile: BN038152.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170427BS	1,4-Dioxane	0.4	0.32	ug/L	80				20 (65)	160 (116)	