

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038173.D
Instrument ID: BNA_N

Contract: REMI02
SDG NO.: Q3552
DFTPP Injection Date: 11/10/2025
DFTPP Injection Time: 09:30

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.7) 1
69	Mass 69 relative abundance	39.7
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	12.5
442	Greater than 50% of mass 198	83.2
443	15.0 - 24.0% of mass 442	16.3 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038174.D	11/10/2025	10:14
PB170427BL	PB170427BL	BN038175.D	11/10/2025	10:50
MW-4DL	Q3552-01DL	BN038177.D	11/10/2025	12:02
MW-3DL	Q3552-05DL	BN038178.D	11/10/2025	12:39
MW-9DL	Q3552-03DL	BN038179.D	11/10/2025	13:16
MW-8DL	Q3552-02DL	BN038180.D	11/10/2025	13:52