

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110625\
 Data File : BP026076.D
 Acq On : 06 Nov 2025 13:25
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 06 14:05:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	106	0.00
2	1,4-Dioxane	40.000	39.674	0.8	105	0.00
3	Pyridine	40.000	40.407	-1.0	105	0.00
4	n-Nitrosodimethylamine	40.000	38.321	4.2	101	0.00
5 S	2-Fluorophenol	80.000	83.617	-4.5	108	0.00
6	Aniline	40.000	41.175	-2.9	107	0.00
7 S	Phenol-d6	80.000	85.713	-7.1	109	0.00
8	2-Chlorophenol	40.000	44.113	-10.3	112	0.00
9	Benzaldehyde	40.000	40.347	-0.9	106	0.00
10 C	Phenol	40.000	41.994	-5.0	108	0.00
11	bis(2-Chloroethyl)ether	40.000	40.414	-1.0	106	-0.01
12	1,3-Dichlorobenzene	40.000	40.667	-1.7	106	0.00
13 C	1,4-Dichlorobenzene	40.000	40.395	-1.0	106	0.00
14	1,2-Dichlorobenzene	40.000	40.515	-1.3	106	0.00
15	Benzyl Alcohol	40.000	44.066	-10.2	114	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.148	4.6	99	0.00
17	2-Methylphenol	40.000	42.660	-6.6	108	0.00
18	Hexachloroethane	40.000	42.683	-6.7	110	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.460	-8.7	108	-0.01
20	3+4-Methylphenols	40.000	43.241	-8.1	111	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	40.276	-0.7	107	0.00
23 S	Nitrobenzene-d5	80.000	88.059	-10.1	114	0.00
24	Nitrobenzene	40.000	42.489	-6.2	110	0.00
25	Isophorone	40.000	42.702	-6.8	111	0.00
26 C	2-Nitrophenol	40.000	57.288	-43.2#	154	0.00
27	2,4-Dimethylphenol	40.000	41.178	-2.9	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.192	-3.0	109	0.00
29 C	2,4-Dichlorophenol	40.000	44.521	-11.3	113	0.01
30	1,2,4-Trichlorobenzene	40.000	40.751	-1.9	108	0.00
31	Naphthalene	40.000	40.013	-0.0	107	0.00
32	Benzoic acid	40.000	56.811	-42.0#	172	-0.01
33	4-Chloroaniline	40.000	41.880	-4.7	109	0.00
34 C	Hexachlorobutadiene	40.000	40.824	-2.1	108	0.00
35	Caprolactam	40.000	47.099	-17.7	122	-0.01
36 C	4-Chloro-3-methylphenol	40.000	45.979	-14.9	116	0.00
37	2-Methylnaphthalene	40.000	41.848	-4.6	109	0.01
38	1-Methylnaphthalene	40.000	41.569	-3.9	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.053	-0.1	112	0.01
41 P	Hexachlorocyclopentadiene	40.000	49.229	-23.1	135	0.00
42 S	2,4,6-Tribromophenol	80.000	99.438	-24.3	130	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.922	-17.3	123	0.01
44	2,4,5-Trichlorophenol	40.000	45.337	-13.3	119	0.02
45 S	2-Fluorobiphenyl	80.000	77.423	3.2	108	0.00
46	1,1'-Biphenyl	40.000	39.499	1.3	111	0.01
47	2-Chloronaphthalene	40.000	39.734	0.7	111	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110625\
 Data File : BP026076.D
 Acq On : 06 Nov 2025 13:25
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 06 14:05:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.750	-11.9	130	0.00
49	Acenaphthylene	40.000	40.331	-0.8	110	0.00
50	Dimethylphthalate	40.000	42.710	-6.8	117	0.00
51	2,6-Dinitrotoluene	40.000	45.582	-14.0	134	0.01
52 C	Acenaphthene	40.000	38.666	3.3	107	0.00
53	3-Nitroaniline	40.000	44.369	-10.9	128	0.01
54 P	2,4-Dinitrophenol	40.000	58.251	-45.6#	186	0.01
55	Dibenzofuran	40.000	39.688	0.8	111	0.00
56 P	4-Nitrophenol	40.000	44.118	-10.3	128	0.01
57	2,4-Dinitrotoluene	40.000	45.938	-14.8	133	0.02
58	Fluorene	40.000	39.967	0.1	112	0.01
59	2,3,4,6-Tetrachlorophenol	40.000	49.851	-24.6	129	0.00
60	Diethylphthalate	40.000	43.300	-8.2	117	0.00
61	4-Chlorophenyl-phenylether	40.000	40.144	-0.4	112	0.00
62	4-Nitroaniline	40.000	47.630	-19.1	124	0.00
63	Azobenzene	40.000	39.832	0.4	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	116	0.02
65	4,6-Dinitro-2-methylphenol	40.000	55.466	-38.7#	178	0.02
66 c	n-Nitrosodiphenylamine	40.000	40.304	-0.8	113	0.00
67	4-Bromophenyl-phenylether	40.000	42.144	-5.4	118	0.00
68	Hexachlorobenzene	40.000	41.596	-4.0	118	0.00
69	Atrazine	40.000	42.711	-6.8	119	0.00
70 C	Pentachlorophenol	40.000	72.285	-80.7#	201	0.01
71	Phenanthrene	40.000	39.502	1.2	112	0.01
72	Anthracene	40.000	40.254	-0.6	114	0.01
73	Carbazole	40.000	40.684	-1.7	114	0.01
74	Di-n-butylphthalate	40.000	45.338	-13.3	120	0.01
75 C	Fluoranthene	40.000	40.678	-1.7	113	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	114	0.02
77	Benzydine	40.000	45.504	-13.8	127	0.01
78	Pyrene	40.000	41.190	-3.0	111	0.01
79 S	Terphenyl-d14	80.000	81.258	-1.6	112	0.00
80	Butylbenzylphthalate	40.000	47.684	-19.2	138	0.00
81	Benzo(a)anthracene	40.000	40.402	-1.0	111	0.02
82	3,3'-Dichlorobenzidine	40.000	45.223	-13.1	123	0.01
83	Chrysene	40.000	39.940	0.2	110	0.02
84	Bis(2-ethylhexyl)phthalate	40.000	45.664	-14.2	126	0.00
85 c	Di-n-octyl phthalate	40.000	54.353	-35.9#	143	0.01
86 I	Perylene-d12	20.000	20.000	0.0	121	0.04
87	Indeno(1,2,3-cd)pyrene	40.000	40.876	-2.2	117	0.05
88	Benzo(b)fluoranthene	40.000	40.323	-0.8	116	0.03
89	Benzo(k)fluoranthene	40.000	38.719	3.2	111	0.01
90 C	Benzo(a)pyrene	40.000	40.569	-1.4	115	0.02
91	Dibenzo(a,h)anthracene	40.000	41.186	-3.0	117	0.05
92	Benzo(g,h,i)perylene	40.000	40.404	-1.0	116	0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110625\
Data File : BP026076.D
Acq On : 06 Nov 2025 13:25
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 06 14:05:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 30 11:51:34 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 3