

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110525

Review By	rahul	Review On	11/5/2025 2:56:50 PM
Supervise By	Jagrut	Supervise On	11/6/2025 10:21:33 AM
SubDirectory	BF110525	HP Acquire Method	BNA_F
		HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144168.D	05 Nov 2025 12:21		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF144169.D	05 Nov 2025 12:50		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF144170.D	05 Nov 2025 13:20		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF144171.D	05 Nov 2025 13:50		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF144172.D	05 Nov 2025 14:20		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF144173.D	05 Nov 2025 14:50	Compound #41 kept on QR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF144174.D	05 Nov 2025 15:49		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF144175.D	05 Nov 2025 16:19		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF144176.D	05 Nov 2025 16:49		RC/JU	Ok
10	SSTDICV040	ICVBF110525	BF144177.D	05 Nov 2025 17:21		RC/JU	Ok,M
11	PB170372BL	PB170372BL	BF144178.D	05 Nov 2025 18:10		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110625

Review By	rahul	Review On	11/6/2025 3:51:49 PM		
Supervise By	Jagrut	Supervise On	11/10/2025 2:04:15 PM		
SubDirectory	BF110625	HP Acquire Method	BNA_P	HP Processing Method	BF110525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144179.D	06 Nov 2025 10:44		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144180.D	06 Nov 2025 11:13		RC/JU	Ok
3	PB170402BL	PB170402BL	BF144181.D	06 Nov 2025 11:42		RC/JU	Ok
4	PB170402BS	PB170402BS	BF144182.D	06 Nov 2025 12:12		RC/JU	Ok,M
5	Q3549-02	RT4922	BF144183.D	06 Nov 2025 12:56		RC/JU	Ok
6	Q3537-01	MW-17-SOIL	BF144184.D	06 Nov 2025 13:26		RC/JU	Ok
7	Q3532-01	TP-1	BF144185.D	06 Nov 2025 13:56		RC/JU	Ok
8	Q3532-05	TP-2	BF144186.D	06 Nov 2025 14:25		RC/JU	Ok
9	Q3532-05MS	TP-2MS	BF144187.D	06 Nov 2025 14:55		RC/JU	Ok,M
10	Q3532-05MSD	TP-2MSD	BF144188.D	06 Nov 2025 15:25		RC/JU	Ok,M
11	Q3541-01	N48969	BF144189.D	06 Nov 2025 15:54		RC/JU	Ok
12	Q3534-01	MW-6	BF144190.D	06 Nov 2025 16:24	Internal standard fail	RC/JU	ReRun
13	Q3534-02	MW-1	BF144191.D	06 Nov 2025 16:54		RC/JU	Ok,M
14	Q3534-05	MW-EB-1	BF144192.D	06 Nov 2025 17:24		RC/JU	Ok
15	Q3552-03	MW-9	BF144193.D	06 Nov 2025 17:53		RC/JU	Ok
16	Q3552-05	MW-3	BF144194.D	06 Nov 2025 18:23		RC/JU	Ok
17	Q3532-08	TP-2	BF144195.D	06 Nov 2025 18:53		RC/JU	Ok
18	Q3544-04	ONYX-1	BF144196.D	06 Nov 2025 19:23		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110625

Review By	rahul	Review On	11/6/2025 3:51:49 PM		
Supervise By	Jagrut	Supervise On	11/10/2025 2:04:15 PM		
SubDirectory	BF110625	HP Acquire Method	BNA_P	HP Processing Method	BF110525
STD. NAME	STD REF.#				
Tune/Reschk	SP6875				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13181,10ul/1000ul sample				
ICV/I.BLK	SP6872				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP102925

Review By	rahul	Review On	10/30/2025 8:52:33 AM
Supervise By	Jagrut	Supervise On	11/4/2025 3:53:07 PM
SubDirectory	BP102925	HP Acquire Method	BNA_P
		HP Processing Method	BP102925

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026027.D	29 Oct 2025 08:45		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP026028.D	29 Oct 2025 09:26		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP026029.D	29 Oct 2025 10:07		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BP026030.D	29 Oct 2025 10:48		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP026031.D	29 Oct 2025 11:29		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP026032.D	29 Oct 2025 12:10	Compound #26,48,51,53,57,80,84 are kept on LR & Compound #32,54,65 are kept on QR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP026033.D	29 Oct 2025 12:51		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP026034.D	29 Oct 2025 13:32		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BP026035.D	29 Oct 2025 14:13	Compound #26 is removed from 80ppm & 60PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP102925	BP026036.D	29 Oct 2025 16:09		RC/JU	Ok,M
11	PB170259BL	PB170259BL	BP026037.D	29 Oct 2025 16:50		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP110625

Review By	rahul	Review On	11/6/2025 3:53:26 PM		
Supervise By	Jagrut	Supervise On	11/10/2025 2:04:52 PM		
SubDirectory	BP110625	HP Acquire Method	BNA_P	HP Processing Method	BP102925
STD. NAME		STD REF.#			
Tune/Reschk		SP6875			
Initial Calibration Stds		SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864			
CCC		SP6860			
Internal Standard/PEM		S13181,10ul/1000ul sample			
ICV/I.BLK		SP6872			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026075.D	06 Nov 2025 12:44		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP026076.D	06 Nov 2025 13:25	CCC Fail high for Com#26,32,41,42,54,59,65,70 Need #85	RC/JU	Ok,M
3	PB170424BL	PB170424BL	BP026077.D	06 Nov 2025 14:13		RC/JU	Ok
4	PB170424BS	PB170424BS	BP026078.D	06 Nov 2025 14:54		RC/JU	Ok,M
5	PB170424BSD	PB170424BSD	BP026079.D	06 Nov 2025 15:35		RC/JU	Ok,M
6	Q3534-09	FB-1	BP026080.D	06 Nov 2025 17:04	Split volume (True value adjusted).	RC/JU	Ok
7	Q3534-10	EB-1	BP026081.D	06 Nov 2025 17:45	Split volume (True value adjusted).	RC/JU	Ok
8	Q3534-03	MW-11	BP026082.D	06 Nov 2025 18:26		RC/JU	Ok
9	Q3534-04	MW-5	BP026083.D	06 Nov 2025 19:07		RC/JU	Ok
10	Q3534-06	MW-EB-2	BP026084.D	06 Nov 2025 19:48		RC/JU	Ok
11	Q3537-03	MW-17-PURGE-WATE	BP026085.D	06 Nov 2025 20:30		RC/JU	Ok
12	Q3552-01	MW-4	BP026086.D	06 Nov 2025 21:11		RC/JU	Ok
13	Q3552-02	MW-8	BP026087.D	06 Nov 2025 21:52		RC/JU	Ok
14	Q3552-04	MW-10	BP026088.D	06 Nov 2025 22:33		RC/JU	Ok

M : Manual Integration