

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP026075.D
Instrument ID: BNA_P

Contract: REMI02
SDG NO.: Q3552
DFTPP Injection Date: 11/06/2025
DFTPP Injection Time: 12:44

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	31.2
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	12.9
442	Greater than 50% of mass 198	84.8
443	15.0 - 24.0% of mass 442	16.4 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP026076.D	11/06/2025	13:25
PB170424BL	PB170424BL	BP026077.D	11/06/2025	14:13
PB170424BS	PB170424BS	BP026078.D	11/06/2025	14:54
PB170424BSD	PB170424BSD	BP026079.D	11/06/2025	15:35
MW-4	Q3552-01	BP026086.D	11/06/2025	21:11
MW-8	Q3552-02	BP026087.D	11/06/2025	21:52
MW-10	Q3552-04	BP026088.D	11/06/2025	22:33