

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
Lab Code: ACE SDG NO.: Q3552
Lab File ID: VV039328.D Date Analyzed: 11/05/2025
Instrument ID: MSVOA_V Time Analyzed: 11:34
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	644517	4.60	1006970	5.53	993503	8.77
UPPER LIMIT	1289030	5.096	2013950	6.029	1987010	9.273
LOWER LIMIT	322259	4.096	503487	5.029	496752	8.273
EPA SAMPLE NO.						
MW-4	698591	4.60	1269452	5.53	1250476	8.77
MW-8	905091	4.60	1744740	5.53	1630594	8.77
MW-9	677683	4.60	1221893	5.53	1164477	8.77
MW-10	602545	4.60	1106968	5.53	1064655	8.77
MW-3	929578	4.60	1730000	5.53	1622898	8.77
TB	856878	4.60	1636404	5.53	1498704	8.77
VV1105WBL01	856812	4.60	1628447	5.53	1466262	8.77
VV1105WBS01	708659	4.60	1110235	5.53	1080434	8.77
VV1105WBSD01	690271	4.60	1108460	5.53	1064992	8.77

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.