

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

RunNo.: LB137801

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.96	1	96	90-110	11/06/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	11/06/2025
Sample ID: CCV2 Ammonia as N	mg/L	1.1	1	110	90-110	11/06/2025
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	11/06/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

RunNo.: LB137825

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.493	0.50	99	90-110	11/07/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.504	0.5	101	90-110	11/07/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.501	0.5	100	90-110	11/07/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

RunNo.: LB137826

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.509	0.50	102	90-110	11/07/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.495	0.50	99	90-110	11/07/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.506	0.50	101	90-110	11/07/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

RunNo.: LB137852

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.97	1	97	90-110	11/11/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	11/11/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.96	1	96	90-110	11/11/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.97	1	97	90-110	11/11/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.96	1	96	90-110	11/11/2025
Sample ID: CCV5 Ammonia as N	mg/L	0.99	1	99	90-110	11/11/2025
Sample ID: CCV6 Ammonia as N	mg/L	0.94	1	94	90-110	11/11/2025