

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

Run No.: LB137795

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137795BS							
BOD5	mg/L	198	191		97	1	84.6-115.4	11/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

Run No.: LB137812

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137812BS							
TSS	mg/L	550	531		96	1	90-110	11/07/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

Run No.: LB137814

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137814BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	11/07/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

Run No.: LB137825

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137825BS							
Orthophosphate as P	mg/L	0.5	0.49		98	1	90-110	11/07/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

Run No.: LB137801

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170403BS							
Ammonia as N	mg/L	1	0.95		95	1	90-110	11/06/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

Run No.: LB137826

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170438BS							
Phosphorus, Total	mg/L	0.50	0.51		101	1	90-110	11/07/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3554

Project: Pre Treatment Plant 2025

Run No.: LB137852

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170466BS							
Ammonia as N	mg/L	1	0.97		97	1	90-110	11/11/2025