

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: RTPE01
 Lab Code: ACE SDG NO.: Q3555
 Lab File ID: VL043155.D Date Analyzed: 11/06/2025
 Instrument ID: MSVOA_L Time Analyzed: 07:19
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	111679	2.79	341877	3.96	299829	8.89
UPPER LIMIT	156351	3.12	478628	4.29	419761	9.22
LOWER LIMIT	67007.4	2.46	205126	3.63	179897	8.56
EPA SAMPLE NO.						
151-IA-1	115856	2.79	347710	3.96	297659	8.88
151-IA-2	116276	2.79	360094	3.96	301802	8.89
151-IA-2DUP	118024	2.79	354784	3.96	303947	8.88
151-SG-1	125587	2.79	343141	3.96	286902	8.88
151-SG-1DL	109979	2.79	312659	3.96	255972	8.88
151-SG-2	116388	2.78	320264	3.95	276449	8.87
151-SG-2DL	112780	2.79	323131	3.96	266609	8.88
VL1106ABL01	111463	2.79	342563	3.96	292913	8.88
VL1106ABS01	111380	2.79	340583	3.96	299526	8.88

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.