

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137806

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/06/2025	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	11/06/2025	16:01
3	CAL3	1	Water	NA	NA	20.2	10.02	11/06/2025	16:02
4	ICV	1	Water	NA	NA	20.3	7.02	11/06/2025	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	11/06/2025	16:07
6	Q3560-01	1	Water	NA	NA	20.4	6.11	11/06/2025	16:14
7	Q3560-01DUP	1	Water	NA	NA	20.5	6.13	11/06/2025	16:17
8	Q3560-03	1	Water	NA	NA	20.4	5.88	11/06/2025	16:20
9	Q3560-05	1	Water	NA	NA	20.9	5.74	11/06/2025	16:25
10	Q3560-07	1	Water	NA	NA	20.4	6.54	11/06/2025	16:35
11	Q3560-09	1	Water	NA	NA	20.5	5.80	11/06/2025	16:39
12	Q3561-02	1	Water	NA	NA	20.7	5.27	11/06/2025	16:44
13	Q3561-03	1	Water	NA	NA	22.8	5.23	11/06/2025	16:50
14	Q3561-04	1	Water	NA	NA	21.4	5.19	11/06/2025	16:55
15	Q3561-05	1	Water	NA	NA	21.9	5.14	11/06/2025	16:59
16	CCV2	1	Water	NA	NA	20.3	12.02	11/06/2025	17:00

43 137806

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tds q3560 WorkList ID : 192952 Department : Wet-Chemistry Date : 11-06-2025 15:43:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3552-01	H MW-4	Water	TDS	Cool 4 deg C	REMI02	D41	11/04/2025	SM2540 C
Q3552-02	H MW-8	Water	TDS	Cool 4 deg C	REMI02	D41	11/04/2025	SM2540 C
Q3552-03	H MW-9	Water	TDS	Cool 4 deg C	REMI02	D41	11/04/2025	SM2540 C
Q3552-04	H MW-10	Water	TDS	Cool 4 deg C	REMI02	D41	11/05/2025	SM2540 C
Q3552-05	H MW-3	Water	TDS	Cool 4 deg C	REMI02	D41	11/04/2025	SM2540 C
Q3560-01	M SY-10D	Water	TDS	Cool 4 deg C	LOCK01	J22	11/05/2025	SM2540 C
Q3560-03	M SY-10S	Water	TDS	Cool 4 deg C	LOCK01	J22	11/05/2025	SM2540 C
Q3560-05	M SY-10I	Water	TDS	Cool 4 deg C	LOCK01	J22	11/05/2025	SM2540 C
Q3560-07	M SY-12I	Water	TDS	Cool 4 deg C	LOCK01	J22	11/05/2025	SM2540 C
Q3560-09	M SY-12D	Water	TDS	Cool 4 deg C	LOCK01	J22	11/05/2025	SM2540 C

Date/Time 11/06/25 15:50
Raw Sample Received by: sd wca
Raw Sample Relinquished by: sd sn

Date/Time 11/06/25 18:00
Raw Sample Received by: sd sn
Raw Sample Relinquished by: sd sn