

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3560

Contract: LOCK01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170434BS							
Antimony	ug/L	400	413		103	80 - 120	P
Arsenic	ug/L	400	402		100	80 - 120	P
Barium	ug/L	100	99.8		100	80 - 120	P
Beryllium	ug/L	100	109		109	80 - 120	P
Cadmium	ug/L	100	98.7		99	80 - 120	P
Chromium	ug/L	200	210		105	80 - 120	P
Cobalt	ug/L	100	100		100	80 - 120	P
Copper	ug/L	150	159		106	80 - 120	P
Iron	ug/L	1500	1500		100	80 - 120	P
Lead	ug/L	500	484		97	80 - 120	P
Manganese	ug/L	100	104		104	80 - 120	P
Nickel	ug/L	250	253		101	80 - 120	P
Potassium	ug/L	5000	4960		99	80 - 120	P
Selenium	ug/L	1000	1050		105	80 - 120	P
Silver	ug/L	37.5	37.9		101	80 - 120	P
Sodium	ug/L	1500	1790		119	80 - 120	P
Thallium	ug/L	1000	1000		100	80 - 120	P
Vanadium	ug/L	150	154		103	80 - 120	P
Zinc	ug/L	100	107		107	80 - 120	P