

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102325

Review By	John Carlone	Review On	10/24/2025 8:16:57 AM
Supervise By	Maresh Dadoda	Supervise On	10/27/2025 2:53:14 PM
SubDirectory	VN102325	HP Acquire Method	HP Processing Method 82N102325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135909 VP135921,VP135922,VP135923,VP135924,VP135925,VP135926 VP135927

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088077.D	23 Oct 2025 09:09		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN088078.D	23 Oct 2025 09:42	COM.#74 Peak not detected in 1PPB	JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN088079.D	23 Oct 2025 10:38		JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN088080.D	23 Oct 2025 10:59		JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN088081.D	23 Oct 2025 11:20	Method failed for Comp. #74	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN088082.D	23 Oct 2025 11:41		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN088083.D	23 Oct 2025 12:02		JC\MD	Ok,M
8	IBLK	IBLK	VN088084.D	23 Oct 2025 12:23		JC\MD	Ok
9	VSTDICV050	ICVVN102325	VN088085.D	23 Oct 2025 14:54		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN110625

Review By	Amit Patel	Review On	11/7/2025 1:33:58 PM
Supervise By	Maresh Dadoda	Supervise On	11/7/2025 1:47:08 PM
SubDirectory	VN110625	HP Acquire Method	MSVOA_N HP Processing Method 82N102325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136072
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136073,VP136074 VP133935

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088196.D	06 Nov 2025 09:20		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088197.D	06 Nov 2025 11:00	pH#lot#v14222	JC\MD	Ok,M
3	VN1106MBL01	VN1106MBL01	VN088198.D	06 Nov 2025 11:35		JC\MD	Ok
4	VN1106WBL01	VN1106WBL01	VN088199.D	06 Nov 2025 12:09		JC\MD	Ok
5	VN1106MBS01	VN1106MBS01	VN088200.D	06 Nov 2025 12:32		JC\MD	Ok,M
6	VN1106WBS01	VN1106WBS01	VN088201.D	06 Nov 2025 12:53		JC\MD	Ok,M
7	Q3549-01	RT4922	VN088202.D	06 Nov 2025 13:48		JC\MD	Ok
8	Q3544-04	ONYX-1	VN088203.D	06 Nov 2025 14:08	pH#6.0 A,Surrogate fail	JC\MD	ReRun
9	Q3537-02	MW-17-SOIL	VN088204.D	06 Nov 2025 14:29	pH#6.0 A ,Surrogate fail	JC\MD	ReRun
10	Q3532-08	TP-2	VN088205.D	06 Nov 2025 14:49	pH#6.0 A	JC\MD	Ok
11	Q3532-04	TP-1	VN088206.D	06 Nov 2025 15:10	pH#6.0 A	JC\MD	Ok
12	PB170395TB	PB170395TB	VN088207.D	06 Nov 2025 15:31		JC\MD	Ok
13	VIBLK	VIBLK	VN088208.D	06 Nov 2025 15:52		JC\MD	Ok
14	Q3561-05	1027-D	VN088209.D	06 Nov 2025 16:12	pH#6.0 A ,Surrogate fail	JC\MD	Ok,M
15	Q3536-03	2061	VN088210.D	06 Nov 2025 16:33		JC\MD	Ok
16	Q3536-04	2052	VN088211.D	06 Nov 2025 16:54		JC\MD	Ok
17	Q3560-01	SY-10D	VN088212.D	06 Nov 2025 17:14	pH#1.0 A	JC\MD	Ok
18	Q3560-03	SY-10S	VN088213.D	06 Nov 2025 17:35	pH#1.0 A	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN110625

Review By	Amit Patel	Review On	11/7/2025 1:33:58 PM
Supervise By	Maresh Dadoda	Supervise On	11/7/2025 1:47:08 PM
SubDirectory	VN110625	HP Acquire Method	MSVOA_N HP Processing Method 82N102325W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136072		
CCC	VP136073,VP136074		
Internal Standard/PEM	VP133935		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3560-05	SY-10I	VN088214.D	06 Nov 2025 17:56	pH#1.0 A	JC\MD	Ok
20	Q3560-07	SY-12I	VN088215.D	06 Nov 2025 18:16	pH#1.0 A	JC\MD	Ok
21	Q3560-09	SY-12D	VN088216.D	06 Nov 2025 18:37	pH#1.0 A ,Surrogate fail	JC\MD	ReRun
22	VN1106MBSD01	VN1106MBSD01	VN088217.D	06 Nov 2025 18:58		JC\MD	Ok,M
23	VN1106WBSD01	VN1106WBSD01	VN088218.D	06 Nov 2025 19:18		JC\MD	Ok,M
24	Q3536-01	FILTER-COMP	VN088219.D	06 Nov 2025 19:39		JC\MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VN088220.D	06 Nov 2025 20:00		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN110725

Review By	John Carlone	Review On	11/10/2025 8:26:16 AM
Supervise By	Amit Patel	Supervise On	11/10/2025 9:55:09 AM
SubDirectory	VN110725	HP Acquire Method	MSVOA_N HP Processing Method 82N102325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136083
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136084,VP136085

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088221.D	07 Nov 2025 09:12		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088222.D	07 Nov 2025 09:49	pH#Lot#V12668	JC\MD	Ok,M
3	VN1107MBL01	VN1107MBL01	VN088223.D	07 Nov 2025 10:39		JC\MD	Ok
4	VN1107WBL01	VN1107WBL01	VN088224.D	07 Nov 2025 11:00		JC\MD	Ok
5	VN1107WBS01	VN1107WBS01	VN088225.D	07 Nov 2025 11:47		JC\MD	Ok,M
6	VN1107WBSD01	VN1107WBSD01	VN088226.D	07 Nov 2025 12:08		JC\MD	Ok,M
7	Q3537-02	MW-17-SOIL	VN088227.D	07 Nov 2025 12:49	vial B pH#5.0	JC\MD	Ok
8	Q3544-04	ONYX-1	VN088228.D	07 Nov 2025 13:09	vial B pH#5.0	JC\MD	Ok
9	VIBLK	VIBLK	VN088229.D	07 Nov 2025 13:30		JC\MD	Ok
10	Q3518-02	50831	VN088230.D	07 Nov 2025 13:51	vial B pH<2	JC\MD	Not Ok
11	Q3560-09	SY-12D	VN088231.D	07 Nov 2025 14:12	vial B pH<2	JC\MD	Ok,M
12	VIBLK	VIBLK	VN088232.D	07 Nov 2025 15:15		JC\MD	Ok
13	VSTDCCC050	VSTDCCC050EC	VN088233.D	07 Nov 2025 15:41		JC\MD	Ok,M

M : Manual Integration