

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3560</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Lockwood, Kessler & Bartlett, Inc.</u>	Datafile :	<u>VN088225.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1107WBS01	Vinyl chloride	20	18.8	ug/L	94			65	117	
	Chloroethane	20	17.2	ug/L	86			56	128	
	Trichlorofluoromethane	20	16.9	ug/L	85			73	115	
	1,1-Dichloroethene	20	16.7	ug/L	84			74	110	
	Acrylonitrile	100	110	ug/L	110			73	113	
	Acetone	100	110	ug/L	110			60	125	
	Carbon disulfide	20	17.6	ug/L	88			64	112	
	Methylene Chloride	20	20.6	ug/L	103			72	114	
	Vinyl Acetate	100	110	ug/L	110			76	115	
	2-Butanone	100	120	ug/L	120			65	122	
	Carbon Tetrachloride	20	19.1	ug/L	96			77	113	
	cis-1,2-Dichloroethene	20	20.8	ug/L	104			77	110	
	Bromochloromethane	20	21.5	ug/L	108			70	124	
	Chloroform	20	21.4	ug/L	107			79	113	
	1,1,1-Trichloroethane	20	19.9	ug/L	100			80	108	
	Benzene	20	20.3	ug/L	102			82	109	
	1,2-Dichloropropane	20	20.9	ug/L	104			83	111	
	Dibromomethane	20	20.6	ug/L	103			82	110	
	Bromodichloromethane	20	21.6	ug/L	108			83	110	
	4-Methyl-2-Pentanone	100	110	ug/L	110			74	118	
	Toluene	20	19.5	ug/L	98			82	110	
	t-1,3-Dichloropropene	20	20.6	ug/L	103			79	110	
	cis-1,3-Dichloropropene	20	21.3	ug/L	106			82	110	
	2-Hexanone	100	110	ug/L	110			73	117	
	Dibromochloromethane	20	20.1	ug/L	101			82	110	
	1,2-Dibromoethane	20	20.3	ug/L	102			81	110	
	Tetrachloroethene	20	17.9	ug/L	90			67	123	
	Chlorobenzene	20	19.8	ug/L	99			82	109	
	1,1,1,2-Tetrachloroethane	20	20.4	ug/L	102			84	111	
	Ethyl Benzene	20	19.5	ug/L	98			83	109	
	m/p-Xylenes	40	38.1	ug/L	95			82	110	
	o-Xylene	20	20.2	ug/L	101			83	109	
	Styrene	20	20.4	ug/L	102			80	111	
	Bromoform	20	20.3	ug/L	102			79	109	
	1,1,2,2-Tetrachloroethane	20	20.7	ug/L	104			76	118	
	1,2,3-Trichloropropane	20	19.0	ug/L	95			75	112	
	1,4-Dichlorobenzene	20	19.3	ug/L	97			82	107	
	1,2-Dichlorobenzene	20	19.7	ug/L	99			82	109	
	1,2-Dibromo-3-Chloropropane	20	19.9	ug/L	100			68	112	
	Methyl iodide	20	21.0	ug/L	105			70	130	
	trans-1,4-Dichloro-2-butene	20	17.3	ug/L	86			79	102	