

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: LOCK01
 Lab Code: ACE SDG NO.: Q3560
 Lab File ID: VN088222.D Date Analyzed: 11/07/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:49
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	312291	8.21	595955	9.08	524310	11.84
UPPER LIMIT	624582	8.706	1191910	9.582	1048620	12.341
LOWER LIMIT	156146	7.706	297978	8.582	262155	11.341
EPA SAMPLE NO.						
SY-12D	211115	8.21	474557	9.08	456150	11.84
VN1107WBL01	218414	8.21	497944	9.08	470093	11.85
VN1107WBS01	316837	8.21	610543	9.08	528698	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.