

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: LOCK01
 Lab Code: ACE SDG NO.: Q3560
 Lab File ID: VN088197.D Date Analyzed: 11/06/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	283428	13.77				
UPPER LIMIT	566856	14.27				
LOWER LIMIT	141714	13.27				
EPA SAMPLE NO.						
SY-10D	297689	13.77				
SY-10S	234197	13.77				
SY-10I	233723	13.77				
SY-12I	309131	13.77				
VN1106WBL01	362059	13.77				
VN1106WBS01	278028	13.77				
VN1106WBSD01	246663	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.