



SAMPLE DATA

Report of Analysis

Client: VERINA CONSULTING GROUP, LLC
Project: Rotor Clip NJ WTD - 2025
Client Sample ID: WATER-TREATMENT DISCHARGE
Lab Sample ID: Q3566-01

Date Collected: 11/06/25 09:32
Date Received: 11/06/25
SDG No.: Q3566
Matrix: WATER
% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Ammonia as N	1.00		1	0.030	0.10	mg/L	11/10/25 11:40	11/11/25 12:07	SM 4500-NH3 B plus G-21
Residual Chlorine	0.046	HJ	1	0.023	0.10	mg/L		11/07/25 11:03	SM 4500-Cl G-11

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



QC RESULT SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3566

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB137823

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.382	0.4	96	90-110	11/07/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.413	0.4	103	90-110	11/07/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.403	0.4	101	90-110	11/07/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3566

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB137852

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.97	1	97	90-110	11/11/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	11/11/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.96	1	96	90-110	11/11/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.97	1	97	90-110	11/11/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.96	1	96	90-110	11/11/2025
Sample ID: CCV5 Ammonia as N	mg/L	0.99	1	99	90-110	11/11/2025
Sample ID: CCV6 Ammonia as N	mg/L	0.94	1	94	90-110	11/11/2025



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3566

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB137823

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	11/07/2025
Sample ID: CCB1 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	11/07/2025
Sample ID: CCB2 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	11/07/2025

Initial and Continuing Calibration Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3566

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB137852

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/11/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/11/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/11/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/11/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/11/2025
Sample ID: CCB5 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/11/2025
Sample ID: CCB6 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	11/11/2025

Preparation Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3566

Project: Rotor Clip NJ WTD - 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137823BL							
Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	11/07/2025
Sample ID: PB170466BL							
Ammonia as N	mg/L	< 0.0500	0.0500	U	0.03	0.1	11/11/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3560-01
Client ID:	SY-10DMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	9.10	OR	8.10	OR	1	1	100		11/11/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3560-01
Client ID:	SY-10DMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	9.10	OR	8.10	OR	1	1	100		11/11/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3566-01
Client ID:	WATER-TREATMENT DISCHARGEMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	0.41		0.046	J	0.4	1	92		11/07/2025

Matrix Spike Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3566-01
Client ID:	WATER-TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	0.42		0.046	J	0.4	1	94		11/07/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3560-01
Client ID:	SY-10DDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	8.10	OR	8.10	OR	1	0		11/11/2025
Ammonia as N	mg/L	+/-20	7.40	D	7.40	D	10	0		11/11/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3560-01
Client ID:	SY-10DMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	9.10	OR	9.10	OR	1	0		11/11/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3566-01
Client ID:	WATER-TREATMENT DISCHARGEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	0.046	J	0.046	J	1	0		11/07/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3566-01
Client ID:	WATER-TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	0.41		0.42		1	2.39		11/07/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137823

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137823BS							
Residual Chlorine	mg/L	0.4	0.39		98	1	90-110	11/07/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137852

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170466BS							
Ammonia as N	mg/L	1	0.97		97	1	90-110	11/11/2025

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137823

Review By	Iwona	Review On	11/10/2025 11:19:12 AM
Supervise By	jignesh	Supervise On	11/10/2025 4:45:26 PM
SubDirectory	LB137823	Test	Residual Chlorine
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP115542,WP115541,WP115538,WP115539,WP115555,WP115537,WP115556,WP115543,WP115540,W3147		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	11/07/25 10:20		Iwona	OK
2	CAL2	CAL2	CAL	11/07/25 10:23		Iwona	OK
3	CAL3	CAL3	CAL	11/07/25 10:26		Iwona	OK
4	CAL4	CAL4	CAL	11/07/25 10:29		Iwona	OK
5	CAL5	CAL5	CAL	11/07/25 10:32		Iwona	OK
6	CAL6	CAL6	CAL	11/07/25 10:35		Iwona	OK
7	ICV	ICV	ICV	11/07/25 10:38		Iwona	OK
8	ICB	ICB	ICB	11/07/25 10:41		Iwona	OK
9	CCV1	CCV1	CCV	11/07/25 10:44		Iwona	OK
10	CCB1	CCB1	CCB	11/07/25 10:47		Iwona	OK
11	LB137823BL	LB137823BL	MB	11/07/25 10:50		Iwona	OK
12	LB137823BS	LB137823BS	LCS	11/07/25 10:53		Iwona	OK
13	Q3530-07	LOD-MDL-WATER-01	SAM	11/07/25 10:57		Iwona	OK
14	Q3530-08	LOQ-WATER-02-QT4	SAM	11/07/25 11:00		Iwona	OK
15	Q3566-01	WATER-TREATMENT	SAM	11/07/25 11:03		Iwona	OK
16	Q3566-01DUP	WATER-TREATMENT	DUP	11/07/25 11:06		Iwona	OK
17	Q3566-01MS	WATER-TREATMENT	MS	11/07/25 11:09		Iwona	OK
18	Q3566-01MSD	WATER-TREATMENT	MSD	11/07/25 11:12		Iwona	OK

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137823

Review By	Iwona	Review On	11/10/2025 11:19:12 AM
Supervise By	jignesh	Supervise On	11/10/2025 4:45:26 PM
SubDirectory	LB137823	Test	Residual Chlorine

STD. NAME	STD REF.#
ICAL Standard	N/A
ICV Standard	N/A
CCV Standard	N/A
ICSA Standard	N/A
CRI Standard	N/A
LCS Standard	N/A
Chk Standard	WP115542,WP115541,WP115538,WP115539,WP115555,WP115537,WP115556,WP115543,WP115540,W3147

19	CCV2	CCV2	CCV	11/07/25 11:15		Iwona	OK
20	CCB2	CCB2	CCB	11/07/25 11:18		Iwona	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137852

Review By	rubina	Review On	11/12/2025 1:03:52 PM
Supervise By	Iwona	Supervise On	11/12/2025 1:04:18 PM
SubDirectory	LB137852	Test	Ammonia
STD. NAME	STD REF.#		
ICAL Standard	WP115598		
ICV Standard	WP115600		
CCV Standard	WP115599		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115589		
Chk Standard	WP115290,WP114133,WP113929,WP114132,WP115590		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPM	0.0PPM	CAL1	11/11/25 10:21		rubina	OK
2	0.1PPM	0.1PPM	CAL2	11/11/25 10:21		rubina	OK
3	0.2PPM	0.2PPM	CAL3	11/11/25 10:21		rubina	OK
4	0.4PPM	0.4PPM	CAL4	11/11/25 10:21		rubina	OK
5	1.0PPM	1.0PPM	CAL5	11/11/25 10:21		rubina	OK
6	1.3PPM	1.3PPM	CAL6	11/11/25 10:21		rubina	OK
7	2.0PPM	2.0PPM	CAL7	11/11/25 10:21		rubina	OK
8	ICV1	ICV1	ICV	11/11/25 11:35		rubina	OK
9	ICB1	ICB1	ICB	11/11/25 11:35		rubina	OK
10	CCV1	CCV1	CCV	11/11/25 11:35		rubina	OK
11	CCB1	CCB1	CCB	11/11/25 11:35		rubina	OK
12	RL	RL	LOQ	11/11/25 11:35		rubina	OK
13	PB170466BL	PB170466BL	MB	11/11/25 11:45		rubina	OK
14	PB170466BS	PB170466BS	LCS	11/11/25 11:46		rubina	OK
15	Q3530-08	LOQ-WATER-02-QT4	LOQ	11/11/25 11:46		rubina	OK
16	Q3554-01	EFFLUENT	SAM	11/11/25 11:46	NH3 is high, need dilution	rubina	Dilution
17	Q3560-01	SY-10D	SAM	11/11/25 11:46	NH3 is high, need dilution	rubina	Dilution
18	Q3560-01DUP	SY-10DDUP	DUP	11/11/25 11:56	NH3 is high, need dilution	rubina	Dilution

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137852

Review By	rubina	Review On	11/12/2025 1:03:52 PM
Supervise By	Iwona	Supervise On	11/12/2025 1:04:18 PM
SubDirectory	LB137852	Test	Ammonia
STD. NAME	STD REF.#		
ICAL Standard	WP115598		
ICV Standard	WP115600		
CCV Standard	WP115599		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115589		
Chk Standard	WP115290,WP114133,WP113929,WP114132,WP115590		

19	Q3560-01MS	SY-10DMS	MS	11/11/25 11:56		rubina	OK
20	Q3560-01MSD	SY-10DMSD	MSD	11/11/25 11:56		rubina	OK
21	CCV2	CCV2	CCV	11/11/25 11:56		rubina	OK
22	CCB2	CCB2	CCB	11/11/25 11:56		rubina	OK
23	Q3560-03	SY-10S	SAM	11/11/25 11:56		rubina	OK
24	Q3560-05	SY-10I	SAM	11/11/25 11:56		rubina	OK
25	Q3560-07	SY-12I	SAM	11/11/25 12:07	NH3 is high, need dilution	rubina	Dilution
26	Q3560-09	SY-12D	SAM	11/11/25 12:07		rubina	OK
27	Q3566-01	WATER-TREATMENT	SAM	11/11/25 12:07		rubina	OK
28	Q3575-01	001 Willets Pt Blvd (N	SAM	11/11/25 12:07	NH3 is high, need dilution	rubina	Dilution
29	Q3575-02	002 35th Ave (Nov)	SAM	11/11/25 12:07	NH3 is high, need dilution	rubina	Dilution
30	PB170467BL	PB170467BL	MB	11/11/25 12:07		rubina	OK
31	PB170467BS	PB170467BS	LCS	11/11/25 12:07		rubina	OK
32	CCV3	CCV3	CCV	11/11/25 12:18		rubina	OK
33	CCB3	CCB3	CCB	11/11/25 12:18		rubina	OK
34	Q3530-02	LOQ-SOIL-02-QT4-20	LOQ	11/11/25 12:18		rubina	OK
35	Q3483-07	HW1025-PT-NUT-SO	SAM	11/11/25 12:18	NH3 is high, need dilution	rubina	Dilution
36	CCV4	CCV4	CCV	11/11/25 12:28		rubina	OK
37	CCB4	CCB4	CCB	11/11/25 12:28		rubina	OK
38	Q3554-01DL	EFFLUENTDL	SAM	11/11/25 13:06	10X For NH3	rubina	Confirms

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QCBatch ID # LB137852

Review By	rubina	Review On	11/12/2025 1:03:52 PM
Supervise By	Iwona	Supervise On	11/12/2025 1:04:18 PM
SubDirectory	LB137852	Test	Ammonia
STD. NAME	STD REF.#		
ICAL Standard	WP115598		
ICV Standard	WP115600		
CCV Standard	WP115599		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115589		
Chk Standard	WP115290,WP114133,WP113929,WP114132,WP115590		

39	Q3560-01DL	SY-10DDL	SAM	11/11/25 13:06	10X For NH3	rubina	Confirms
40	Q3560-01DUPDL	SY-10DDUPDL	DUP	11/11/25 13:06	10X For NH3	rubina	Confirms
41	Q3560-07DL	SY-12IDL	SAM	11/11/25 13:06	20x For NH3 Still high	rubina	Dilution
42	Q3575-01DL	001 Willets Pt Blvd (N	SAM	11/11/25 13:16	5X For NH3	rubina	Confirms
43	Q3575-02DL	002 35th Ave (Nov)DL	SAM	11/11/25 13:16	5X For NH3	rubina	Confirms
44	Q3483-07DL	HW1025-PT-NUT-SO	SAM	11/11/25 13:16	20x For NH3	rubina	Confirms
45	CCV5	CCV5	CCV	11/11/25 13:22		rubina	OK
46	CCB5	CCB5	CCB	11/11/25 13:22		rubina	OK
47	Q3530-07	LOD-MDL-WATER-01	SAM	11/11/25 13:46		rubina	OK
48	Q3530-01	LOD-MDL-SOIL-01-Q	SAM	11/11/25 13:46		rubina	OK
49	Q3560-07DL2	SY-12IDL2	SAM	11/11/25 13:46	40X For NH3	rubina	Confirms
50	CCV6	CCV6	CCV	11/11/25 13:46		rubina	OK
51	CCB6	CCB6	CCB	11/11/25 13:46		rubina	OK

LAB CHRONICLE

OrderID:	Q3566	OrderDate:	11/6/2025 1:42:00 PM
Client:	VERINA CONSULTING GROUP, LLC	Project:	Rotor Clip NJ WTD - 2025
Contact:	Michael Valenzi	Location:	J23

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3566-01	WATER-TREATMENT DISCHARGE	WATER			11/06/25 09:32			11/06/25
			Ammonia	SM4500-NH3		11/10/25	11/11/25 12:07	
			Residual Chlorine	SM4500 Cl G			11/07/25 11:03	

SOP ID : MSM4500-NH3 B,G-Ammonia-18

SDG No : N/A

Matrix : WATER

Pipette ID : WC

Balance ID : N/A

Hood ID : HOOD#2

Block ID : WC-DIST-BLOCK-1

Weigh By : N/A

Start Digest Date: 11/10/2025 Time : 11:40 Temp : 150 °C

End Digest Date: 11/10/2025 Time : 12:40 Temp : 160 °C

11 batch
11/10/2025 13:15 150g
11/10/2025 14:15 160g

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Filter paper ID : N/A

Prep Technician Signature: *PM*

pH Meter ID : N/A

Supervisor Signature: *12*

Standard Name	MLS USED	STD REF. # FROM LOG
LCSW	1.0ML	WP115589
MS/MSD SPIKE SOL.	1.0ML	WP115588
PBW	50.0ML	W3112
LOD	0.8ML	WP115585
LOQ	1.0ML	WP115585

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP113886
NAOH 6N	0.5-2.0ML	WP113887
H2SO4 0.04N	5.0ML	WP115336
pH strip-Ammonia	N/A	W3133
KI-starch paper	N/A	W3155
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT WP114104. Due to bad matrix and client history 1ML was taken as an initial volume for Q3554-01

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/10/2025 14:30	<i>RM (wc)</i>	<i>RM (wc)</i>
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB170466BL	PBW466	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB170466BS	LCS466	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3530-07	LOD-MDL-WATER-01-QT4-20 25	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3530-08	LOQ-WATER-02-QT4-2025	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3554-01	EFFLUENT	1	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3560-01	SY-10D	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3560-01DUP	SY-10DDUP	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3560-01MS	SY-10DMS	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3560-01MSD	SY-10DMSD	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3560-03	SY-10S	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3560-05	SY-10I	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3560-07	SY-12I	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3560-09	SY-12D	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3566-01	WATER-TREATMENT DISCHARGE	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3575-01	001 WILLETS PT BLVD (NOV)	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q3575-02	002 35TH AVE (NOV)	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A