

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137823

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137823BS							
Residual Chlorine	mg/L	0.4	0.39		98	1	90-110	11/07/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3566
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137852

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170466BS							
Ammonia as N	mg/L	1	0.97		97	1	90-110	11/11/2025