



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3566

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	9.05	10.0	90	65 - 135	P	11/10/2025	12:16	LB137841
	Copper	21.9	20.0	110	65 - 135	P	11/10/2025	12:16	LB137841
	Nickel	39.7	40.0	99	65 - 135	P	11/10/2025	12:16	LB137841
	Zinc	43.2	40.0	108	65 - 135	P	11/10/2025	12:16	LB137841