

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q3566
Contract: VERI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3566-01 **Client ID:** WATER-TREATMENT DISCHARGEDU
Percent Solids for Sample: NA **Duplicate ID** Q3566-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	160		156		3		P
Copper	ug/L	20	6.36	J	5.81	J	9		P
Nickel	ug/L	20	6.37	J	5.95	J	7		P
Zinc	ug/L	20	49.7		48.8		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q3566
Contract: VERI01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3566-01MS **Client ID:** WATER-TREATMENT DISCHARGEMS
Percent Solids for Sample: NA **Duplicate ID** Q3566-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	377		422		11		P
Copper	ug/L	20	159		154		3		P
Nickel	ug/L	20	277		273		1		P
Zinc	ug/L	20	160		155		3		P