

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3566

Contract: VERI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170434BS							
Chromium	ug/L	200	210		105	80 - 120	P
Copper	ug/L	150	159		106	80 - 120	P
Nickel	ug/L	250	253		101	80 - 120	P
Zinc	ug/L	100	107		107	80 - 120	P