

Metals

-9 -

ICP SERIAL DILUTIONS

SAMPLE NO.

WATER-TREATMENT DISCHARGEL

Lab Name: Alliance

Contract: VERI01

Lab Code: ACE **Lb No.:** lb137841 **Lab Sample ID :** Q3566-01L **SDG No.:** Q3566

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence		Q	M
	C		C					
Chromium	160		155		3			P
Copper	6.36	J	50.0	U	100.0			P
Nickel	6.37	J	100	U	100.0			P
Zinc	49.7		49.7	J	0			P