



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3567

Order ID: Q3567

Client: ATC Group Services LLC

Project ID: 2025 SCA 083

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

- A
- B
- C
- D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ATC Group Services LLC		Date Collected:	11/06/25
Project:	2025 SCA 083		Date Received:	11/06/25
Client Sample ID:	1A1B1C		SDG No.:	Q3567
Lab Sample ID:	Q3567-01		Matrix:	CAULK
Analytical Method:	8082A		% Solid:	100
Sample Wt/Vol:	1.1 g	Final Vol: 10000 uL	Test:	PCB Group1
Prep Method:	SW3541B	Prep Date: 11/07/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	108	U	1	108	464	ug/kg	11/07/25 12:49	PB170432
11104-28-2	Aroclor-1221	110	U	1	110	464	ug/kg	11/07/25 12:49	PB170432
11141-16-5	Aroclor-1232	101	U	1	101	464	ug/kg	11/07/25 12:49	PB170432
53469-21-9	Aroclor-1242	109	U	1	109	464	ug/kg	11/07/25 12:49	PB170432
12672-29-6	Aroclor-1248	161	U	1	161	464	ug/kg	11/07/25 12:49	PB170432
11097-69-1	Aroclor-1254	87.5	U	1	87.5	464	ug/kg	11/07/25 12:49	PB170432
37324-23-5	Aroclor-1262	137	U	1	137	464	ug/kg	11/07/25 12:49	PB170432
11100-14-4	Aroclor-1268	98.2	U	1	98.2	464	ug/kg	11/07/25 12:49	PB170432
11096-82-5	Aroclor-1260	88.1	U	1	88.1	464	ug/kg	11/07/25 12:49	PB170432
Total PCBs	Total PCBs	161	U	1	161	464	ug/kg	11/07/25 12:49	PB170432
SURROGATES									
877-09-8	Tetrachloro-m-xylene	28.1			32 - 144	141%	SPK: 20		
2051-24-3	Decachlorobiphenyl	26.4			32 - 175	132%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ATC Group Services LLC		Date Collected:	11/06/25
Project:	2025 SCA 083		Date Received:	11/06/25
Client Sample ID:	2A2B2C		SDG No.:	Q3567
Lab Sample ID:	Q3567-02		Matrix:	CAULK
Analytical Method:	8082A		% Solid:	100
Sample Wt/Vol:	0.73 g	Final Vol: 10000 uL	Test:	PCB Group1
Prep Method:	SW3541B		Prep Date:	11/07/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	162	U	1	162	699	ug/kg	11/07/25 13:05	PB170432
11104-28-2	Aroclor-1221	166	U	1	166	699	ug/kg	11/07/25 13:05	PB170432
11141-16-5	Aroclor-1232	153	U	1	153	699	ug/kg	11/07/25 13:05	PB170432
53469-21-9	Aroclor-1242	165	U	1	165	699	ug/kg	11/07/25 13:05	PB170432
12672-29-6	Aroclor-1248	243	U	1	243	699	ug/kg	11/07/25 13:05	PB170432
11097-69-1	Aroclor-1254	132	U	1	132	699	ug/kg	11/07/25 13:05	PB170432
37324-23-5	Aroclor-1262	206	U	1	206	699	ug/kg	11/07/25 13:05	PB170432
11100-14-4	Aroclor-1268	148	U	1	148	699	ug/kg	11/07/25 13:05	PB170432
11096-82-5	Aroclor-1260	133	U	1	133	699	ug/kg	11/07/25 13:05	PB170432
Total PCBs	Total PCBs	243	U	1	243	699	ug/kg	11/07/25 13:05	PB170432
SURROGATES									
877-09-8	Tetrachloro-m-xylene	28.4			32 - 144	142%	SPK: 20		
2051-24-3	Decachlorobiphenyl	28.4			32 - 175	142%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ATC Group Services LLC		Date Collected:	11/06/25
Project:	2025 SCA 083		Date Received:	11/06/25
Client Sample ID:	3A3B3C		SDG No.:	Q3567
Lab Sample ID:	Q3567-03		Matrix:	CAULK
Analytical Method:	8082A		% Solid:	100
Sample Wt/Vol:	1.58 g	Final Vol: 10000 uL	Test:	PCB Group1
Prep Method:	SW3541B		Prep Date:	11/07/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	75.0	U	1	75.0	323	ug/kg	11/07/25 13:21	PB170432
11104-28-2	Aroclor-1221	76.5	U	1	76.5	323	ug/kg	11/07/25 13:21	PB170432
11141-16-5	Aroclor-1232	70.6	U	1	70.6	323	ug/kg	11/07/25 13:21	PB170432
53469-21-9	Aroclor-1242	76.1	U	1	76.1	323	ug/kg	11/07/25 13:21	PB170432
12672-29-6	Aroclor-1248	112	U	1	112	323	ug/kg	11/07/25 13:21	PB170432
11097-69-1	Aroclor-1254	60.9	U	1	60.9	323	ug/kg	11/07/25 13:21	PB170432
37324-23-5	Aroclor-1262	95.3	U	1	95.3	323	ug/kg	11/07/25 13:21	PB170432
11100-14-4	Aroclor-1268	68.4	U	1	68.4	323	ug/kg	11/07/25 13:21	PB170432
11096-82-5	Aroclor-1260	61.3	U	1	61.3	323	ug/kg	11/07/25 13:21	PB170432
Total PCBs	Total PCBs	112	U	1	112	323	ug/kg	11/07/25 13:21	PB170432
SURROGATES									
877-09-8	Tetrachloro-m-xylene	29.3	*		32 - 144	146%	SPK: 20		
2051-24-3	Decachlorobiphenyl	26.0			32 - 175	130%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID: Q3567
Client: ATC Group Services LLC
Contact: Olga Seldinas

OrderDate: 11/6/2025 2:49:00 PM
Project: 2025 SCA 083
Location: A11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3567-01	1A1B1C	CAULK	PCB Group1	8082A	11/06/25	11/07/25	11/07/25	11/06/25
Q3567-02	2A2B2C	CAULK	PCB Group1	8082A	11/06/25	11/07/25	11/07/25	11/06/25
Q3567-03	3A3B3C	CAULK	PCB Group1	8082A	11/06/25	11/07/25	11/07/25	11/06/25