



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170432BS

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q3567

Lab Sample ID: PB170432BS

Date(s) Analyzed: 11/07/2025 11/07/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP076380.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.792	5.742	5.842	185	176	8.9
COLUMN 1	2	5.813	5.763	5.863	170		
	3	5.876	5.826	5.926	173		
	4	5.973	5.923	6.023	176		
	5	6.265	6.215	6.315	173		
	1	4.873	4.823	4.923	162		
COLUMN 2	2	4.931	4.881	4.981	156	161	
	3	5.051	5.001	5.101	152		
	4	5.09	5.04	5.14	183		
	5	5.303	5.253	5.353	150		
Aroclor-1260	1	7.383	7.333	7.433	194	181	5.38
COLUMN 1	2	7.638	7.588	7.688	194		
	3	7.996	7.946	8.046	169		
	4	8.222	8.172	8.272	186		
	5	8.549	8.499	8.599	163		
	1	6.332	6.282	6.382	196		
COLUMN 2	2	6.519	6.469	6.569	196	191	
	3	6.671	6.621	6.721	196		
	4	7.141	7.091	7.191	196		
	5	7.382	7.332	7.432	173		