

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3567
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076373.D **Time Analyzed:** 09:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.787	5.695	5.895	576.250	500.000	15.3
Aroclor-1016-2	5.809	5.717	5.917	503.130	500.000	0.6
Aroclor-1016-3	5.872	5.780	5.980	506.540	500.000	1.3
Aroclor-1016-4	5.969	5.877	6.077	520.000	500.000	4.0
Aroclor-1016-5	6.261	6.169	6.369	537.980	500.000	7.6
Aroclor-1260-1	7.380	7.287	7.487	570.690	500.000	14.1
Aroclor-1260-2	7.633	7.540	7.740	544.650	500.000	8.9
Aroclor-1260-3	7.991	7.899	8.099	559.970	500.000	12.0
Aroclor-1260-4	8.219	8.127	8.327	560.670	500.000	12.1
Aroclor-1260-5	8.544	8.452	8.652	530.860	500.000	6.2
Decachlorobiphenyl	10.407	10.319	10.519	49.920	50.000	-0.2
Tetrachloro-m-xylene	4.637	4.542	4.742	53.560	50.000	7.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3567
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076373.D **Time Analyzed:** 09:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.870	4.778	4.978	524.510	500.000	4.9
Aroclor-1016-2	4.930	4.836	5.036	483.130	500.000	-3.4
Aroclor-1016-3	5.048	4.954	5.154	482.550	500.000	-3.5
Aroclor-1016-4	5.090	4.996	5.196	545.090	500.000	9.0
Aroclor-1016-5	5.303	5.210	5.410	504.070	500.000	0.8
Aroclor-1260-1	6.330	6.238	6.438	577.760	500.000	15.6
Aroclor-1260-2	6.517	6.425	6.625	566.710	500.000	13.3
Aroclor-1260-3	6.670	6.578	6.778	575.490	500.000	15.1
Aroclor-1260-4	7.138	7.047	7.247	584.730	500.000	16.9
Aroclor-1260-5	7.380	7.287	7.487	568.140	500.000	13.6
Decachlorobiphenyl	8.772	8.682	8.882	47.390	50.000	-5.2
Tetrachloro-m-xylene	3.777	3.681	3.881	56.240	50.000	12.5