

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3567
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076387.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.792	5.695	5.895	558.350	500.000	11.7
Aroclor-1016-2	5.814	5.717	5.917	487.610	500.000	-2.5
Aroclor-1016-3	5.876	5.780	5.980	499.540	500.000	-0.1
Aroclor-1016-4	5.973	5.877	6.077	514.760	500.000	3.0
Aroclor-1016-5	6.265	6.169	6.369	510.060	500.000	2.0
Aroclor-1260-1	7.385	7.287	7.487	568.790	500.000	13.8
Aroclor-1260-2	7.638	7.540	7.740	547.040	500.000	9.4
Aroclor-1260-3	7.997	7.899	8.099	569.590	500.000	13.9
Aroclor-1260-4	8.225	8.127	8.327	578.240	500.000	15.6
Aroclor-1260-5	8.550	8.452	8.652	539.090	500.000	7.8
Decachlorobiphenyl	10.414	10.319	10.519	50.610	50.000	1.2
Tetrachloro-m-xylene	4.641	4.542	4.742	52.800	50.000	5.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3567
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076387.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.874	4.778	4.978	511.950	500.000	2.4
Aroclor-1016-2	4.933	4.836	5.036	486.820	500.000	-2.6
Aroclor-1016-3	5.053	4.954	5.154	451.830	500.000	-9.6
Aroclor-1016-4	5.094	4.996	5.196	514.350	500.000	2.9
Aroclor-1016-5	5.307	5.210	5.410	489.330	500.000	-2.1
Aroclor-1260-1	6.334	6.238	6.438	561.890	500.000	12.4
Aroclor-1260-2	6.522	6.425	6.625	568.140	500.000	13.6
Aroclor-1260-3	6.675	6.578	6.778	558.300	500.000	11.7
Aroclor-1260-4	7.141	7.047	7.247	565.190	500.000	13.0
Aroclor-1260-5	7.384	7.287	7.487	569.140	500.000	13.8
Decachlorobiphenyl	8.777	8.682	8.882	48.150	50.000	-3.7
Tetrachloro-m-xylene	3.780	3.681	3.881	52.830	50.000	5.7