

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137803

Review By	rubina	Review On	11/11/2025 2:03:13 PM
Supervise By	Iwona	Supervise On	11/11/2025 2:07:31 PM
SubDirectory	LB137803	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP115441,WP115442,WP115443,WP115444,WP115445,WP115446,WP115447		
ICV Standard	WP115448		
CCV Standard	WP115505,WP115574		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115504,WP115575		
Chk Standard	WP115449,WP115450		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	11/03/25 15:58	All standards, samples, and	RM/IZ	OK
2	STD2	STD2	CAL2	11/03/25 16:22	QC are filtered through	RM/IZ	OK
3	STD3	STD3	CAL3	11/03/25 17:04	0.45um, filter lot W3160	RM/IZ	OK
4	STD4	STD4	CAL4	11/03/25 17:47		RM/IZ	OK
5	STD5	STD5	CAL5	11/03/25 18:08		RM/IZ	OK
6	STD6	STD6	CAL6	11/03/25 18:30		RM/IZ	OK
7	STD7	STD7	CAL7	11/03/25 18:51		RM/IZ	OK
8	ICV1	ICV1	ICV	11/03/25 19:13		RM/IZ	OK
9	ICB1	ICB1	ICB	11/03/25 19:34		RM/IZ	OK
10	CCV1	CCV1	CCV	11/06/25 09:15		RM/IZ	OK
11	CCB1	CCB1	CCB	11/06/25 09:37		RM/IZ	OK
12	LB137803BLW	LB137803BLW	MB	11/06/25 09:58		RM/IZ	OK
13	LB137803BSW	LB137803BSW	LCS	11/06/25 10:20		RM/IZ	OK
14	Q3551-01	MANHOLE	SAM	11/06/25 10:46		RM/IZ	OK
15	Q3551-01MS	MANHOLEMS	MS	11/06/25 11:09	9.5ml of sample, 0.5mL W3096	RM/IZ	OK
16	Q3551-01MSD	MANHOLEMSD	MSD	11/06/25 11:31	9.5ml of sample, 0.5mL W3096	RM/IZ	OK
17	CCV2	CCV2	CCV	11/06/25 12:01		RM/IZ	OK
18	CCB2	CCB2	CCB	11/06/25 12:25		RM/IZ	OK

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137803

Review By	rubina	Review On	11/11/2025 2:03:13 PM
Supervise By	Iwona	Supervise On	11/11/2025 2:07:31 PM
SubDirectory	LB137803	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP115441,WP115442,WP115443,WP115444,WP115445,WP115446,WP115447		
ICV Standard	WP115448		
CCV Standard	WP115505,WP115574		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115504,WP115575		
Chk Standard	WP115449,WP115450		

19	Q3560-01	SY-10D	SAM	11/06/25 15:13	Cl is high, need dilution	RM/IZ	Dilution
20	Q3560-03	SY-10S	SAM	11/06/25 15:34	Cl is high, need dilution	RM/IZ	Dilution
21	Q3560-05	SY-10I	SAM	11/06/25 15:56	Cl,NO3 is high, need dilution	RM/IZ	Dilution
22	Q3560-07	SY-12I	SAM	11/06/25 16:17	Cl, SO4 is high, need dilution	RM/IZ	Dilution
23	Q3560-09	SY-12D	SAM	11/06/25 16:39	Cl,NO3,SO4 is high, need dilution	RM/IZ	Dilution
24	CCV3	CCV3	CCV	11/06/25 17:00		RM/IZ	OK
25	CCB3	CCB3	CCB	11/06/25 17:43		RM/IZ	OK
26	CCV4	CCV4	CCV	11/07/25 09:00		RM/IZ	OK
27	CCB4	CCB4	CCB	11/07/25 09:22		RM/IZ	OK
28	LB137803BLW2	LB137803BLW2	MB	11/07/25 09:43		RM/IZ	OK
29	LB137803BSW2	LB137803BSW2	LCS	11/07/25 10:05		RM/IZ	OK
30	Q3569-01	MW-2	SAM	11/07/25 10:26		RM/IZ	OK
31	Q3569-02	MW-12	SAM	11/07/25 10:48	High	RM/IZ	Dilution
32	Q3569-01DL	MW-2DL	SAM	11/07/25 11:09	For Nitrate	RM/IZ	Confirms
33	Q3560-01DL	SY-10DDL	SAM	11/07/25 11:31	200X For Cl	RM/IZ	Confirms
34	Q3560-03DL	SY-10SDL	SAM	11/07/25 11:52	5X For Cl	RM/IZ	Confirms
35	Q3560-05DL	SY-10IDL	SAM	11/07/25 12:14	20X For Cl,NO3	RM/IZ	Confirms
36	Q3560-07DL	SY-12IDL	SAM	11/07/25 12:35	5X for Cl	RM/IZ	Dilution
37	Q3560-07DL2	SY-12IDL2	SAM	11/07/25 12:57	50X For Cl	RM/IZ	Confirms

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137803

Review By	rubina	Review On	11/11/2025 2:03:13 PM
Supervise By	Iwona	Supervise On	11/11/2025 2:07:31 PM
SubDirectory	LB137803	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP115441,WP115442,WP115443,WP115444,WP115445,WP115446,WP115447		
ICV Standard	WP115448		
CCV Standard	WP115505,WP115574		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115504,WP115575		
Chk Standard	WP115449,WP115450		

38	CCV5	CCV5	CCV	11/07/25 13:19		RM/IZ	OK
39	CCB5	CCB5	CCB	11/07/25 13:40		RM/IZ	OK
40	Q3560-09DL	SY-12DDL	SAM	11/07/25 14:02	2X For Cl, SO4 ,still high	RM/IZ	Dilution
41	Q3560-09DL2	SY-12DDL2	SAM	11/07/25 14:45	20X For Cl, SO4	RM/IZ	Confirms
42	CCV6	CCV6	CCV	11/07/25 15:28		RM/IZ	OK
43	CCB6	CCB6	CCB	11/07/25 15:49		RM/IZ	OK

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137821

Review By	rubina	Review On	11/7/2025 3:29:49 PM
Supervise By	Iwona	Supervise On	11/7/2025 4:25:26 PM
SubDirectory	LB137821	Test	Hexavalent Chromium
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP115550,WP115549,WP115547,WP115546,WP115554,WP115340,WP115548,WP115553,WP115551,WP115552		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	11/07/25 09:30		Rubina	OK
2	CAL2	CAL2	CAL	11/07/25 09:30		Rubina	OK
3	CAL3	CAL3	CAL	11/07/25 09:31		Rubina	OK
4	CAL4	CAL4	CAL	11/07/25 09:31		Rubina	OK
5	CAL5	CAL5	CAL	11/07/25 09:32		Rubina	OK
6	CAL6	CAL6	CAL	11/07/25 09:32		Rubina	OK
7	CAL7	CAL7	CAL	11/07/25 09:33		Rubina	OK
8	ICV	ICV	ICV	11/07/25 09:33		Rubina	OK
9	ICB	ICB	ICB	11/07/25 09:34		Rubina	OK
10	CCV1	CCV1	CCV	11/07/25 09:34		Rubina	OK
11	CCB1	CCB1	CCB	11/07/25 09:35		Rubina	OK
12	RL Check	RL Check	RL	11/07/25 09:35		Rubina	OK
13	LB137821BL	LB137821BL	MB	11/07/25 09:36		Rubina	OK
14	LB137821BS	LB137821BS	LCS	11/07/25 09:36		Rubina	OK
15	Q3569-01	MW-2	SAM	11/07/25 09:37		Rubina	OK
16	Q3569-01DUP	MW-2DUP	DUP	11/07/25 09:37		Rubina	OK
17	Q3569-01MS	MW-2MS	MS	11/07/25 09:38		Rubina	OK
18	Q3569-01MSD	MW-2MSD	MSD	11/07/25 09:38		Rubina	OK

Instrument ID: SPECTROPHOTOMETER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137821

Review By	rubina	Review On	11/7/2025 3:29:49 PM
Supervise By	Iwona	Supervise On	11/7/2025 4:25:26 PM
SubDirectory	LB137821	Test	Hexavalent Chromium
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP115550,WP115549,WP115547,WP115546,WP115554,WP115340,WP115548,WP115553,WP115551,WP115552		

19	Q3569-02	MW-12	SAM	11/07/25 09:39		Rubina	OK
20	CCV2	CCV2	CCV	11/07/25 09:39		Rubina	OK
21	CCB2	CCB2	CCB	11/07/25 09:40		Rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137840

Review By	rubina	Review On	11/11/2025 12:55:24 PM
Supervise By	Iwona	Supervise On	11/11/2025 2:08:01 PM
SubDirectory	LB137840	Test	Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP115577,WP115578,WP115579,WP115580,WP115581,WP115582,WP115583		
ICV Standard	W3012		
CCV Standard	WP115578		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115585		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPBCN	0.0PPBCN	CAL1	11/10/25 10:56		rubina	OK
2	5.0PPBCN	5.0PPBCN	CAL2	11/10/25 10:56		rubina	OK
3	10PPBCN	10PPBCN	CAL3	11/10/25 10:56		rubina	OK
4	50PPBCN	50PPBCN	CAL4	11/10/25 10:56		rubina	OK
5	100PPBCN	100PPBCN	CAL5	11/10/25 10:56		rubina	OK
6	250PPBCN	250PPBCN	CAL6	11/10/25 10:56		rubina	OK
7	500PPBCN	500PPBCN	CAL7	11/10/25 10:56		rubina	OK
8	ICV1	ICV1	ICV	11/10/25 14:17		rubina	OK
9	ICB1	ICB1	ICB	11/10/25 14:17		rubina	OK
10	CCV1	CCV1	CCV	11/10/25 14:17		rubina	OK
11	CCB1	CCB1	CCB	11/10/25 14:17		rubina	OK
12	PB170470BL	PB170470BL	MB	11/10/25 14:17		rubina	OK
13	PB170470BS	PB170470BS	LCS	11/10/25 14:17		rubina	OK
14	LOWPB170440	LOWPB170440	SAM	11/10/25 14:24		rubina	OK
15	HIGHPB170440	HIGHPB170440	SAM	11/10/25 14:24		rubina	OK
16	PB170442BL	PB170442BL	MB	11/10/25 14:24		rubina	OK
17	PB170442BS	PB170442BS	LCS	11/10/25 14:24		rubina	OK
18	Q3560-01	SY-10D	SAM	11/10/25 14:24		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137840

Review By	rubina	Review On	11/11/2025 12:55:24 PM
Supervise By	Iwona	Supervise On	11/11/2025 2:08:01 PM
SubDirectory	LB137840	Test	Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP115577,WP115578,WP115579,WP115580,WP115581,WP115582,WP115583		
ICV Standard	W3012		
CCV Standard	WP115578		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115585		

19	Q3560-01DUP	SY-10DDUP	DUP	11/10/25 14:32		rubina	OK
20	Q3560-01MS	SY-10DMS	MS	11/10/25 14:32		rubina	OK
21	Q3560-01MSD	SY-10DMSD	MSD	11/10/25 14:32		rubina	OK
22	CCV2	CCV2	CCV	11/10/25 14:32		rubina	OK
23	CCB2	CCB2	CCB	11/10/25 14:32		rubina	OK
24	Q3560-03	SY-10S	SAM	11/10/25 14:32		rubina	OK
25	Q3560-05	SY-10I	SAM	11/10/25 14:32		rubina	OK
26	Q3560-07	SY-12I	SAM	11/10/25 14:39		rubina	OK
27	Q3560-09	SY-12D	SAM	11/10/25 14:39		rubina	OK
28	Q3569-01	MW-2	SAM	11/10/25 14:39		rubina	OK
29	Q3569-02	MW-12	SAM	11/10/25 14:39		rubina	OK
30	Q3580-01	346	SAM	11/10/25 14:39		rubina	OK
31	Q3580-01DUP	346DUP	DUP	11/10/25 14:39		rubina	OK
32	Q3580-01MS	346MS	MS	11/10/25 14:39		rubina	OK
33	Q3580-01MSD	346MSD	MSD	11/10/25 14:40		rubina	OK
34	CCV3	CCV3	CCV	11/10/25 14:40		rubina	OK
35	CCB3	CCB3	CCB	11/10/25 14:47		rubina	OK
36	Q3586-02	SED-2	SAM	11/10/25 14:47		rubina	OK
37	Q3586-03	SED-3	SAM	11/10/25 14:47		rubina	OK
38	Q3587-01	NB-08-110725	SAM	11/10/25 14:47		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137840

Review By	rubina	Review On	11/11/2025 12:55:24 PM
Supervise By	Iwona	Supervise On	11/11/2025 2:08:01 PM
SubDirectory	LB137840	Test	Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP115577,WP115578,WP115579,WP115580,WP115581,WP115582,WP115583		
ICV Standard	W3012		
CCV Standard	WP115578		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115585		

39	Q3588-01	TR-06-11072025	SAM	11/10/25 14:47		rubina	OK
40	Q3588-03	TR-01-11072025	SAM	11/10/25 14:47		rubina	OK
41	Q3586-01	SED-1	SAM	11/10/25 14:47		rubina	OK
42	Q3582-04	RT4912	SAM	11/10/25 15:00		rubina	OK
43	Q3583-04	CHRT27080	SAM	11/10/25 15:05		rubina	OK
44	CCV4	CCV4	CCV	11/10/25 15:05		rubina	OK
45	CCB4	CCB4	CCB	11/10/25 15:05		rubina	OK

Instrument ID: WC PH METER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137843

Review By	jignesh	Review On	11/11/2025 10:24:38 AM
Supervise By	Iwona	Supervise On	11/11/2025 1:02:22 PM
SubDirectory	LB137843	Test	pH
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3178,W3093,W3191,W3217,W3161,W3200		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	11/10/25 13:33		jignesh	OK
2	CAL2	CAL2	CAL	11/10/25 13:35		jignesh	OK
3	CAL3	CAL3	CAL	11/10/25 13:36		jignesh	OK
4	ICV	ICV	ICV	11/10/25 13:37		jignesh	OK
5	CCV1	CCV1	CCV	11/10/25 13:39		jignesh	OK
6	Q3569-01	MW-2	SAM	11/10/25 13:44		jignesh	OK
7	Q3569-02	MW-12	SAM	11/10/25 13:47		jignesh	OK
8	Q3574-01	304641-01 Liquid	SAM	11/10/25 13:55		jignesh	OK
9	Q3578-01	MH-1172025	SAM	11/10/25 14:00		jignesh	OK
10	Q3584-01	SW-1	SAM	11/10/25 14:10		jignesh	OK
11	Q3584-02	SW-2	SAM	11/10/25 14:15		jignesh	OK
12	Q3584-03	SW-3	SAM	11/10/25 14:22		jignesh	OK
13	Q3584-04	EB-2	SAM	11/10/25 14:25		jignesh	OK
14	Q3584-05	FB-2	SAM	11/10/25 14:30		jignesh	OK
15	Q3584-07	SEEP-1	SAM	11/10/25 14:37		jignesh	OK
16	CCV2	CCV2	CCV	11/10/25 14:40		jignesh	OK
17	Q3595-01	LAW-25-0176	SAM	11/10/25 14:48		jignesh	OK
18	Q3595-01DUP	LAW-25-0176DUP	DUP	11/10/25 14:50		jignesh	OK

Instrument ID: WC PH METER-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137843

Review By	jignesh	Review On	11/11/2025 10:24:38 AM
Supervise By	Iwona	Supervise On	11/11/2025 1:02:22 PM
SubDirectory	LB137843	Test	pH
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3178,W3093,W3191,W3217,W3161,W3200		

19	CCV3	CCV3	CCV	11/10/25 15:00		jignesh	OK
----	------	------	-----	----------------	--	---------	----

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB137845

Review By	jignesh	Review On	11/11/2025 12:57:35 PM
Supervise By	Iwona	Supervise On	11/11/2025 1:06:25 PM
SubDirectory	LB137845	Test	TDS
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	LB137845BL	LB137845BL	MB	11/10/25 11:30		jignesh	OK
2	LB137845BS	LB137845BS	LCS	11/10/25 11:30	WP115591	jignesh	OK
3	Q3569-01	MW-2	SAM	11/10/25 11:30		jignesh	OK
4	Q3569-02	MW-12	SAM	11/10/25 11:30		jignesh	OK
5	Q3584-01	SW-1	SAM	11/10/25 11:30		jignesh	OK
6	Q3584-02	SW-2	SAM	11/10/25 11:30		jignesh	OK
7	Q3584-03	SW-3	SAM	11/10/25 11:30		jignesh	OK
8	Q3584-03DUP	SW-3DUP	DUP	11/10/25 11:30		jignesh	OK
9	Q3584-07	SEEP-1	SAM	11/10/25 11:30		jignesh	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137870

Review By	rubina	Review On	11/12/2025 5:30:38 PM
Supervise By	Iwona	Supervise On	11/13/2025 1:15:12 PM
SubDirectory	LB137870	Test	TKN
STD. NAME	STD REF.#		
ICAL Standard	WP115592		
ICV Standard	WP115594		
CCV Standard	WP115593		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115595		
Chk Standard	WP115290,WP114133,WP113929,WP114132,WP115596,WP115597		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPM	0.0PPM	CAL1	11/12/25 10:23		rubina	OK
2	0.5PPM	0.5PPM	CAL2	11/12/25 10:23		rubina	OK
3	1.0PPM	1.0PPM	CAL3	11/12/25 10:23		rubina	OK
4	2.5PPM	2.5PPM	CAL4	11/12/25 10:23		rubina	OK
5	5.0PPM	5.0PPM	CAL5	11/12/25 10:23		rubina	OK
6	6.7PPM	6.7PPM	CAL6	11/12/25 10:23		rubina	OK
7	10.0PPM	10.0PPM	CAL7	11/12/25 10:23		rubina	OK
8	ICV1	ICV1	ICV	11/12/25 11:32		rubina	OK
9	ICB1	ICB1	ICB	11/12/25 11:32		rubina	OK
10	CCV1	CCV1	CCV	11/12/25 11:32		rubina	OK
11	CCB1	CCB1	CCB	11/12/25 11:32		rubina	OK
12	RL	RL	LOQ	11/12/25 11:32		rubina	OK
13	PB170510BL	PB170510BL	MB	11/12/25 11:32		rubina	OK
14	PB170510BS	PB170510BS	LCS	11/12/25 11:43		rubina	OK
15	Q3530-07	LOD-MDL-WATER-01	SAM	11/12/25 11:43		rubina	OK
16	Q3530-08	LOQ-WATER-02-QT4	LOQ	11/12/25 11:43		rubina	OK
17	Q3552-01	MW-4	SAM	11/12/25 11:43	TKN high, need dilution	rubina	Dilution
18	Q3552-02	MW-8	SAM	11/12/25 11:43	TKN high, need dilution	rubina	Dilution

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137870

Review By	rubina	Review On	11/12/2025 5:30:38 PM
Supervise By	Iwona	Supervise On	11/13/2025 1:15:12 PM
SubDirectory	LB137870	Test	TKN
STD. NAME	STD REF.#		
ICAL Standard	WP115592		
ICV Standard	WP115594		
CCV Standard	WP115593		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115595		
Chk Standard	WP115290,WP114133,WP113929,WP114132,WP115596,WP115597		

19	Q3552-03	MW-9	SAM	11/12/25 11:43	TKN high, need dilution	rubina	Dilution
20	Q3552-04	MW-10	SAM	11/12/25 11:43		rubina	OK
21	Q3552-04DUP	MW-10DUP	DUP	11/12/25 11:43		rubina	OK
22	CCV2	CCV2	CCV	11/12/25 11:54		rubina	OK
23	CCB2	CCB2	CCB	11/12/25 11:54		rubina	OK
24	Q3552-04MS	MW-10MS	MS	11/12/25 11:54		rubina	OK
25	Q3552-04MSD	MW-10MSD	MSD	11/12/25 11:54		rubina	OK
26	Q3552-05	MW-3	SAM	11/12/25 11:54	TKN high, need dilution	rubina	Dilution
27	Q3560-01	SY-10D	SAM	11/12/25 11:54		rubina	OK
28	Q3560-03	SY-10S	SAM	11/12/25 11:54		rubina	OK
29	Q3560-05	SY-10I	SAM	11/12/25 11:54		rubina	OK
30	Q3560-07	SY-12I	SAM	11/12/25 11:54	TKN high, need dilution	rubina	Dilution
31	Q3560-09	SY-12D	SAM	11/12/25 12:04		rubina	OK
32	Q3569-01	MW-2	SAM	11/12/25 12:04	TKN high, need dilution	rubina	Dilution
33	Q3569-02	MW-12	SAM	11/12/25 12:04	TKN high, need dilution	rubina	Dilution
34	CCV3	CCV3	CCV	11/12/25 12:04		rubina	OK
35	CCB3	CCB3	CCB	11/12/25 12:04		rubina	OK
36	Q3584-01	SW-1	SAM	11/12/25 12:04		rubina	OK
37	Q3584-02	SW-2	SAM	11/12/25 12:04		rubina	OK
38	Q3584-03	SW-3	SAM	11/12/25 12:04		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137870

Review By	rubina	Review On	11/12/2025 5:30:38 PM
Supervise By	Iwona	Supervise On	11/13/2025 1:15:12 PM
SubDirectory	LB137870	Test	TKN
STD. NAME	STD REF.#		
ICAL Standard	WP115592		
ICV Standard	WP115594		
CCV Standard	WP115593		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115595		
Chk Standard	WP115290,WP114133,WP113929,WP114132,WP115596,WP115597		

39	Q3584-07	SEEP-1	SAM	11/12/25 12:04	TKN high, need dilution	rubina	Dilution
40	PB170482BL	PB170482BL	MB	11/12/25 12:05		rubina	OK
41	PB170482BS	PB170482BS	LCS	11/12/25 12:05		rubina	OK
42	Q3483-07	HW1025-PT-NUT-SO	SAM	11/12/25 12:05	TKN high, need dilution	rubina	Dilution
43	Q3530-01	LOD-MDL-SOIL-01-Q	SAM	11/12/25 12:15		rubina	OK
44	Q3530-02	LOQ-SOIL-02-QT4-20	LOQ	11/12/25 12:15		rubina	OK
45	Q3586-02	SED-2	SAM	11/12/25 12:15	TKN high, need dilution	rubina	Dilution
46	CCV4	CCV4	CCV	11/12/25 12:15		rubina	OK
47	CCB4	CCB4	CCB	11/12/25 12:15		rubina	OK
48	Q3586-02DUP	SED-2DUP	DUP	11/12/25 12:15	TKN high, need dilution	rubina	Dilution
49	Q3586-02MS	SED-2MS	MS	11/12/25 12:15		rubina	OK
50	Q3586-02MSD	SED-2MSD	MSD	11/12/25 12:15		rubina	OK
51	Q3552-01DL	MW-4DL	SAM	11/12/25 12:15	20X For TKN	rubina	Confirms
52	Q3552-02DL	MW-8DL	SAM	11/12/25 12:15	2X For TKN	rubina	Confirms
53	Q3552-03DL	MW-9DL	SAM	11/12/25 12:15	10X For TKN	rubina	Confirms
54	Q3552-05DL	MW-3DL	SAM	11/12/25 12:26	10X For TKN	rubina	Confirms
55	Q3560-07DL	SY-12IDL	SAM	11/12/25 12:26	10X For TKN	rubina	Confirms
56	Q3569-01DL	MW-2DL	SAM	11/12/25 12:26	10X For TKN	rubina	Confirms
57	CCV5	CCV5	CCV	11/12/25 13:42		rubina	OK
58	CCB5	CCB5	CCB	11/12/25 13:42		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QCBatch ID # LB137870

Review By	rubina	Review On	11/12/2025 5:30:38 PM
Supervise By	Iwona	Supervise On	11/13/2025 1:15:12 PM
SubDirectory	LB137870	Test	TKN
STD. NAME	STD REF.#		
ICAL Standard	WP115592		
ICV Standard	WP115594		
CCV Standard	WP115593		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115595		
Chk Standard	WP115290,WP114133,WP113929,WP114132,WP115596,WP115597		

59	Q3569-02DL	MW-12DL	SAM	11/12/25 13:42	10X For TKN	rubina	Confirms
60	Q3483-07DL	HW1025-PT-NUT-SO	SAM	11/12/25 13:42	10X For TKN	rubina	Confirms
61	Q3586-02DL	SED-2DL	SAM	11/12/25 13:42	5X For TKN	rubina	Confirms
62	Q3586-02DUPDL	SED-2DUPDL	DUP	11/12/25 13:42	5X For TKN	rubina	Confirms
63	Q3584-07DL	SEEP-1DL	SAM	11/12/25 13:42	20X For TKN	rubina	Confirms
64	CCV6	CCV6	CCV	11/12/25 13:42		rubina	OK
65	CCB6	CCB6	CCB	11/12/25 13:45		rubina	OK

Instrument ID: DO METER

Daily Analysis Runlog For Sequence/QC Batch ID # LB137873

Review By	rubina	Review On	11/12/2025 5:01:53 PM
Supervise By	Iwona	Supervise On	11/12/2025 5:02:07 PM
SubDirectory	LB137873	Test	BOD5
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP115513,W3149,WP115342,W3103,W3109,W3248,WP115515,WP115514,WP113878		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB137873BL	LB137873BL	MB	11/07/25 11:30		rubina	OK
2	LB137873BS	LB137873BS	LCS	11/07/25 11:30		rubina	OK
3	Q3569-01	MW-2	SAM	11/07/25 11:30		rubina	OK
4	Q3569-01DUP	MW-2DUP	DUP	11/07/25 11:30		rubina	OK
5	Q3569-02	MW-12	SAM	11/07/25 11:30		rubina	OK

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB137877

Review By	jignesh	Review On	11/13/2025 3:41:31 PM
Supervise By	Iwona	Supervise On	11/13/2025 4:58:34 PM
SubDirectory	LB137877	Test	Oil and Grease
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3240,M6069,EP2655,WP115016,N/A,N/A,WP115017,N/A,WP115018		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB137877BL	LB137877BL	MB	11/13/25 10:00		jignesh	OK
2	LB137877BS	LB137877BS	LCS	11/13/25 10:00		jignesh	OK
3	Q3530-09	MDL-WATER-03-QT4	SAM	11/13/25 10:00		jignesh	OK
4	Q3569-01	MW-2	SAM	11/13/25 10:00		jignesh	OK
5	Q3569-02	MW-12	SAM	11/13/25 10:00		jignesh	OK
6	Q3575-01	001 Willets Pt Blvd (N	SAM	11/13/25 10:00		jignesh	OK
7	Q3575-02	002 35th Ave (Nov)	SAM	11/13/25 10:00		jignesh	OK
8	Q3578-01	MH-1172025	SAM	11/13/25 10:00		jignesh	OK
9	Q3578-02	Q3578-01MS	MS	11/13/25 10:00		jignesh	OK
10	Q3578-03	Q3578-01MSD	MSD	11/13/25 10:00		jignesh	OK
11	Q3584-07	SEEP-1	SAM	11/13/25 10:00		jignesh	OK
12	Q3616-01	OIL AND GREASE-1	SAM	11/13/25 10:00		jignesh	OK
13	Q3616-02	OIL AND GREASE-2	SAM	11/13/25 10:00		jignesh	OK
14	Q3616-03	OIL AND GREASE-3	SAM	11/13/25 10:00		jignesh	OK

Instrument ID: SPECTROPHOTOMETER-2

Daily Analysis Runlog For Sequence/QC Batch ID # LB137904

Review By	Iwona	Review On	11/14/2025 2:58:02 PM
Supervise By	JIGNESH	Supervise On	11/14/2025 3:20:04 PM
SubDirectory	LB137904	Test	COD
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP114659,WP114658,WP114656,WP114655,WP114654,WP114661,WP114657,WP115676,WP115677,WP115678,V		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	CAL1	CAL1	CAL	09/08/25 13:30		Iwona	OK
2	CAL2	CAL2	CAL	09/08/25 13:30		Iwona	OK
3	CAL3	CAL3	CAL	09/08/25 13:31		Iwona	OK
4	CAL4	CAL4	CAL	09/08/25 13:31		Iwona	OK
5	CAL5	CAL5	CAL	09/08/25 13:32		Iwona	OK
6	CAL6	CAL6	CAL	09/08/25 13:32		Iwona	OK
7	ICV	ICV	ICV	09/08/25 13:33		Iwona	OK
8	ICB	ICB	ICB	09/08/25 13:33		Iwona	OK
9	CCV1	CCV1	CCV	11/14/25 14:00		Iwona	OK
10	CCB1	CCB1	CCB	11/14/25 14:00		Iwona	OK
11	RL Check	RL Check	RL	11/14/25 14:01		Iwona	OK
12	LB137904BL	LB137904BL	MB	11/14/25 14:01		Iwona	OK
13	LB137904BS	LB137904BS	LCS	11/14/25 14:02		Iwona	OK
14	Q3551-04	MANHOLE	SAM	11/14/25 14:02		Iwona	OK
15	Q3552-01	MW-4	SAM	11/14/25 14:03		Iwona	OK
16	Q3552-02	MW-8	SAM	11/14/25 14:03		Iwona	OK
17	Q3552-03	MW-9	SAM	11/14/25 14:04		Iwona	OK
18	Q3552-04	MW-10	SAM	11/14/25 14:04		Iwona	OK

Instrument ID: SPECTROPHOTOMETER-2

Daily Analysis Runlog For Sequence/QC Batch ID # LB137904

Review By	Iwona	Review On	11/14/2025 2:58:02 PM
Supervise By	JIGNESH	Supervise On	11/14/2025 3:20:04 PM
SubDirectory	LB137904	Test	COD
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP114659,WP114658,WP114656,WP114655,WP114654,WP114661,WP114657,WP115676,WP115677,WP115678,V		

19	Q3552-05	MW-3	SAM	11/14/25 14:05		Iwona	OK
20	Q3569-01	MW-2	SAM	11/14/25 14:05		Iwona	OK
21	Q3569-02	MW-12	SAM	11/14/25 14:06		Iwona	OK
22	CCV2	CCV2	CCV	11/14/25 14:06		Iwona	OK
23	CCB2	CCB2	CCB	11/14/25 14:07		Iwona	OK
24	Q3578-01	MH-1172025	SAM	11/14/25 14:07		Iwona	OK
25	Q3584-01	SW-1	SAM	11/14/25 14:08		Iwona	OK
26	Q3584-02	SW-2	SAM	11/14/25 14:08		Iwona	OK
27	Q3584-03	SW-3	SAM	11/14/25 14:09		Iwona	OK
28	Q3584-07	SEEP-1	SAM	11/14/25 14:09		Iwona	OK
29	Q3630-01	DSN002	SAM	11/14/25 14:10		Iwona	OK
30	Q3630-03	DSN001	SAM	11/14/25 14:10		Iwona	OK
31	Q3630-04	DSN001	SAM	11/14/25 14:11		Iwona	OK
32	Q3630-05	DSN003	SAM	11/14/25 14:11		Iwona	OK
33	Q3630-05DUP	DSN003DUP	DUP	11/14/25 14:12		Iwona	OK
34	CCV3	CCV3	CCV	11/14/25 14:12		Iwona	OK
35	CCB3	CCB3	CCB	11/14/25 14:13		Iwona	OK
36	Q3630-05MS	DSN003MS	MS	11/14/25 14:13	0.5ml WP115676 + 9.5ml Sample	Iwona	OK
37	Q3630-05MSD	DSN003MSD	MSD	11/14/25 14:13	0.5ml WP115676 + 9.5ml Sample	Iwona	OK

Instrument ID: SPECTROPHOTOMETER-2

Daily Analysis Runlog For Sequence/QC Batch ID # LB137904

Review By	Iwona	Review On	11/14/2025 2:58:02 PM
Supervise By	JIGNESH	Supervise On	11/14/2025 3:20:04 PM
SubDirectory	LB137904	Test	COD
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP114659,WP114658,WP114656,WP114655,WP114654,WP114661,WP114657,WP115676,WP115677,WP115678,V		

38	CCV4	CCV4	CCV	11/14/25 14:14		Iwona	OK
39	CCB4	CCB4	CCB	11/14/25 14:14		Iwona	OK