

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3569

Project: Edison Landfill

RunNo.: LB137803

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.7	10	97	90-110	11/03/2025
Chloride	mg/L	2.9	3	97	90-110	11/03/2025
Fluoride	mg/L	2	2	100	90-110	11/03/2025
Nitrite	mg/L	2.9	3	97	90-110	11/03/2025
Nitrate	mg/L	2.4	2.5	96	90-110	11/03/2025
Sulfate	mg/L	14.5	15	97	90-110	11/03/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	11/03/2025
Sample ID: CCV1						
Bromide	mg/L	10.1	10	101	90-110	11/06/2025
Chloride	mg/L	3	3	100	90-110	11/06/2025
Fluoride	mg/L	2	2	100	90-110	11/06/2025
Nitrite	mg/L	3	3	100	90-110	11/06/2025
Nitrate	mg/L	2.5	2.5	100	90-110	11/06/2025
Sulfate	mg/L	15.1	15	101	90-110	11/06/2025
Orthophosphate as P	mg/L	4.6	5	92	90-110	11/06/2025
Sample ID: CCV2						
Bromide	mg/L	10.2	10	102	90-110	11/06/2025
Chloride	mg/L	3.1	3	103	90-110	11/06/2025
Fluoride	mg/L	2	2	100	90-110	11/06/2025
Nitrite	mg/L	3	3	100	90-110	11/06/2025
Nitrate	mg/L	2.5	2.5	100	90-110	11/06/2025
Sulfate	mg/L	15.2	15	101	90-110	11/06/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	11/06/2025
Sample ID: CCV3						
Bromide	mg/L	10.2	10	102	90-110	11/06/2025
Chloride	mg/L	3.1	3	103	90-110	11/06/2025
Fluoride	mg/L	2	2	100	90-110	11/06/2025
Nitrite	mg/L	3	3	100	90-110	11/06/2025
Nitrate	mg/L	2.5	2.5	100	90-110	11/06/2025
Sulfate	mg/L	15.1	15	101	90-110	11/06/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	11/06/2025
Sample ID: CCV4						
Bromide	mg/L	10.2	10	102	90-110	11/07/2025
Chloride	mg/L	3.1	3	103	90-110	11/07/2025
Fluoride	mg/L	2.1	2	105	90-110	11/07/2025
Nitrite	mg/L	3.1	3	103	90-110	11/07/2025
Nitrate	mg/L	2.5	2.5	100	90-110	11/07/2025
Sulfate	mg/L	15.1	15	101	90-110	11/07/2025

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3569

Project: Edison Landfill

RunNo.: LB137803

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Orthophosphate as P	mg/L	5	5	100	90-110	11/07/2025
Sample ID: CCV5						
Bromide	mg/L	10.3	10	103	90-110	11/07/2025
Chloride	mg/L	3.1	3	103	90-110	11/07/2025
Fluoride	mg/L	2.1	2	105	90-110	11/07/2025
Nitrite	mg/L	3.1	3	103	90-110	11/07/2025
Nitrate	mg/L	2.6	2.5	104	90-110	11/07/2025
Sulfate	mg/L	15.2	15	101	90-110	11/07/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	11/07/2025
Sample ID: CCV6						
Bromide	mg/L	10.3	10	103	90-110	11/07/2025
Chloride	mg/L	3.1	3	103	90-110	11/07/2025
Fluoride	mg/L	2.1	2	105	90-110	11/07/2025
Nitrite	mg/L	3.1	3	103	90-110	11/07/2025
Nitrate	mg/L	2.6	2.5	104	90-110	11/07/2025
Sulfate	mg/L	15.2	15	101	90-110	11/07/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	11/07/2025

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3569

Project: Edison Landfill

RunNo.: LB137821

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	11/07/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	11/07/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	11/07/2025

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3569

Project: Edison Landfill

RunNo.: LB137840

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.096	0.099	97	90-110	11/10/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	11/10/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	11/10/2025
Sample ID: Cyanide	CCV3	mg/L	0.24	0.25	96	90-110	11/10/2025
Sample ID: Cyanide	CCV4	mg/L	0.25	0.25	100	90-110	11/10/2025

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3569

Project: Edison Landfill

RunNo.: LB137843

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.02	7	100	90-110	11/10/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	11/10/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	11/10/2025
Sample ID:	CCV3						
pH		pH	2.01	2.00	101	90-110	11/10/2025

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3569

Project: Edison Landfill

RunNo.: LB137870

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TKN	mg/L	5.1	5	102	90-110	11/12/2025
Sample ID: CCV1 TKN	mg/L	5	5	100	90-110	11/12/2025
Sample ID: CCV2 TKN	mg/L	5	5	100	90-110	11/12/2025
Sample ID: CCV3 TKN	mg/L	5.1	5	102	90-110	11/12/2025
Sample ID: CCV4 TKN	mg/L	5.1	5	102	90-110	11/12/2025
Sample ID: CCV5 TKN	mg/L	5	5	100	90-110	11/12/2025
Sample ID: CCV6 TKN	mg/L	5.2	5	104	90-110	11/12/2025

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3569

Project: Edison Landfill

RunNo.: LB137904

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	51.184	50	102	95-105	11/14/2025
Sample ID: CCV2 COD	mg/L	48.150	50	96	95-105	11/14/2025
Sample ID: CCV3 COD	mg/L	48.150	50	96	95-105	11/14/2025
Sample ID: CCV4 COD	mg/L	49.161	50	98	95-105	11/14/2025