

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3569
Project: Edison Landfill	Sample ID: Q3551-01
Client ID: MANHOLEMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Chloride	mg/L	+/-20	166	OR	166	OR	1	0		11/06/2025
Sulfate	mg/L	+/-20	40.4	OR	41.5	OR	1	3		11/06/2025
Orthophosphate as P	mg/L	+/-20	10.4		10.0		1	4		11/06/2025
Nitrite	mg/L	+/-20	2.90		3.10		1	7		11/06/2025
Bromide	mg/L	+/-20	9.50		10.3		1	8		11/06/2025
Fluoride	mg/L	+/-20	2.40		2.60		1	8		11/06/2025
Nitrate	mg/L	+/-20	2.50		2.70		1	8		11/06/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3569
Project:	Edison Landfill	Sample ID:	Q3552-04
Client ID:	MW-10DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TKN	mg/L	+/-20	2.90		2.70		1	7		11/12/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3569
Project:	Edison Landfill	Sample ID:	Q3552-04
Client ID:	MW-10MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TKN	mg/L	+/-20	7.30		7.20		1	1		11/12/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3569
Project: Edison Landfill	Sample ID: Q3560-01
Client ID: SY-10DDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.047	U	0.047	U	1	0		11/10/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3569
Project: Edison Landfill	Sample ID: Q3560-01
Client ID: SY-10DMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	1.90		2.00		1	5		11/10/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Edison Landfill Client ID: MW-2DUP	SDG No.: Q3569 Sample ID: Q3569-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		11/07/2025
BOD5	mg/L	+/-20	50.0		49.2		1	1.73		11/07/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers	SDG No.: Q3569
Project: Edison Landfill	Sample ID: Q3569-01
Client ID: MW-2MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.94		0.94		2	0.43		11/07/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Edison Landfill Client ID: MH-1172025MSD	SDG No.: Q3569 Sample ID: Q3578-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	115		113		1	1.23		11/13/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3569
Project:	Edison Landfill	Sample ID:	Q3584-03
Client ID:	SW-3DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	12500		13000		1	3.22		11/10/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3569
Project:	Edison Landfill	Sample ID:	Q3595-01
Client ID:	LAW-25-0176DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	7.13		7.15		1	0.28		11/10/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3569
Project:	Edison Landfill	Sample ID:	Q3630-05
Client ID:	DSN003DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	23.9		22.9		1	4.27		11/14/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3569
Project:	Edison Landfill	Sample ID:	Q3630-05
Client ID:	DSN003MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	68.4		67.4		1	1.47		11/14/2025