



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB108	Mercury	0.076	+/-0.2	U	0.20	CV	11/10/2025	14:32	LB137835

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB53	Mercury	0.076	+/-0.2	U	0.20	CV	11/10/2025	14:36	LB137835
CCB54	Mercury	0.076	+/-0.2	U	0.20	CV	11/10/2025	15:04	LB137835
CCB55	Mercury	0.076	+/-0.2	U	0.20	CV	11/11/2025	08:38	LB137835

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	11/10/2025	12:11	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	12:11	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	12:11	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	12:11	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	12:11	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	12:11	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	12:11	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	12:11	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	12:11	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	12:11	LB137841
	Iron	32.9	+/-50	J	100	P	11/10/2025	12:11	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	12:11	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	12:11	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	12:11	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	12:11	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	12:11	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	12:11	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	12:11	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	12:11	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	12:11	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	12:11	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	12:11	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	14.6	+/-50	J	100	P	11/10/2025	12:42	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	12:42	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	12:42	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	12:42	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	12:42	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	12:42	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	12:42	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	12:42	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	12:42	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	12:42	LB137841
	Iron	23.4	+/-50	U	100	P	11/10/2025	12:42	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	12:42	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	12:42	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	12:42	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	12:42	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	12:42	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	12:42	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	12:42	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	12:42	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	12:42	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	12:42	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	12:42	LB137841
CCB02	Aluminum	20.4	+/-50	J	100	P	11/10/2025	13:54	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	13:54	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	13:54	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	13:54	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	13:54	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	13:54	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	13:54	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	13:54	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	13:54	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	13:54	LB137841
	Iron	29.8	+/-50	J	100	P	11/10/2025	13:54	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	13:54	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	13:54	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	13:54	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	13:54	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	13:54	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	13:54	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/10/2025	13:54	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	13:54	LB137841
	Thallium	4.43	+/-20	J	40.0	P	11/10/2025	13:54	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	13:54	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	13:54	LB137841
CCB03	Aluminum	11.3	+/-50	U	100	P	11/10/2025	14:50	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	14:50	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	14:50	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	14:50	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	14:50	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	14:50	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	14:50	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	14:50	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	14:50	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	14:50	LB137841
	Iron	29.3	+/-50	J	100	P	11/10/2025	14:50	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	14:50	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	14:50	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	14:50	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	14:50	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	14:50	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	14:50	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	14:50	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	14:50	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	14:50	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	14:50	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	14:50	LB137841
CCB04	Aluminum	37.9	+/-50	J	100	P	11/10/2025	15:47	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	15:47	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	15:47	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	15:47	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	15:47	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	15:47	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	15:47	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	15:47	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	15:47	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	15:47	LB137841
	Iron	37.7	+/-50	J	100	P	11/10/2025	15:47	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	15:47	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/10/2025	15:47	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	15:47	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	15:47	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	15:47	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	15:47	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	15:47	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	15:47	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	15:47	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	15:47	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	15:47	LB137841
CCB05	Aluminum	19.1	+/-50	J	100	P	11/10/2025	16:38	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	16:38	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	16:38	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	16:38	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	16:38	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	16:38	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	16:38	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	16:38	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	16:38	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	16:38	LB137841
	Iron	24.8	+/-50	J	100	P	11/10/2025	16:38	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	16:38	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	16:38	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	16:38	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	16:38	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	16:38	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	16:38	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	16:38	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	16:38	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	16:38	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	16:38	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	16:38	LB137841
CCB06	Aluminum	23.5	+/-50	J	100	P	11/10/2025	17:31	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	17:31	LB137841
	Arsenic	7.21	+/-10	J	20.0	P	11/10/2025	17:31	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	17:31	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	17:31	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	17:31	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	17:31	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	17:31	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	17:31	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	17:31	LB137841
	Iron	24.6	+/-50	J	100	P	11/10/2025	17:31	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	17:31	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	17:31	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	17:31	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	17:31	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	17:31	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	17:31	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	17:31	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	17:31	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	17:31	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	17:31	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	17:31	LB137841
CCB07	Aluminum	18.8	+/-50	J	100	P	11/10/2025	19:18	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	19:18	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	19:18	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	19:18	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	19:18	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	19:18	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	19:18	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	19:18	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	19:18	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	19:18	LB137841
	Iron	38.0	+/-50	J	100	P	11/10/2025	19:18	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	19:18	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	19:18	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	19:18	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	19:18	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	19:18	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	19:18	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	19:18	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	19:18	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	19:18	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	19:18	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	19:18	LB137841
CCB08	Aluminum	49.0	+/-50	J	100	P	11/10/2025	20:26	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	20:26	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	20:26	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	20:26	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	20:26	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	20:26	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	20:26	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	20:26	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	20:26	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	20:26	LB137841
	Iron	27.2	+/-50	J	100	P	11/10/2025	20:26	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	20:26	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	20:26	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	20:26	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	20:26	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	20:26	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	20:26	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	20:26	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	20:26	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	20:26	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	20:26	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	20:26	LB137841
CCB09	Aluminum	11.3	+/-50	U	100	P	11/10/2025	21:41	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	21:41	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	21:41	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	21:41	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	21:41	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	21:41	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	21:41	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	21:41	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	21:41	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	21:41	LB137841
	Iron	23.4	+/-50	U	100	P	11/10/2025	21:41	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	21:41	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	21:41	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	21:41	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	21:41	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	21:41	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	21:41	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	21:41	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	21:41	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	21:41	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	21:41	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	21:41	LB137841
CCB10	Aluminum	21.4	+/-50	J	100	P	11/10/2025	22:47	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	22:47	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	22:47	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	22:47	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	22:47	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	22:47	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	22:47	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	22:47	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	22:47	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	22:47	LB137841
	Iron	23.4	+/-50	U	100	P	11/10/2025	22:47	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	22:47	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	22:47	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	22:47	LB137841
	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	22:47	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	22:47	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	22:47	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	22:47	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	22:47	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	22:47	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	22:47	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	22:47	LB137841
	Aluminum	39.4	+/-50	J	100	P	11/10/2025	23:34	LB137841
	Antimony	6.76	+/-25	U	50.0	P	11/10/2025	23:34	LB137841
	Arsenic	5.12	+/-10	U	20.0	P	11/10/2025	23:34	LB137841
	Barium	14.6	+/-50	U	100	P	11/10/2025	23:34	LB137841
	Beryllium	0.56	+/-3	U	6.00	P	11/10/2025	23:34	LB137841
	Cadmium	0.50	+/-3	U	6.00	P	11/10/2025	23:34	LB137841
	Calcium	234	+/-1000	U	2000	P	11/10/2025	23:34	LB137841
	Chromium	2.12	+/-5	U	10.0	P	11/10/2025	23:34	LB137841
	Cobalt	2.26	+/-15	U	30.0	P	11/10/2025	23:34	LB137841
	Copper	4.60	+/-10	U	20.0	P	11/10/2025	23:34	LB137841
	Iron	23.4	+/-50	U	100	P	11/10/2025	23:34	LB137841
	Lead	2.30	+/-6	U	12.0	P	11/10/2025	23:34	LB137841
	Magnesium	244	+/-1000	U	2000	P	11/10/2025	23:34	LB137841
	Manganese	5.94	+/-10	U	20.0	P	11/10/2025	23:34	LB137841

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	3.06	+/-20	U	40.0	P	11/10/2025	23:34	LB137841
	Potassium	918	+/-1000	U	2000	P	11/10/2025	23:34	LB137841
	Selenium	9.64	+/-10	U	20.0	P	11/10/2025	23:34	LB137841
	Silver	1.62	+/-5	U	10.0	P	11/10/2025	23:34	LB137841
	Sodium	868	+/-1000	U	2000	P	11/10/2025	23:34	LB137841
	Thallium	4.38	+/-20	U	40.0	P	11/10/2025	23:34	LB137841
	Vanadium	6.26	+/-20	U	40.0	P	11/10/2025	23:34	LB137841
	Zinc	16.7	+/-20	U	40.0	P	11/10/2025	23:34	LB137841