

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Remington & Vernick Engineers</u>	SDG No.: <u>Q3569</u>
Contract: <u>REMI02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	247000	255000	97	216000	294000	11/10/2025	12:21	LB137841
	Antimony	1.44			-50	50	11/10/2025	12:21	LB137841
	Arsenic	4.11			-20	20	11/10/2025	12:21	LB137841
	Barium	1.49	6.0	25	-94	106	11/10/2025	12:21	LB137841
	Beryllium	1.48			-6	6	11/10/2025	12:21	LB137841
	Cadmium	0.069	1.0	7	-5	7	11/10/2025	12:21	LB137841
	Calcium	238000	245000	97	208000	282000	11/10/2025	12:21	LB137841
	Chromium	49.5	52.0	95	42	62	11/10/2025	12:21	LB137841
	Cobalt	1.52			-30	30	11/10/2025	12:21	LB137841
	Copper	18.7	2.0	935	-18	22	11/10/2025	12:21	LB137841
	Iron	103000	101000	102	85600	116500	11/10/2025	12:21	LB137841
	Lead	-2.05			-12	12	11/10/2025	12:21	LB137841
	Magnesium	249000	255000	98	216000	294000	11/10/2025	12:21	LB137841
	Manganese	8.79	7.0	126	-13	27	11/10/2025	12:21	LB137841
	Nickel	6.36	2.0	318	-38	42	11/10/2025	12:21	LB137841
	Potassium	291			0	0	11/10/2025	12:21	LB137841
	Selenium	3.23			-20	20	11/10/2025	12:21	LB137841
	Silver	0.36			-10	10	11/10/2025	12:21	LB137841
	Sodium	-245			0	0	11/10/2025	12:21	LB137841
	Thallium	5.19			-40	40	11/10/2025	12:21	LB137841
	Vanadium	-3.12			-40	40	11/10/2025	12:21	LB137841
	Zinc	3.40			-40	40	11/10/2025	12:21	LB137841
ICSAB01	Aluminum	250000	247000	101	209000	285000	11/10/2025	12:26	LB137841
	Antimony	608	618	98	525	711	11/10/2025	12:26	LB137841
	Arsenic	101	104	97	88.4	120	11/10/2025	12:26	LB137841
	Barium	497	537	93	437	637	11/10/2025	12:26	LB137841
	Beryllium	515	495	104	420	570	11/10/2025	12:26	LB137841
	Cadmium	996	972	102	826	1120	11/10/2025	12:26	LB137841
	Calcium	239000	235000	102	199000	271000	11/10/2025	12:26	LB137841
	Chromium	584	542	108	460	624	11/10/2025	12:26	LB137841
	Cobalt	505	476	106	404	548	11/10/2025	12:26	LB137841
	Copper	501	511	98	434	588	11/10/2025	12:26	LB137841
	Iron	103000	99300	104	84400	114500	11/10/2025	12:26	LB137841
	Lead	49.4	49.0	101	37	61	11/10/2025	12:26	LB137841
	Magnesium	250000	248000	101	210000	286000	11/10/2025	12:26	LB137841
	Manganese	505	507	100	430	584	11/10/2025	12:26	LB137841
	Nickel	1020	954	107	810	1100	11/10/2025	12:26	LB137841
	Potassium	-9.62			0	0	11/10/2025	12:26	LB137841
	Selenium	51.0	46.0	111	26	66	11/10/2025	12:26	LB137841
	Silver	204	201	102	170	232	11/10/2025	12:26	LB137841
	Sodium	-230			0	0	11/10/2025	12:26	LB137841
	Thallium	98.9	108	92	68	148	11/10/2025	12:26	LB137841

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Remington & Vernick Engineers

SDG No.: Q3569

Contract: REMI02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	511	491	104	417	565	11/10/2025	12:26	LB137841
	Zinc	1030	952	108	809	1095	11/10/2025	12:26	LB137841