

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Remington & Vernick Engineers **SDG No.:** Q3569
Contract: REMI02 **Lab Code:** ACE
Matrix: Water **Level:** LOW **Client ID:** WATER-TREATMENT DISCHARGE A
Sample ID: Q3566-01 **Spiked ID:** Q3566-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	360		160		200	100		P
Iron	ug/L	75 - 125	1680		55.0		1500	108		P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Remington & Vernick Engineers **SDG No.:** Q3569
Contract: REMI02 **Lab Code:** ACE
Matrix: Water **Level:** LOW **Client ID:** SW-1A
Sample ID: Q3584-01 **Spiked ID:** Q3584-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	82 - 119	4.72		0.20	U	4.00	118		CV