

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3569

Continuing Calib Date: 11/11/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 10:42

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1248-2	(2)	5.76	5.76	5.66	5.86	0.00
Aroclor-1248-3	(3)	5.96	5.96	5.86	6.06	0.00
Aroclor-1248-4	(4)	6.35	6.35	6.25	6.45	0.00
Aroclor-1248-5	(5)	6.39	6.39	6.29	6.49	0.00
Tetrachloro-m-xylene		4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3569

Continuing Calib Date: 11/11/2025

Initial Calibration Date(s): 11/04/2025

11/04/2025

Continuing Calib Time: 10:42

Initial Calibration Time(s): 09:39

18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1248-2 (2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1248-3 (3)	5.00	5.01	4.91	5.11	0.01
Aroclor-1248-4 (4)	5.18	5.18	5.08	5.28	0.00
Aroclor-1248-5 (5)	5.57	5.57	5.47	5.67	0.00
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.64	8.64	8.54	8.74	0.00