

Report of Analysis

Client:	Remington & Vernick Engineers	Date Collected:	11/05/25
Project:	Edison Landfill	Date Received:	11/06/25
Client Sample ID:	MW-12	SDG No.:	Q3569
Lab Sample ID:	Q3569-02	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	475 mL	Test:	PCB
Prep Method:	3510C		
	Final Vol: 5000 uL		
	Prep Date: 11/11/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.10	U	1	0.10	0.53	ug/L	11/11/25 19:11	PB170484
11104-28-2	Aroclor-1221	0.14	U	1	0.14	0.53	ug/L	11/11/25 19:11	PB170484
11141-16-5	Aroclor-1232	0.10	U	1	0.10	0.53	ug/L	11/11/25 19:11	PB170484
53469-21-9	Aroclor-1242	0.13	U	1	0.13	0.53	ug/L	11/11/25 19:11	PB170484
12672-29-6	Aroclor-1248	0.075	U	1	0.075	0.53	ug/L	11/11/25 19:11	PB170484
11097-69-1	Aroclor-1254	0.099	U	1	0.099	0.53	ug/L	11/11/25 19:11	PB170484
37324-23-5	Aroclor-1262	0.15	U	1	0.15	0.53	ug/L	11/11/25 19:11	PB170484
11100-14-4	Aroclor-1268	0.12	U	1	0.12	0.53	ug/L	11/11/25 19:11	PB170484
11096-82-5	Aroclor-1260	0.085	U	1	0.085	0.53	ug/L	11/11/25 19:11	PB170484
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.9			30 (30) - 150 (173)	125%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.0			30 (10) - 150 (173)	110%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit