

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3569

Analytical Method: 8081B

Client: Remington & Vernick Engineers

Datafile : PD091093.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170485BSD (Column 1)	alpha-BHC	0.5	0.47	ug/L	93	2			40 (85)	140 (130)	20 (20)
	beta-BHC	0.5	0.45	ug/L	89	3			40 (83)	140 (126)	20 (20)
	delta-BHC	0.5	0.46	ug/L	92	3			40 (69)	140 (141)	20 (20)
	gamma-BHC (Lindane)	0.5	0.46	ug/L	92	2			40 (82)	140 (129)	20 (20)
	Heptachlor	0.5	0.55	ug/L	109	4			40 (79)	140 (127)	20 (20)
	Aldrin	0.5	0.45	ug/L	90	5			40 (79)	140 (126)	20 (20)
	Heptachlor epoxide	0.5	0.45	ug/L	90	7			40 (81)	140 (124)	20 (20)
	Endosulfan I	0.5	0.45	ug/L	89	7			40 (85)	140 (122)	20 (20)
	Dieldrin	0.5	0.43	ug/L	85	12			40 (83)	140 (125)	20 (20)
	4,4'-DDE	0.5	0.43	ug/L	87	7			40 (80)	140 (127)	20 (20)
	Endrin	0.5	0.41	ug/L	83	6			40 (81)	140 (128)	20 (20)
	Endosulfan II	0.5	0.44	ug/L	88	9			40 (82)	140 (123)	20 (20)
	4,4'-DDD	0.5	0.47	ug/L	95	8			40 (77)	140 (131)	20 (20)
	Endosulfan sulfate	0.5	0.44	ug/L	88	8			40 (76)	140 (129)	20 (20)
	4,4'-DDT	0.5	0.45	ug/L	90	5			40 (80)	140 (133)	20 (20)
	Methoxychlor	0.5	0.44	ug/L	88	9			40 (78)	140 (108)	20 (20)
	Endrin ketone	0.5	0.46	ug/L	93	7			40 (80)	140 (131)	20 (20)
	Endrin aldehyde	0.5	0.46	ug/L	91	8			40 (82)	140 (127)	20 (20)
	alpha-Chlordane	0.5	0.43	ug/L	86	11			40 (82)	140 (125)	20 (20)
	gamma-Chlordane	0.5	0.45	ug/L	90	4			40 (82)	140 (125)	20 (20)
PB170485BSD (Column 2)	alpha-BHC	0.5	0.51	ug/L	102	5			40 (85)	140 (130)	20 (20)
	beta-BHC	0.5	0.50	ug/L	99	6			40 (83)	140 (126)	20 (20)
	delta-BHC	0.5	0.50	ug/L	100	6			40 (69)	140 (141)	20 (20)
	gamma-BHC (Lindane)	0.5	0.51	ug/L	102	4			40 (82)	140 (129)	20 (20)
	Heptachlor	0.5	0.52	ug/L	103	7			40 (79)	140 (127)	20 (20)
	Aldrin	0.5	0.50	ug/L	100	6			40 (79)	140 (126)	20 (20)
	Heptachlor epoxide	0.5	0.48	ug/L	95	9			40 (81)	140 (124)	20 (20)
	Endosulfan I	0.5	0.48	ug/L	97	9			40 (85)	140 (122)	20 (20)
	Dieldrin	0.5	0.48	ug/L	95	13			40 (83)	140 (125)	20 (20)
	4,4'-DDE	0.5	0.48	ug/L	97	10			40 (80)	140 (127)	20 (20)
	Endrin	0.5	0.46	ug/L	91	9			40 (81)	140 (128)	20 (20)
	Endosulfan II	0.5	0.48	ug/L	97	12			40 (82)	140 (123)	20 (20)
	4,4'-DDD	0.5	0.50	ug/L	99	14			40 (77)	140 (131)	20 (20)
	Endosulfan sulfate	0.5	0.46	ug/L	93	17			40 (76)	140 (129)	20 (20)
	4,4'-DDT	0.5	0.47	ug/L	93	14			40 (80)	140 (133)	20 (20)
	Methoxychlor	0.5	0.45	ug/L	91	17			40 (78)	140 (108)	20 (20)
	Endrin ketone	0.5	0.49	ug/L	98	23			40 (80)	140 (131)	20 (20)
	Endrin aldehyde	0.5	0.48	ug/L	96	15			40 (82)	140 (127)	20 (20)
	alpha-Chlordane	0.5	0.48	ug/L	97	9			40 (82)	140 (125)	20 (20)
	gamma-Chlordane	0.5	0.49	ug/L	97	9			40 (82)	140 (125)	20 (20)