

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102825

Review By	rahul	Review On	10/28/2025 3:41:07 PM
Supervise By	mohammad	Supervise On	10/30/2025 5:29:45 AM
SubDirectory	BN102825	HP Acquire Method	BNA_N, 8270_HP Processing Method BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038110.D	28 Oct 2025 12:05		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN038111.D	28 Oct 2025 13:00		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN038112.D	28 Oct 2025 13:36		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN038113.D	28 Oct 2025 14:13	compound #20 kept on LR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN038114.D	28 Oct 2025 14:49		RC/JU	Ok,M
6	SSTDICC1.6	SSTDICC1.6	BN038115.D	28 Oct 2025 15:25		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN038116.D	28 Oct 2025 16:02		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN038117.D	28 Oct 2025 16:38	Compound #20 removed from 5PPM	RC/JU	Ok
9	SSTDICV0.4	ICVBN102825	BN038118.D	28 Oct 2025 17:51		RC/JU	Ok
10	PB170167BL	PB170167BL	BN038119.D	28 Oct 2025 19:04	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN111225

Review By	rahul	Review On	11/12/2025 1:44:37 PM
Supervise By	Jagrut	Supervise On	11/13/2025 10:06:02 AM
SubDirectory	BN111225	HP Acquire Method	BNA_N
		HP Processing Method	BN102825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6897,SP6898,SP6900,SP6901,SP6902,SP6903,SP6904
CCC	SP6902
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6906
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038182.D	12 Nov 2025 08:52		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038183.D	12 Nov 2025 09:31		RC/JU	Ok
3	PB170483BL	PB170483BL	BN038184.D	12 Nov 2025 10:07		RC/JU	Ok
4	PB170483BS	PB170483BS	BN038185.D	12 Nov 2025 10:44		RC/JU	Ok
5	PB170483BSD	PB170483BSD	BN038186.D	12 Nov 2025 11:20		RC/JU	Ok
6	Q3569-02	MW-12	BN038187.D	12 Nov 2025 11:56	Need 40X Dilution	RC/JU	Dilution
7	Q3584-01	SW-1	BN038188.D	12 Nov 2025 12:32		RC/JU	Ok
8	Q3569-01	MW-2	BN038189.D	12 Nov 2025 13:09	Need 10X Dilution	RC/JU	Dilution
9	Q3584-03	SW-3	BN038190.D	12 Nov 2025 13:45		RC/JU	Ok
10	Q3584-02	SW-2	BN038191.D	12 Nov 2025 14:21		RC/JU	Ok
11	Q3584-04	EB-2	BN038192.D	12 Nov 2025 14:58		RC/JU	Ok
12	Q3584-05	FB-2	BN038193.D	12 Nov 2025 15:34		RC/JU	Ok
13	Q3569-02DL	MW-12DL	BN038194.D	12 Nov 2025 16:11		RC/JU	Ok
14	Q3569-01DL	MW-2DL	BN038195.D	12 Nov 2025 16:47		RC/JU	Ok
15	SSTDCCC0.4	SSTDCCC0.4EC	BN038196.D	12 Nov 2025 18:40		RC/JU	Ok

M : Manual Integration