

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3569

Client ID : SSTDCCC0.4

Date Analyzed: 11/12/2025

Lab File ID: BN038183.D

Time Analyzed: 09:31

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9292	7.768	28984	10.57	16601	14.43
UPPER LIMIT	18584	8.268	57968	11.073	33202	14.93
LOWER LIMIT	4646	7.268	14492	10.073	8300.5	13.93
EPA SAMPLE NO.						
01 PB170483BL	9479	7.77	27842	10.57	15234	14.43
02 PB170483BS	11079	7.77	33638	10.57	18281	14.43
03 PB170483BSD	9171	7.78	26214	10.57	13453	14.43
04 MW-2	8899	7.77	29516	10.57	20929	14.43
05 MW-12	9896	7.77	33487	10.57	32721	14.43
06 MW-2DL	7979	7.78	25314	10.57	14740	14.43
07 MW-12DL	7912	7.78	24631	10.57	14683	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.