

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102925\
 Data File : BP026036.D
 Acq On : 29 Oct 2025 16:09
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP102925

Quant Time: Oct 30 11:55:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	-0.05
2	1,4-Dioxane	40.000	33.303	16.7	79	-0.02
3	Pyridine	40.000	32.670	18.3	77	-0.03
4	n-Nitrosodimethylamine	40.000	35.142	12.1	84	-0.03
5 S	2-Fluorophenol	80.000	68.429	14.5	80	-0.05
6	Aniline	40.000	35.839	10.4	84	-0.05
7 S	Phenol-d6	80.000	68.090	14.9	79	-0.05
8	2-Chlorophenol	40.000	34.596	13.5	80	-0.05
9	Benzaldehyde	40.000	46.027	-15.1	110	-0.05
10 C	Phenol	40.000	33.813	15.5	79	-0.05
11	bis(2-Chloroethyl)ether	40.000	33.880	15.3	80	-0.05
12	1,3-Dichlorobenzene	40.000	34.538	13.7	81	-0.05
13 C	1,4-Dichlorobenzene	40.000	35.099	12.3	83	-0.05
14	1,2-Dichlorobenzene	40.000	35.092	12.3	83	-0.04
15	Benzyl Alcohol	40.000	35.185	12.0	83	-0.05
16	2,2'-oxybis(1-Chloropropane	40.000	33.526	16.2	78	-0.03
17	2-Methylphenol	40.000	34.998	12.5	80	-0.04
18	Hexachloroethane	40.000	34.560	13.6	80	-0.05
19 P	n-Nitroso-di-n-propylamine	40.000	36.340	9.1	82	-0.05
20	3+4-Methylphenols	40.000	34.245	14.4	80	-0.05
21 I	Naphthalene-d8	20.000	20.000	0.0	97	-0.04
22	Acetophenone	40.000	35.756	10.6	85	-0.05
23 S	Nitrobenzene-d5	80.000	68.652	14.2	79	-0.04
24	Nitrobenzene	40.000	34.825	12.9	80	-0.04
25	Isophorone	40.000	35.069	12.3	81	-0.05
26 C	2-Nitrophenol	40.000	36.106	9.7	83	-0.04
27	2,4-Dimethylphenol	40.000	37.573	6.1	87	-0.05
28	bis(2-Chloroethoxy)methane	40.000	34.454	13.9	81	-0.03
29 C	2,4-Dichlorophenol	40.000	34.599	13.5	78	-0.03
30	1,2,4-Trichlorobenzene	40.000	35.441	11.4	84	-0.04
31	Naphthalene	40.000	33.888	15.3	80	-0.04
32	Benzoic acid	40.000	37.646	5.9	87	-0.09
33	4-Chloroaniline	40.000	38.877	2.8	90	-0.03
34 C	Hexachlorobutadiene	40.000	33.876	15.3	80	-0.04
35	Caprolactam	40.000	34.833	12.9	80	-0.07
36 C	4-Chloro-3-methylphenol	40.000	34.690	13.3	78	-0.04
37	2-Methylnaphthalene	40.000	30.988	22.5	72	-0.03
38	1-Methylnaphthalene	40.000	32.934	17.7	77	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	34.481	13.8	81	-0.03
41 P	Hexachlorocyclopentadiene	40.000	47.197	-18.0	109	-0.03
42 S	2,4,6-Tribromophenol	80.000	72.809	9.0	80	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.282	9.3	80	-0.02
44	2,4,5-Trichlorophenol	40.000	35.326	11.7	78	-0.03
45 S	2-Fluorobiphenyl	80.000	67.037	16.2	79	-0.02
46	1,1'-Biphenyl	40.000	35.073	12.3	83	-0.02
47	2-Chloronaphthalene	40.000	33.277	16.8	78	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102925\
 Data File : BP026036.D
 Acq On : 29 Oct 2025 16:09
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP102925

Quant Time: Oct 30 11:55:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	33.821	15.4	81	-0.02
49	Acenaphthylene	40.000	38.923	2.7	90	0.00
50	Dimethylphthalate	40.000	35.112	12.2	81	-0.01
51	2,6-Dinitrotoluene	40.000	37.528	6.2	91	-0.01
52 C	Acenaphthene	40.000	32.961	17.6	76	0.00
53	3-Nitroaniline	40.000	36.037	9.9	86	-0.01
54 P	2,4-Dinitrophenol	40.000	36.040	9.9	83	0.00
55	Dibenzofuran	40.000	33.938	15.2	79	0.00
56 P	4-Nitrophenol	40.000	32.703	18.2	80	-0.01
57	2,4-Dinitrotoluene	40.000	35.013	12.5	83	0.00
58	Fluorene	40.000	35.342	11.6	84	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.197	2.0	85	-0.01
60	Diethylphthalate	40.000	35.015	12.5	79	0.00
61	4-Chlorophenyl-phenylether	40.000	33.928	15.2	79	0.00
62	4-Nitroaniline	40.000	37.891	5.3	83	0.00
63	Azobenzene	40.000	35.051	12.4	81	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	35.409	11.5	84	0.00
66 c	n-Nitrosodiphenylamine	40.000	33.661	15.8	80	0.00
67	4-Bromophenyl-phenylether	40.000	32.800	18.0	78	0.00
68	Hexachlorobenzene	40.000	32.804	18.0	79	0.00
69	Atrazine	40.000	37.108	7.2	88	0.00
70 C	Pentachlorophenol	40.000	33.716	15.7	80	0.00
71	Phenanthrene	40.000	32.726	18.2	79	0.00
72	Anthracene	40.000	33.409	16.5	81	0.00
73	Carbazole	40.000	33.421	16.4	79	0.00
74	Di-n-butylphthalate	40.000	35.126	12.2	79	0.00
75 C	Fluoranthene	40.000	33.126	17.2	78	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.01
77	Benzidine	40.000	48.151	-20.4	115	0.00
78	Pyrene	40.000	32.912	17.7	76	0.00
79 S	Terphenyl-d14	80.000	67.515	15.6	80	0.00
80	Butylbenzylphthalate	40.000	33.630	15.9	81	0.00
81	Benzo(a)anthracene	40.000	33.052	17.4	78	0.01
82	3,3'-Dichlorobenzidine	40.000	37.253	6.9	87	0.00
83	Chrysene	40.000	32.526	18.7	77	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	33.750	15.6	78	0.01
85 c	Di-n-octyl phthalate	40.000	33.668	15.8	76	0.02
86 I	Perylene-d12	20.000	20.000	0.0	96	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	33.295	16.8	75	0.01
88	Benzo(b)fluoranthene	40.000	34.098	14.8	78	0.00
89	Benzo(k)fluoranthene	40.000	34.381	14.0	79	0.00
90 C	Benzo(a)pyrene	40.000	35.024	12.4	79	0.00
91	Dibenzo(a,h)anthracene	40.000	33.879	15.3	77	0.02
92	Benzo(g,h,i)perylene	40.000	33.862	15.3	77	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP102925\
Data File : BP026036.D
Acq On : 29 Oct 2025 16:09
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP102925

Quant Time: Oct 30 11:55:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 30 11:51:34 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0